

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 11/13/2018

TEMP1 IN: 103 11/12/2018 16:00 TEMP OUT: 104 11/13/2018 07:20

TEMP2 IN: 104 11/13/2018 08:00 TEMP OUT: 104 11/13/2018 09:10

TEMP3 IN: 105 11/13/2018 11:00 TEMP OUT: 105 11/13/2018 12:40

TEMP4 IN: 105 11/13/2018 13:30 TEMP OUT: 105 11/13/2018 15:10

Run Number: LB99174

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB99174BL	LB99174BL	1.3930	1.3930	100	1.3930	1.3930	1.3930	0.0000	0
2	LB99174BS	LB99174BS	1.3962	1.3962	100	1.4470	1.4470	1.4470	0.0508	508
3	J5916-01	110918-SIDI-F-D-B	1.3952	1.3952	250	1.4146	1.4146	1.4146	0.0194	77.6
4	J5916-01DUP	110918-SIDI-F-D-BDUP	1.3935	1.3935	250	1.4130	1.4130	1.4130	0.0195	78
5	J5916-02	110918-SIDI-F-D-M	1.3964	1.3964	300	1.4144	1.4144	1.4144	0.0180	60
6	J5916-03	110918-SIDI-F-D-S	1.3978	1.3978	300	1.4139	1.4139	1.4139	0.0161	53.7
7	J5916-04	110918-SIDI-F-U-B	1.3939	1.3939	300	1.4163	1.4163	1.4163	0.0224	74.7
8	J5916-05	110918-SIDI-F-U-M	1.3900	1.3900	300	1.4072	1.4072	1.4072	0.0172	57.3
9	J5916-06	110918-SIDI-F-U-S	1.3941	1.3941	300	1.4076	1.4076	1.4076	0.0135	45
10	J5916-07	110918-DBRI-F-D-B	1.3846	1.3846	300	1.4021	1.4021	1.4021	0.0175	58.3
11	J5916-08	110918-DBRI-F-D-M	1.3890	1.3890	200	1.4048	1.4048	1.4048	0.0158	79
12	J5916-09	110918-DBRI-F-D-S	1.3995	1.3995	300	1.4158	1.4158	1.4158	0.0163	54.3
13	J5916-10	110918-DBRI-F-U-B	1.3968	1.3968	200	1.4190	1.4190	1.4190	0.0222	111
14	J5916-11	110918-DBRI-F-U-M	1.3831	1.3831	200	1.4011	1.4011	1.4011	0.0180	90
15	J5916-12	110918-DBRI-F-U-S	1.4055	1.4055	300	1.4206	1.4206	1.4206	0.0151	50.3

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 11/13/2018

TEMP1 IN: 103 11/12/2018 16:00 TEMP OUT: 104 11/13/2018 07:20

TEMP2 IN: 104 11/13/2018 08:00 TEMP OUT: 104 11/13/2018 09:10

TEMP3 IN: 105 11/13/2018 11:00 TEMP OUT: 105 11/13/2018 12:40

TEMP4 IN: 105 11/13/2018 13:30 TEMP OUT: 105 11/13/2018 15:10

Run Number: LB99174

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
16	J5916-13	110918-TIPC-F-D-B	1.3926	1.3926	200	1.4092	1.4092	1.4092	0.0166	83
17	J5916-14	110918-TIPC-F-D-S	1.3961	1.3961	200	1.4100	1.4100	1.4100	0.0139	69.5
18	J5916-15	110918-TIPC-F-U-B	1.3878	1.3878	200	1.4068	1.4068	1.4068	0.0190	95
19	J5916-16	110918-TIPC-F-U-S	1.4001	1.4001	300	1.4136	1.4136	1.4136	0.0135	45
20	J5916-17	110918-JETT-F-D-B	1.3990	1.3990	200	1.4167	1.4167	1.4167	0.0177	88.5
21	J5916-18	110918-JETT-F-D-M	1.4029	1.4029	200	1.4184	1.4184	1.4184	0.0155	77.5
22	J5916-19	110918-JETT-F-D-S	1.3983	1.3983	200	1.4115	1.4115	1.4115	0.0132	66
23	J5916-20	110918-JETT-F-U-B	1.4069	1.4069	200	1.4322	1.4322	1.4322	0.0253	126.5
24	J5916-20DUP	110918-JETT-F-U-BDUP	1.4012	1.4012	200	1.4264	1.4264	1.4264	0.0252	126

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS111318

WorkList ID : 119164

Date : 11/13/2018 7:37:36 AM

11/13/18

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/16/2018	Water	J5916-01	TSS	Cool 4 deg C	HDR103	K41	110918-SIDI-F-D-B	11/09/2018	SM2540 D
11/16/2018	Water	J5916-02	TSS	Cool 4 deg C	HDR103	K41	110918-SIDI-F-D-M	11/09/2018	SM2540 D
11/16/2018	Water	J5916-03	TSS	Cool 4 deg C	HDR103	K41	110918-SIDI-F-D-S	11/09/2018	SM2540 D
11/16/2018	Water	J5916-04	TSS	Cool 4 deg C	HDR103	K41	110918-SIDI-F-U-B	11/09/2018	SM2540 D
11/16/2018	Water	J5916-05	TSS	Cool 4 deg C	HDR103	K41	110918-SIDI-F-U-M	11/09/2018	SM2540 D
11/16/2018	Water	J5916-06	TSS	Cool 4 deg C	HDR103	K41	110918-SIDI-F-U-S	11/09/2018	SM2540 D
11/16/2018	Water	J5916-07	TSS	Cool 4 deg C	HDR103	K41	110918-DBRI-F-D-B	11/09/2018	SM2540 D
11/16/2018	Water	J5916-08	TSS	Cool 4 deg C	HDR103	K41	110918-DBRI-F-D-M	11/09/2018	SM2540 D
11/16/2018	Water	J5916-09	TSS	Cool 4 deg C	HDR103	K41	110918-DBRI-F-D-S	11/09/2018	SM2540 D
11/16/2018	Water	J5916-10	TSS	Cool 4 deg C	HDR103	K41	110918-DBRI-F-U-B	11/09/2018	SM2540 D
11/16/2018	Water	J5916-11	TSS	Cool 4 deg C	HDR103	K41	110918-DBRI-F-U-M	11/09/2018	SM2540 D
11/16/2018	Water	J5916-12	TSS	Cool 4 deg C	HDR103	K41	110918-DBRI-F-U-S	11/09/2018	SM2540 D
11/14/2018	Water	J5916-13	TSS	Cool 4 deg C	HDR103	K41	110918-TIPC-F-D-B	11/09/2018	SM2540 D
11/14/2018	Water	J5916-14	TSS	Cool 4 deg C	HDR103	K41	110918-TIPC-F-D-S	11/09/2018	SM2540 D
11/14/2018	Water	J5916-15	TSS	Cool 4 deg C	HDR103	K41	110918-TIPC-F-U-B	11/09/2018	SM2540 D
11/14/2018	Water	J5916-16	TSS	Cool 4 deg C	HDR103	K41	110918-TIPC-F-U-S	11/09/2018	SM2540 D
11/16/2018	Water	J5916-17	TSS	Cool 4 deg C	HDR103	K41	110918-JETT-F-D-B	11/09/2018	SM2540 D
11/16/2018	Water	J5916-18	TSS	Cool 4 deg C	HDR103	K41	110918-JETT-F-D-M	11/09/2018	SM2540 D
11/16/2018	Water	J5916-19	TSS	Cool 4 deg C	HDR103	K41	110918-JETT-F-D-S	11/09/2018	SM2540 D
11/16/2018	Water	J5916-20	TSS	Cool 4 deg C	HDR103	K41	110918-JETT-F-U-B	11/09/2018	SM2540 D

EP-7042602

Date/Time 11/13/18, 08:00
 Received by: 
 Relinquished by: 

Date/Time 11/13/18, 13:00
 Received by: 
 Relinquished by: 