


245.1

L399176
INSTRUMENT ID: CV1

Linear

 μ Abs.:

38947



A= 0.0000e+000

B= 2.5830e-004

C= 1.9534e-002

Rho= 0.9995285

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.027	0.027	27	0.000	27				
0.05	0.050	0.048	-0.002	112	0.0 %	112				
0.2	0.200	0.202	0.002	706	0.0 %	706				
2.5	2.500	2.427	-0.073	9322	0.0 %	9322				
5.0	5.000	5.181	0.181	19981	0.0 %	19981				
7.5	7.500	7.286	-0.214	28130	0.0 %	28130				
10.0	10.000	10.080	0.080	38947	0.0 %	38947				

LB99176 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 13 Nov 2018 08:31:28

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	27	-	0.0000	245.1	PPB	13 Nov 2018 08:37:38	S	Std
0.05 - 1	50.05	112	-	0.0500	245.1	PPB	13 Nov 2018 08:39:56	S	Std
0.2 - 1	50.2	706	-	0.2000	245.1	PPB	13 Nov 2018 08:42:14	S	Std
2.5 - 1	52.5	9322	-	2.5000	245.1	PPB	13 Nov 2018 08:44:32	S	Std
5.0 - 1	55	19981	-	5.0000	245.1	PPB	13 Nov 2018 08:46:49	S	Std
7.5 - 1	57.5	28130	-	7.5000	245.1	PPB	13 Nov 2018 08:49:06	S	Std
10.0 - 1	50	38947	-	10.0000	245.1	PPB	13 Nov 2018 08:51:23	S	Std
ICV71 - 1	ICV71	14874	3.8615	-245.1		PPB	13 Nov 2018 09:02:46	U	SMPL
ICB71 - 1	ICB71	-73	0.0007	-245.1		PPB	13 Nov 2018 09:05:02	U	SMPL
CCV72 - 1	CCV72	19538	5.0662	-245.1		PPB	13 Nov 2018 09:07:18	U	SMPL
CCB72 - 1	CCB72	-107	-0.0081	-245.1		PPB	13 Nov 2018 09:09:34	U	SMPL
CRA - 1	CRA	100	0.0454	-245.1		PPB	13 Nov 2018 09:11:50	U	SMPL
HighStd - 1	HighStd	37571	9.7241	-245.1		PPB	13 Nov 2018 09:14:06	U	SMPL
ChkStd - 1	ChkStd	26374	6.8320	-245.1		PPB	13 Nov 2018 09:16:22	U	SMPL
PB114733BL - 1	PBW	-220	-0.0373	-245.1		PPB	13 Nov 2018 09:18:41	U	SMPL
PB114733BS - 1	LCSW	14017	3.6401	-245.1		PPB	13 Nov 2018 09:23:55	U	SMPL
J5904-01 - 1	110818-TIPC-F-D-B	-242	-0.0430	-245.1		PPB	13 Nov 2018 09:26:12	U	SMPL
J5904-02 - 1	110818-TIPC-F-D-S	-155	-0.0205	-245.1		PPB	13 Nov 2018 09:28:29	U	SMPL
J5904-03 - 1	110818-TIPC-F-U-B	-146	-0.0182	-245.1		PPB	13 Nov 2018 09:30:46	U	SMPL
J5904-04 - 1	110818-TIPC-F-U-S	-145	-0.0179	-245.1		PPB	13 Nov 2018 09:33:04	U	SMPL
J5904-05 - 1	110818-DBRI-F-D-B	-129	-0.0138	-245.1		PPB	13 Nov 2018 09:35:20	U	SMPL
CCV73 - 1	CCV73	19866	5.1509	-245.1		PPB	13 Nov 2018 09:37:36	U	SMPL
CCB73 - 1	CCB73	-111	-0.0091	-245.1		PPB	13 Nov 2018 09:39:53	U	SMPL
J5904-06 - 1	110818-DBRI-F-D-M	-154	-0.0202	-245.1		PPB	13 Nov 2018 09:42:08	U	SMPL
J5904-07 - 1	110818-DBRI-F-D-S	-146	-0.0182	-245.1		PPB	13 Nov 2018 09:44:24	U	SMPL
J5904-08 - 1	110818-DBRI-F-U-B	-166	-0.0233	-245.1		PPB	13 Nov 2018 09:46:41	U	SMPL
J5904-09 - 1	110818-DBRI-F-U-M	-143	-0.0174	-245.1		PPB	13 Nov 2018 09:48:57	U	SMPL
J5904-10 - 1	110818-DBRI-F-U-S	-120	-0.0115	-245.1		PPB	13 Nov 2018 09:51:14	U	SMPL
J5904-11 - 1	110818-SIDI-F-D-B	-135	-0.0153	-245.1		PPB	13 Nov 2018 09:53:31	U	SMPL
J5904-12 - 1	110818-SIDI-F-D-M	-101	-0.0066	-245.1		PPB	13 Nov 2018 10:10:58	U	SMPL
J5904-13 - 1	110818-SIDI-F-D-S	-119	-0.0112	-245.1		PPB	13 Nov 2018 10:13:16	U	SMPL
J5904-14 - 1	110818-SIDI-F-U-B	-126	-0.0130	-245.1		PPB	13 Nov 2018 10:15:33	U	SMPL
J5904-15 - 1	110818-SIDI-F-U-M	-131	-0.0143	-245.1		PPB	13 Nov 2018 10:17:49	U	SMPL
CCV74 - 1	CCV74	20294	5.2615	-245.1		PPB	13 Nov 2018 10:20:06	U	SMPL
CCB74 - 1	CCB74	-81	-0.0014	-245.1		PPB	13 Nov 2018 10:22:23	U	SMPL
J5904-16 - 1	110818-SIDI-F-U-S	-157	-0.0210	-245.1		PPB	13 Nov 2018 10:24:39	U	SMPL
J5904-17 - 1	110818-JETT-E-D-B	-142	-0.0171	-245.1		PPB	13 Nov 2018 10:26:56	U	SMPL
J5904-18 - 1	110818-JETT-E-D-M	-133	-0.0148	-245.1		PPB	13 Nov 2018 10:29:12	U	SMPL
J5904-19 - 1	110818-JETT-E-D-S	-137	-0.0159	-245.1		PPB	13 Nov 2018 10:31:29	U	SMPL
J5904-20 - 1	110818-JETT-E-U-B	-135	-0.0153	-245.1		PPB	13 Nov 2018 10:33:46	U	SMPL
J5904-21 - 1	110818-JETT-E-U-M	-120	-0.0115	-245.1		PPB	13 Nov 2018 10:36:03	U	SMPL
J5904-22 - 1	110818-JETT-E-U-S	-131	-0.0143	-245.1		PPB	13 Nov 2018 10:38:20	U	SMPL
J5904-22DUP - 1	110818-JETT-E-U-SDUP	-116	-0.0104	-245.1		PPB	13 Nov 2018 10:40:37	U	SMPL
J5904-22MS - 1	110818-JETT-E-U-SMS	13829	3.5916	-245.1		PPB	13 Nov 2018 10:42:55	U	SMPL
J5904-22MSD - 1	110818-JETT-E-U-SMSD	14013	3.6391	-245.1		PPB	13 Nov 2018 10:45:12	U	SMPL
CCV75 - 1	CCV75	19890	5.1571	-245.1		PPB	13 Nov 2018 10:47:29	U	SMPL
CCB75 - 1	CCB75	-98	-0.0058	-245.1		PPB	13 Nov 2018 10:49:45	U	SMPL
J5904-22LX5 - 1		-40	0.0092	-245.1		PPB	13 Nov 2018 10:52:03	U	SMPL
J5904-22A - 1		14972	3.8868	-245.1		PPB	13 Nov 2018 10:54:20	U	SMPL
CCB76 - 1	CCB76	19956	5.1742	-245.1		PPB	13 Nov 2018 10:56:36	U	SMPL
CCB76 - 1	CCB76	-88	-0.0032	-245.1		PPB	13 Nov 2018 10:58:52	U	SMPL