



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatelANALYST: AparnaDate: 11/13/2018TEMP1 IN: 103 11/12/2018 16:00 TEMP OUT: 104 11/13/2018 07:20TEMP2 IN: 104 11/13/2018 08:10 TEMP OUT: 104 11/13/2018 09:25TEMP3 IN: 105 11/13/2018 11:30 TEMP OUT: 105 11/13/2018 13:10TEMP4 IN: 105 11/13/2018 13:45 TEMP OUT: 105 11/13/2018 15:30Run Number: LB99177BalanceID: WC SC-3OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB99177BL	LB99177BL	1.4010	1.4010	100	1.4010	1.4010	1.4010	0.0000	0
2	LB99177BS	LB99177BS	1.4000	1.4000	100	1.4500	1.4500	1.4500	0.0500	500
3	J5907-17	TP-3-GW	1.3948	1.3948	10	2.7305	2.7305	2.7305	1.3357	133570
4	J5907-17DUP	TP-3-GWDUP	1.4051	1.4051	10	2.7400	2.7400	2.7400	1.3349	133490

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)


WorkList Name : tss111318-2

WorkList ID : 119183

Date : 11/13/2018 9:19:33 AM

2399177

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/14/2018	Water	J5907-17	TSS	Cool 4 deg C	PSEG01	I31	TP-3-GW	11/09/2018	SM2540 D

Date/Time 11/13/18, 08:00
 Received by: 
 Relinquished by: _____

Date/Time 11/13/18 13:00
 Received by: 
 Relinquished by: 