



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: Corrosivity

Supervisor Review By : JIGNESH

Run Number: LB99181

Slope : 98.7

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.3	4.01	11/12/2018	17:00
2	CAL2	1	Water			20.4	7.00	11/12/2018	17:02
3	CAL3	1	Water			20.4	10.04	11/12/2018	17:04
4	ICV	1	Water			20.5	7.01	11/12/2018	17:06
5	CCV1	1	Water			20.2	2.00	11/12/2018	18:10
6	J5909-14	1	Water			16.9	6.60	11/12/2018	18:12
7	J5909-14DUP	1	Water			17.0	6.61	11/12/2018	18:14
8	J5921-03	1	Water			11.6	6.24	11/12/2018	18:16
9	J5921-06	1	Water			11.4	7.15	11/12/2018	18:18
10	J5921-09	1	Water			14.6	6.05	11/12/2018	18:20
11	J5925-18	1	Water			18.5	7.32	11/12/2018	18:22
12	CCV2	1	Water			20.1	12.00	11/12/2018	18:24

WORKLIST(Hardcopy Internal Chain)

LB99181

WorkList Name : corrosivity111218-1

WorkList ID : 119159

Date : 11/12/2018 4:00:31 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/09/2018	Water	J5909-14	Corrosivity	Cool 4 deg C	PSEG01	I31	SB-15	11/09/2018	9040C
11/12/2018	Water	J5921-03	Corrosivity	Cool 4 deg C	PSEG01	I33	SP-TANK1-N29675	11/12/2018	9040C
11/12/2018	Water	J5921-06	Corrosivity	Cool 4 deg C	PSEG01	I33	SP-TANK2-N48963	11/12/2018	9040C
11/12/2018	Water	J5921-09	Corrosivity	Cool 4 deg C	PSEG01	I33	SP-UST-MIDDLE	11/12/2018	9040C
11/12/2018	Water	J5925-18	Corrosivity	Cool 4 deg C	PSEG01	I42	TP-5-GW	11/12/2018	9040C

Date/Time 11/12/18 6:00
 Received by: SP
 Relinquished by: CP

Date/Time 11/12/18 6:30
 Received by: CP
 Relinquished by: SP