

LB99183


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

39269



A= 0.0000e+000

B= 2.5368e-004

C= 1.7250e-002

Rho= 0.9999708

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.023	0.023	21	0.000	21				
0.2	0.200	0.201	0.001	724	0.0 %	724				
2.5	2.500	2.458	-0.042	9621	0.0 %	9621				
5.0	5.000	4.995	-0.005	19621	0.0 %	19621				
7.5	7.500	7.545	0.045	29674	0.0 %	29674				
10.0	10.000	9.979	-0.021	39269	0.0 %	39269				

LB99183 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 13 Nov 2018 11:47:53

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	21	-	0.00007470A	PPB		13 Nov 2018 11:50:30	S	Std
0.2 - 1	50.2	724	-	0.20007470A	PPB		13 Nov 2018 11:52:39	S	Std
2.5 - 1	52.5	9621	-	2.50007470A	PPB		13 Nov 2018 11:54:47	S	Std
5.0 - 1	55	19621	-	5.00007470A	PPB		13 Nov 2018 11:56:56	S	Std
7.5 - 1	57.5	29674	-	7.50007470A	PPB		13 Nov 2018 11:59:04	S	Std
10.0 - 1	510	39269	-	10.00007470A	PPB		13 Nov 2018 12:01:11	S	Std
ICV72 - 1	ICV72	15571	3.9673	-7470A	PPB		13 Nov 2018 12:03:48	U	SMPL
ICB72 - 1	ICB72	-101	-0.0084	-7470A	PPB		13 Nov 2018 12:05:54	U	SMPL
CCV77 - 1	CCV77	18612	4.7387	-7470A	PPB		13 Nov 2018 12:08:01	U	SMPL
CCB77 - 1	CCB77	-101	-0.0084	-7470A	PPB		13 Nov 2018 12:10:07	U	SMPL
CRA - 1	CRA	723	0.2007	-7470A	PPB		13 Nov 2018 12:12:14	U	SMPL
HighStd - 1	HighStd	38216	9.7119	-7470A	PPB		13 Nov 2018 12:14:20	U	SMPL
ChkStd - 1	ChkStd	27215	6.9211	-7470A	PPB		13 Nov 2018 12:16:27	U	SMPL
PB114728BL - 1	PBW	-178	-0.0279	-7470A	PPB		13 Nov 2018 12:18:36	U	SMPL
PB114728BS - 1	LCSW	13865	3.5345	-7470A	PPB		13 Nov 2018 12:20:44	U	SMPL
J5269-10 - 1	SB-12-D	-90	-0.0056	-7470A	PPB		13 Nov 2018 12:22:51	U	SMPL
J5269-14 - 1	SB-13-S	-64	0.0010	-7470A	PPB		13 Nov 2018 12:25:00	U	SMPL
J5269-18 - 1	SB-13-D	-41	0.0068	-7470A	PPB		13 Nov 2018 12:27:07	U	SMPL
J5269-22 - 1	SB-14	-42	0.0066	-7470A	PPB		13 Nov 2018 12:29:16	U	SMPL
J5766-02 - 1	HD-01-110918	-43	0.0063	-7470A	PPB		13 Nov 2018 12:31:23	U	SMPL
CCV78 - 1	CCV78	18283	4.6553	-7470A	PPB		13 Nov 2018 12:33:30	U	SMPL
CCB78 - 1	CCB78	-86	-0.0046	-7470A	PPB		13 Nov 2018 12:35:37	U	SMPL
J5766-02DUP - 1	HD-01-110918DUP	-52	0.0041	-7470A	PPB		13 Nov 2018 12:37:44	U	SMPL
J5766-02MS - 1	HD-01-110918MS	13482	3.4374	-7470A	PPB		13 Nov 2018 12:39:51	U	SMPL
J5766-02MSD - 1	HD-01-110918MSD	13311	3.3940	-7470A	PPB		13 Nov 2018 12:41:58	U	SMPL
J5887-04 - 1	W-1-S	-110	-0.0107	-7470A	PPB		13 Nov 2018 12:44:05	U	SMPL
J5887-08 - 1	W-1-D	-69	-0.0003	-7470A	PPB		13 Nov 2018 12:46:13	U	SMPL
J5887-12 - 1	W-2	-54	0.0036	-7470A	PPB		13 Nov 2018 12:48:21	U	SMPL
J5887-16 - 1	W-3-S	-27	0.0104	-7470A	PPB		13 Nov 2018 12:50:30	U	SMPL
J5888-01 - 1	HG-CON-1	-53	0.0038	-7470A	PPB		13 Nov 2018 12:52:38	U	SMPL
J5888-02 - 1	HG-CON-2	-40	0.0071	-7470A	PPB		13 Nov 2018 12:54:46	U	SMPL
J5890-01 - 1	TOWER-1	-39	0.0074	-7470A	PPB		13 Nov 2018 12:56:53	U	SMPL
CCV79 - 1	CCV79	18675	4.7547	-7470A	PPB		13 Nov 2018 12:59:00	U	SMPL
CCB79 - 1	CCB79	-80	-0.0030	-7470A	PPB		13 Nov 2018 13:01:07	U	SMPL
J5890-02 - 1	TOWER-2	-32	0.0091	-7470A	PPB		13 Nov 2018 13:03:15	U	SMPL
J5890-03 - 1	TOWER-3	2	0.0178	-7470A	PPB		13 Nov 2018 13:05:23	U	SMPL
J5892-02 - 1	PN-STONE	-51	0.0043	-7470A	PPB		13 Nov 2018 13:07:30	U	SMPL
J5892-03 - 1	PN-CONCRETE	-37	0.0079	-7470A	PPB		13 Nov 2018 13:09:38	U	SMPL
J5894-05 - 1	COMP-2456	-14	0.0137	-7470A	PPB		13 Nov 2018 13:11:45	U	SMPL
J5894-10 - 1	COMP-891011	-45	0.0058	-7470A	PPB		13 Nov 2018 13:13:53	U	SMPL
PB114729BL - 1	PBW	-21	0.0119	-7470A	PPB		13 Nov 2018 13:16:01	U	SMPL
PB114729BS - 1	LCSW	12825	3.2707	-7470A	PPB		13 Nov 2018 13:18:09	U	SMPL
J5901-08 - 1	SS-2	-111	-0.0109	-7470A	PPB		13 Nov 2018 13:20:17	U	SMPL
J5905-01 - 1	06-TANK-SOIL	-57	0.0028	-7470A	PPB		13 Nov 2018 13:22:24	U	SMPL
CCV80 - 1	CCV80	18822	4.7920	-7470A	PPB		13 Nov 2018 13:24:32	U	SMPL
CCB80 - 1	CCB80	-49	0.0048	-7470A	PPB		13 Nov 2018 13:26:39	U	SMPL
J5907-04 - 1	TP-2-S	-9	0.0150	-7470A	PPB		13 Nov 2018 13:28:47	U	SMPL
J5907-08 - 1	TP-3-S	-44	0.0061	-7470A	PPB		13 Nov 2018 13:30:54	U	SMPL
J5907-12 - 1	TP-3-D	-26	0.0107	-7470A	PPB		13 Nov 2018 13:33:01	U	SMPL
J5907-16 - 1	TP-4-S	6	0.0188	-7470A	PPB		13 Nov 2018 13:35:08	U	SMPL
J5907-16DUP - 1	TP-4-SDUP	-11	0.0145	-7470A	PPB		13 Nov 2018 13:37:16	U	SMPL
J5907-16MS - 1	TP-4-SMS	13734	3.5013	-7470A	PPB		13 Nov 2018 13:39:24	U	SMPL
J5907-16MSD - 1	TP-4-SMSD	13360	3.4064	-7470A	PPB		13 Nov 2018 13:41:32	U	SMPL
PB114734BL - 1	PBW	-95	-0.0068	-7470A	PPB		13 Nov 2018 13:43:40	U	SMPL
PB114734BS - 1	LCSW	13578	3.4617	-7470A	PPB		13 Nov 2018 13:45:49	U	SMPL
CCV81 - 1	CCV81	18765	4.7775	-7470A	PPB		13 Nov 2018 13:47:56	U	SMPL
CCB81 - 1	CCB81	-58	0.0025	-7470A	PPB		13 Nov 2018 13:50:03	U	SMPL
J5907-17 - 1	TP-3-GW	400900	HIGH	-7470A	PPB		13 Nov 2018 13:52:11	X	
J5907-17DUP - 1	TP-3-GWDUP	363581	HIGH	-7470A	PPB		13 Nov 2018 13:54:02	X	
J5907-17MS - 1	TP-3-GWMS	331518	HIGH	-7470A	PPB		13 Nov 2018 13:55:58	X	
J5907-17MSD - 1	TP-3-GWMSD	345935	HIGH	-7470A	PPB		13 Nov 2018 13:57:53	X	
J5909-13 - 1	SB-15	14087	3.5908	-7470A	PPB		13 Nov 2018 13:59:48	U	SMPL
PB114728TB - 1	PB114677TB	-188	-0.0304	-7470A	PPB		13 Nov 2018 14:01:58	U	SMPL
PB114729TB - 1	PB114703TB	-49	0.0048	-7470A	PPB		13 Nov 2018 14:04:07	U	SMPL
CCV82 - 1	CCV82	18218	4.6388	-7470A	PPB		13 Nov 2018 14:06:15	U	SMPL
CCB82 - 1	CCB82	-84	-0.0041	-7470A	PPB		13 Nov 2018 14:08:23	U	SMPL
J5766-02LX5 - 1		0	0.0173	-7470A	PPB		13 Nov 2018 14:10:32	U	SMPL
J5766-02A - 1		14507	3.6974	-7470A	PPB		13 Nov 2018 14:12:40	U	SMPL
J5907-16LX5 - 1		-87	-0.0048	-7470A	PPB		13 Nov 2018 14:14:47	U	SMPL
J5907-16A - 1		14640	3.7311	-7470A	PPB		13 Nov 2018 14:16:55	U	SMPL
J5907-17LX5 - 1		95193	24.1657	-7470A	PPB		13 Nov 2018 14:19:03	U	SMPL
J5907-17A - 1		380698	HIGH	-7470A	PPB		13 Nov 2018 14:21:10	X	
5907-17DLX10 - 1		46995	11.9389	-7470A	PPB		13 Nov 2018 14:23:04	U	SMPL
5907-17DUPDLX10 - 1		46840	11.8996	-7470A	PPB		13 Nov 2018 14:25:15	U	SMPL
5907-17MSDLX10 - 1		41829	10.6284	-7470A	PPB		13 Nov 2018 14:27:25	U	SMPL
5907-17MSDDLX10 - 1		43634	11.0863	-7470A	PPB		13 Nov 2018 14:29:32	U	SMPL
CCV83 - 1	CCV83	18158	4.6236	-7470A	PPB		13 Nov 2018 14:31:41	U	SMPL
CCB83 - 1	CCB83	-84	-0.0041	-7470A	PPB		13 Nov 2018 14:33:47	U	SMPL
5907-17LDLX50 - 1		9744	2.4891	-7470A	PPB		13 Nov 2018 14:35:54	U	SMPL
5907-17ADLX10 - 1		50047	12.7131	-7470A	PPB		13 Nov 2018 14:38:02	U	SMPL
5907-17DL2X50 - 1		9583	2.4483	-7470A	PPB		13 Nov 2018 14:40:10	U	SMPL

LB99183 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 13 Nov 2018 11:47:53

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
J5907-17DUPDL2X50 - 1		9702	2.4784		-7470A	PPB	13 Nov 2018 14:45:01	U	SMPL
J5907-17MSDL2X50 - 1		10272	2.6230		-7470A	PPB	13 Nov 2018 14:47:09	U	SMPL
J5907-17MSDDL2X50 - 1		11642	2.9706		-7470A	PPB	13 Nov 2018 14:49:17	U	SMPL
J5907-17LDL2X250 - 1		1848	0.4860		-7470A	PPB	13 Nov 2018 14:51:24	U	SMPL
J5907-17ADL2X50 - 1		10266	2.6215		-7470A	PPB	13 Nov 2018 14:53:32	U	SMPL
CCV84 - 1	CCV84	18324	4.6657		-7470A	PPB	13 Nov 2018 14:55:40	U	SMPL
CCB84 - 1	CCB84	-65	0.0008		-7470A	PPB	13 Nov 2018 14:57:48	U	SMPL