



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: pH

Supervisor Review By : JIGNESH

Run Number: LB99196

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

True Value of CCV4 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.1	4.01	11/13/2018	13:40
2	CAL2	1	Water			20.2	7.00	11/13/2018	13:42
3	CAL3	1	Water			20.1	10.04	11/13/2018	13:44
4	ICV	1	Water			20.3	7.01	11/13/2018	13:46
5	CCV1	1	Water			20.2	2.00	11/13/2018	13:48
6	J5865-01	1	Solid	20.02	20	18.4	8.13	11/13/2018	13:50
7	J5865-02	1	Solid	20.04	20	18.5	8.19	11/13/2018	13:52
8	J5865-03	1	Solid	20.02	20	18.6	8.34	11/13/2018	13:54
9	J5865-04	1	Solid	20.03	20	18.0	8.15	11/13/2018	13:56
10	J5865-05	1	Solid	20.03	20	18.1	8.00	11/13/2018	13:58
11	J5865-06	1	Solid	20.02	20	18.1	8.27	11/13/2018	14:00
12	J5865-07	1	Solid	20.04	20	18.0	6.54	11/13/2018	14:02
13	J5865-08	1	Solid	20.02	20	18.2	5.94	11/13/2018	14:04
14	J5865-09	1	Solid	20.03	20	18.1	5.97	11/13/2018	14:06
15	J5865-10	1	Solid	20.03	20	18.0	5.95	11/13/2018	14:08
16	CCV2	1	Water			20.1	12.00	11/13/2018	14:10
17	J5865-10DUP	1	Solid	20.02	20	18.2	5.96	11/13/2018	14:12
18	J5865-13	1	Solid	20.03	20	18.1	4.43	11/13/2018	14:14
19	J5865-14	1	Solid	20.04	20	18.4	4.30	11/13/2018	14:16
20	J5865-15	1	Solid	20.02	20	18.1	4.36	11/13/2018	14:18

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	J5865-16	1	Solid	20.04	20	18.2	3.93	11/13/2018	14:20
22	J5865-17	1	Solid	20.02	20	18.4	3.78	11/13/2018	14:22
23	J5865-18	1	Solid	20.03	20	19.3	4.02	11/13/2018	14:24
24	J5865-19	1	Solid	20.02	20	19.3	3.95	11/13/2018	14:26
25	J5865-20	1	Solid	20.03	20	19.3	4.00	11/13/2018	14:28
26	J5865-21	1	Solid	20.02	20	19.4	4.06	11/13/2018	14:30
27	CCV3	1	Water			20.2	2.00	11/13/2018	14:32
28	J5865-21DUP	1	Solid	20.04	20	19.5	4.07	11/13/2018	14:34
29	CCV4	1	Water			20.3	12.00	11/13/2018	14:36

WORKLIST(Hardcopy Internal Chain)

LB99196

WorkList Name : pH111318

WorkList ID : 119209

Date : 11/13/2018 1:18:25 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/06/2018	Solid	J5865-01	pH	Cool 4 deg C	USEP01	A11	MYAJK5	11/06/2018	9045D
11/06/2018	Solid	J5865-02	pH	Cool 4 deg C	USEP01	A11	MYAJL9	11/06/2018	9045D
11/06/2018	Solid	J5865-03	pH	Cool 4 deg C	USEP01	A11	MYAJM1	11/06/2018	9045D
11/06/2018	Solid	J5865-04	pH	Cool 4 deg C	USEP01	A11	MYAJM2	11/06/2018	9045D
11/06/2018	Solid	J5865-05	pH	Cool 4 deg C	USEP01	A11	MYAJM3	11/06/2018	9045D
11/06/2018	Solid	J5865-06	pH	Cool 4 deg C	USEP01	A11	MYAJM4	11/06/2018	9045D
11/07/2018	Solid	J5865-07	pH	Cool 4 deg C	USEP01	A11	MYAJN4	11/07/2018	9045D
11/07/2018	Solid	J5865-08	pH	Cool 4 deg C	USEP01	A11	MYAJL5	11/07/2018	9045D
11/07/2018	Solid	J5865-09	pH	Cool 4 deg C	USEP01	A11	MYAJL6	11/07/2018	9045D
11/07/2018	Solid	J5865-10	pH	Cool 4 deg C	USEP01	A11	MYAJL7	11/07/2018	9045D
11/09/2018	Solid	J5865-13	pH	Cool 4 deg C	USEP01	A11	MYAJG5	11/09/2018	9045D
11/09/2018	Solid	J5865-14	pH	Cool 4 deg C	USEP01	A11	MYAJG6	11/09/2018	9045D
11/09/2018	Solid	J5865-15	pH	Cool 4 deg C	USEP01	A11	MYAJJ1	11/09/2018	9045D
11/09/2018	Solid	J5865-16	pH	Cool 4 deg C	USEP01	A11	MYAJJ2	11/09/2018	9045D
11/09/2018	Solid	J5865-17	pH	Cool 4 deg C	USEP01	A11	MYAJJ6	11/09/2018	9045D
11/09/2018	Solid	J5865-18	pH	Cool 4 deg C	USEP01	A11	MYAJJ7	11/09/2018	9045D
11/09/2018	Solid	J5865-19	pH	Cool 4 deg C	USEP01	A11	MYAJJ8	11/09/2018	9045D
11/09/2018	Solid	J5865-20	pH	Cool 4 deg C	USEP01	A11	MYAJK2	11/09/2018	9045D
11/09/2018	Solid	J5865-21	pH	Cool 4 deg C	USEP01	A11	MYAJK3	11/09/2018	9045D

Date/Time 11/13/18 1:30
 Received by: SP
 Relinquished by: CP

Date/Time 11/13/18 2:45
 Received by: CP
 Relinquished by: SP