



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: sweta

Run Number: LB99197

Ambient Barometric Pressure (mmHg): 745

Barometric Scale ID: 55971

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	81.5	11/13/2018	10:20
2	J5909-14		1	8	>100.0	>212.0	>212.0	11/13/2018	10:40
3	J5909-14DUP		1	8	>100.0	>212.0	>212.0	11/13/2018	11:02
4	J5921-03		1	9	>100.0	>212.0	>212.0	11/13/2018	11:35
5	J5921-06		1	9	>100.0	>212.0	>212.0	11/13/2018	11:58
6	J5921-09		1	6	>100.0	>212.0	>212.0	11/13/2018	12:20
7	J5925-18		1	8	>100.0	>212.0	>212.0	11/13/2018	12:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB99197

WorkList Name : fp111318

WorkList ID : 119192

Date : 11/13/2018 9:34:16 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/14/2018	Water	J5909-14 B	Flash Point	Cool 4 deg C	PSEG01	I31	SB-15	11/09/2018	1010A
11/15/2018	Water	J5921-03 B	Flash Point	Cool 4 deg C	PSEG01	I33	SP-TANK1-N29675	11/12/2018	1010A
11/15/2018	Water	J5921-06 B	Flash Point	Cool 4 deg C	PSEG01	I33	SP-TANK2-N48963	11/12/2018	1010A
11/15/2018	Water	J5921-09 B	Flash Point	Cool 4 deg C	PSEG01	I33	SP-UST-MIDDLE	11/12/2018	1010A
11/15/2018	Water	J5925-18 B	Flash Point	Cool 4 deg C	PSEG01	I42	TP-5-GW	11/12/2018	1010A

Date/Time 11/13/18 9:45 AM

Received by: KS

Relinquished by: CP

Date/Time 11/13/18 12:50 PM

Received by: CP

Relinquished by: KS