



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : JIGNESH

Run Number: LB99207

Slope : 99.0

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.2	4.01	11/13/2018	17:30
2	CAL2	1	Water			20.4	7.00	11/13/2018	17:32
3	CAL3	1	Water			20.2	10.04	11/13/2018	17:34
4	ICV	1	Water			20.1	7.01	11/13/2018	17:36
5	CCV1	1	Water			20.2	2.00	11/13/2018	17:38
6	J5923-02	1	Solid	20.02	20	18.0	8.84	11/13/2018	17:40
7	J5923-02DUP	1	Solid	20.03	20	18.1	8.85	11/13/2018	17:42
8	J5925-04	1	Solid	20.03	20	19.5	8.42	11/13/2018	17:44
9	J5925-08	1	Solid	20.04	20	19.6	9.14	11/13/2018	17:46
10	J5925-12	1	Solid	20.02	20	19.4	9.02	11/13/2018	17:48
11	J5925-16	1	Solid	20.02	20	19.3	7.94	11/13/2018	17:50
12	J5929-01	1	Solid	20.03	20	19.2	9.27	11/13/2018	17:52
13	J5930-02	1	Solid	20.04	20	19.1	9.26	11/13/2018	17:54
14	J5932-04	1	Solid	20.02	20	19.2	8.44	11/13/2018	17:56
15	J5940-04	1	Solid	20.02	20	19.1	7.92	11/13/2018	17:58
16	CCV2	1	Water			20.1	12.00	11/13/2018	18:06

WORKLIST(Hardcopy Internal Chain)

LB99207

WorkList Name : corrosivity11318

WorkList ID : 119215

Date : 11/13/2018 5:05:09 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/12/2018	Solid	J5923-02	Corrosivity	Cool 4 deg C	PSEG01	I33	MS-OILY-STONE	11/12/2018	9045D
11/12/2018	Solid	J5925-04	Corrosivity	Cool 4 deg C	PSEG01	I42	TP-1S	11/12/2018	9045D
11/12/2018	Solid	J5925-08	Corrosivity	Cool 4 deg C	PSEG01	I42	TP-1D	11/12/2018	9045D
11/12/2018	Solid	J5925-12	Corrosivity	Cool 4 deg C	PSEG01	I42	TP-5S	11/12/2018	9045D
11/12/2018	Solid	J5925-16	Corrosivity	Cool 4 deg C	PSEG01	I42	TP-6S	11/12/2018	9045D
11/08/2018	Solid	J5929-01	Corrosivity	Cool 4 deg C	SKAN05	L11	207-WEST-TRENCH	11/08/2018	9045D
11/13/2018	Solid	J5930-02	Corrosivity	Cool 4 deg C	PSEG01	L11	32727-45914-21716	11/13/2018	9045D
11/13/2018	Solid	J5932-04	Corrosivity	Cool 4 deg C	PSEG01	L11	CG-SOIL	11/13/2018	9045D
11/12/2018	Solid	J5940-04	Corrosivity	Cool 4 deg C	WSPO01	L13	B-25(0-10)	11/12/2018	9045D

Date/Time 11/13/18 5:20
 Received by: SP
 Relinquished by: CP

Date/Time 11/13/18 6:10
 Received by: CP
 Relinquished by: SP