



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: pH

Supervisor Review By : JIGNESH

Run Number: LB99209

Slope : 99.0

pH Meter ID : pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.2	4.01	11/13/2018	17:30
2	CAL2	1	Water			20.4	7.00	11/13/2018	17:32
3	CAL3	1	Water			20.2	10.04	11/13/2018	17:34
4	ICV	1	Water			20.1	7.01	11/13/2018	17:36
5	CCV1	1	Water			20.2	2.00	11/13/2018	18:00
6	J5939-01	1	Water			13.9	9.17	11/13/2018	18:02
7	J5939-01DUP	1	Water			14.0	9.18	11/13/2018	18:04
8	CCV2	1	Water			20.1	12.00	11/13/2018	18:06

WORKLIST(Hardcopy Internal Chain)

LBa209

WorkList Name : pH111318-1

WorkList ID : 119217

Date : 11/13/2018 5:05:31 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/13/2018	Water	J5939-01	pH	Cool 4 deg C	PSEG01	L13	OILY-WATER	11/13/2018	9040C

Date/Time 11/13/18 5:20
 Received by: SP
 Relinquished by: CP

Date/Time 11/13/18 6:10
 Received by: CP
 Relinquished by: SP