

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB99213

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP65054
calibration std. phosphate 1 ppm	WP69554
calibration std. phosphate 0.5 ppm	WP69553
calibration std. phosphate 0.3 ppm	WP69552
calibration std. 0 ppm	WP69549
phosphate ICV std.	WP69556
phosphate CCV std.	WP69555
calibration std. phosphate 0.1 ppm	WP69551
calibration std. phosphate 0.05 ppm	WP69550
Combined reagent	WP69559
11 N sulfuric acid	WP66905
Phenolphthalein indicator	WP66907

Intercept: -0.0009Slope: 0.6504Regression: 0.999829

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	11/14/2018	11:15
2	CAL2	0.05	1	50	50	0.033	11/14/2018	11:15
3	CAL3	0.10	1	50	50	0.065	11/14/2018	11:16
4	CAL4	0.30	1	50	50	0.186	11/14/2018	11:16
5	CAL5	0.50	1	50	50	0.330	11/14/2018	11:17
6	CAL6	1.00	1	50	50	0.649	11/14/2018	11:17



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB99213

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.331	0.510	11/14/2018	11:18
2	ICB		1	50	50	0.000	0.001	11/14/2018	11:18
3	CCV1	0.50	1	50	50	0.329	0.507	11/14/2018	11:29
4	CCB1		1	50	50	0.000	0.001	11/14/2018	11:29
5	PB114816BL		1	50	50	0.000	0.001	11/14/2018	11:30
6	PB114816BS	0.50	1	50	50	0.315	0.486	11/14/2018	11:30
7	J5920-01		1	50	50	0.019	0.031	11/14/2018	11:31
8	J5920-01DUP		1	50	50	0.020	0.032	11/14/2018	11:31
9	J5920-01MS	0.50	1	50	50	0.339	0.523	11/14/2018	11:32
10	J5920-01MSD	0.50	1	50	50	0.342	0.527	11/14/2018	11:32
11	CCV2	0.50	1	50	50	0.328	0.506	11/14/2018	11:33
12	CCB2		1	50	50	0.000	0.001	11/14/2018	11:33