

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus, TotalSUPERVISOR REVIEW BY: apatelRun Number: LB99215BALANCE ID: WC SC-4

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP65054
calibration std. phosphate 1 ppm	WP69554
calibration std. phosphate 0.5 ppm	WP69553
calibration std. phosphate 0.3 ppm	WP69552
calibration std. 0 ppm	WP69549
phosphate ICV std.	WP69556
phosphate CCV std.	WP69555
calibration std. phosphate 0.1 ppm	WP69551
calibration std. phosphate 0.05 ppm	WP69550
Combined reagent	WP69559
11 N sulfuric acid	WP66905
Phenolphthalein indicator	WP66907
Sodium hydroxide, 1N	WP66924

Intercept: -0.0009Slope: 0.6504Regression: 0.999829

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	11/14/2018	11:15
2	CAL2	0.05	1	50	50	0.033	11/14/2018	11:15
3	CAL3	0.10	1	50	50	0.065	11/14/2018	11:16
4	CAL4	0.30	1	50	50	0.186	11/14/2018	11:16
5	CAL5	0.50	1	50	50	0.330	11/14/2018	11:17
6	CAL6	1.00	1	50	50	0.649	11/14/2018	11:17



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus, Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB99215

BALANCE ID: WC SC-4

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.331	0.510	11/14/2018	11:18
2	ICB		1	50	50	0.000	0.001	11/14/2018	11:18
3	CCV1	0.50	1	50	50	0.328	0.506	11/14/2018	11:19
4	CCB1		1	50	50	0.000	0.001	11/14/2018	11:19
5	PB114810BL		1	1.00	50	0.000	0.001	11/14/2018	11:19
6	PB114810BS	0.50	1	1.01	50	0.324	0.500	11/14/2018	11:19
7	J5807-09		1	1.01	50	1.533	2.358	11/14/2018	11:20
8	J5807-09		50	1.01	50	0.630	0.970	11/14/2018	11:28
9	CCV2	0.50	1	50	50	0.327	0.504	11/14/2018	11:29
10	CCB2		1	50	50	0.000	0.001	11/14/2018	11:29