



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: sweta

Run Number: LB99218

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 55971

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.9	11/14/2018	12:45
2	J5939-02		1	8	>100.0	>212.0	>212.0	11/14/2018	13:20
3	J5939-02DUP		1	8	>100.0	>212.0	>212.0	11/14/2018	13:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB99218

WorkList Name : fp111418

WorkList ID : 119249

Date : 11/14/2018 12:10:26 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/16/2018	Water	J5939-02	Flash Point	Cool 4 deg C	PSEG01	L13	OILY-WATER	11/13/2018	1010A

Date/Time 11/14/18 12:15 PM
 Received by: KS
 Relinquished by: CP

Date/Time 11/14/18 13:50
 Received by: CP
 Relinquished by: KS