



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: pH

Supervisor Review By : ketankumar

Run Number: LB99230

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.1	4.01	11/14/2018	14:20
2	CAL2	1	Water			20.2	7.00	11/14/2018	14:22
3	CAL3	1	Water			20.1	10.04	11/14/2018	14:24
4	ICV	1	Water			20.2	7.01	11/14/2018	14:26
5	CCV1	1	Water			20.3	2.00	11/14/2018	14:28
6	J5926-01	1	Solid	20.02	20	17.7	6.72	11/14/2018	14:30
7	J5926-01DUP	1	Solid	20.03	20	17.8	6.73	11/14/2018	14:32
8	J5926-04	1	Solid	20.02	20	17.3	7.19	11/14/2018	14:34
9	J5926-05	1	Solid	20.04	20	17.9	7.70	11/14/2018	14:36
10	J5926-06	1	Solid	20.02	20	17.9	6.80	11/14/2018	14:38
11	J5926-07	1	Solid	20.03	20	17.5	8.43	11/14/2018	14:40
12	J5926-08	1	Solid	20.02	20	17.7	7.64	11/14/2018	14:42
13	J5926-09	1	Solid	20.03	20	17.8	7.62	11/14/2018	14:44
14	J5926-10	1	Solid	20.02	20	17.6	8.52	11/14/2018	14:46
15	J5926-11	1	Solid	20.02	20	17.7	8.33	11/14/2018	14:48
16	CCV2	1	Water			20.1	12.00	11/14/2018	14:50
17	J5926-12	1	Solid	20.03	20	17.7	7.90	11/14/2018	14:52
18	CCV3	1	Water			20.2	2.00	11/14/2018	14:54

WORKLIST(Hardcopy Internal Chain)

1899230

WorkList Name : pH111418

WorkList ID : 119252

Date : 11/14/2018 2:07:02 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/08/2018	Solid	J5926-01	pH	Cool 4 deg C	USEP01	A21	MYAJG9	11/08/2018	9045D
11/08/2018	Solid	J5926-04	pH	Cool 4 deg C	USEP01	A21	MYAJH0	11/08/2018	9045D
11/12/2018	Solid	J5926-05	pH	Cool 4 deg C	USEP01	A21	MYAJG7	11/12/2018	9045D
11/12/2018	Solid	J5926-06	pH	Cool 4 deg C	USEP01	A21	MYAJG8	11/12/2018	9045D
11/12/2018	Solid	J5926-07	pH	Cool 4 deg C	USEP01	A21	MYAJH8	11/12/2018	9045D
11/12/2018	Solid	J5926-08	pH	Cool 4 deg C	USEP01	A21	MYAJJ0	11/12/2018	9045D
11/12/2018	Solid	J5926-09	pH	Cool 4 deg C	USEP01	A21	MYAJL8	11/12/2018	9045D
11/12/2018	Solid	J5926-10	pH	Cool 4 deg C	USEP01	A21	MYAJN1	11/12/2018	9045D
11/12/2018	Solid	J5926-11	pH	Cool 4 deg C	USEP01	A21	MYAJN2	11/12/2018	9045D
11/12/2018	Solid	J5926-12	pH	Cool 4 deg C	USEP01	A21	MYAJN3	11/12/2018	9045D

Date/Time 11/14/18 2:10
 Received by: SP
 Relinquished by: CP

Date/Time 11/14/18 3:05
 Received by: CP
 Relinquished by: SP