



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : ketankumar

Run Number: LB99235

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.3	4.01	11/14/2018	17:05
2	CAL2	1	Water			20.1	7.00	11/14/2018	17:07
3	CAL3	1	Water			20.3	10.03	11/14/2018	17:09
4	ICV	1	Water			20.1	7.01	11/14/2018	17:11
5	CCV1	1	Water			20.3	2.00	11/14/2018	17:13
6	J5767-10	1	Solid	20.02	20	18.2	8.35	11/14/2018	17:15
7	J5767-10DUP	1	Solid	20.03	20	18.3	8.36	11/14/2018	17:17
8	J5767-12	1	Solid	20.02	20	18.2	8.16	11/14/2018	17:19
9	J5767-14	1	Solid	20.02	20	18.0	8.36	11/14/2018	17:21
10	J5767-16	1	Solid	20.04	20	18.0	8.25	11/14/2018	17:23
11	J5942-04	1	Solid	20.02	20	18.0	8.14	11/14/2018	17:25
12	J5942-08	1	Solid	20.03	20	17.9	8.53	11/14/2018	17:27
13	J5942-12	1	Solid	20.02	20	17.8	8.55	11/14/2018	17:29
14	CCV2	1	Water			20.1	12.00	11/14/2018	17:31

WORKLIST(Hardcopy Internal Chain)

UP 99235

WorkList Name : corrosivity111418

WorkList ID : 119259

Date : 11/14/2018 4:52:28 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/14/2018	Solid	J5767-10	Corrosivity	Cool 4 deg C	PSEG05		CL-01-111418-A	11/14/2018	9045D
11/14/2018	Solid	J5767-12	Corrosivity	Cool 4 deg C	PSEG05		CL-01-111418-B	11/14/2018	9045D
11/14/2018	Solid	J5767-14	Corrosivity	Cool 4 deg C	PSEG05		CL-01-111418-C	11/14/2018	9045D
11/14/2018	Solid	J5767-16	Corrosivity	Cool 4 deg C	PSEG05		CL-01-111418-D	11/14/2018	9045D
11/14/2018	Solid	J5942-04	Corrosivity	Cool 4 deg C	PSEG01	L13	VW-D1	11/14/2018	9045D
11/14/2018	Solid	J5942-08	Corrosivity	Cool 4 deg C	PSEG01	L13	VW-F1	11/14/2018	9045D
11/14/2018	Solid	J5942-12	Corrosivity	Cool 4 deg C	PSEG01	L13	VW-F2	11/14/2018	9045D

Date/Time 11/14/18 4:55
 Received by: SP
 Relinquished by: SP

Date/Time 11/14/18 5:35
 Received by: SP
 Relinquished by: SP