

LB 99242


7470A

INSTRUMENT ID : CV1

Linear

 μ Abs.:

37060



A= 0.0000e+000

B= 2.6648e-004

C= -8.7797e-003

Rho= 0.9997135

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.010	-0.010	-6	0.000	-6				
0.2	0.200	0.182	-0.018	716	0.0 %	716				
2.5	2.500	2.474	-0.026	9318	0.0 %	9318				
5.0	5.000	5.021	0.021	18874	0.0 %	18874				
7.5	7.500	7.666	0.166	28801	0.0 %	28801				
10.0	10.000	9.867	-0.133	37060	0.0 %	37060				

LB99242 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 15 Nov 2018 09:20:03

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-6	-	0.0000	7470A	PPB	15 Nov 2018 09:26:51	S	Std
0.2 - 1	50.2	716	-	0.2000	7470A	PPB	15 Nov 2018 09:33:01	S	Std
2.5 - 1	52.5	9318	-	2.5000	7470A	PPB	15 Nov 2018 09:35:10	S	Std
5.0 - 1	55	18874	-	5.0000	7470A	PPB	15 Nov 2018 09:37:18	S	Std
7.5 - 1	57.5	28801	-	7.5000	7470A	PPB	15 Nov 2018 09:39:26	S	Std
10.0 - 1	510	37060	-	10.0000	7470A	PPB	15 Nov 2018 09:41:33	S	Std
ICV75 - 1	ICV75	13632	3.6239	-7470A	PPB	15 Nov 2018 09:44:12	U	SMPL	
ICB75 - 1	ICB75	-103	-0.0362	-7470A	PPB	15 Nov 2018 09:46:19	U	SMPL	
CCV11 - 1	CCV11	18946	5.0400	-7470A	PPB	15 Nov 2018 09:48:25	U	SMPL	
CCB11 - 1	CCB11	-111	-0.0384	-7470A	PPB	15 Nov 2018 09:50:32	U	SMPL	
CRA - 1	CRA	716	0.1820	-7470A	PPB	15 Nov 2018 09:52:38	U	SMPL	
HighStd - 1	HighStd	37334	9.9401	-7470A	PPB	15 Nov 2018 09:54:45	U	SMPL	
ChkStd - 1	ChkStd	27061	7.2025	-7470A	PPB	15 Nov 2018 09:56:52	U	SMPL	
PB114823BL - 1	PBW	-159	-0.0512	-7470A	PPB	15 Nov 2018 09:59:00	U	SMPL	
PB114823BS - 1	LCSW	15146	4.0274	-7470A	PPB	15 Nov 2018 10:01:09	U	SMPL	
J5925-04 - 1	TP-1S	-109	-0.0378	-7470A	PPB	15 Nov 2018 10:03:16	U	SMPL	
J5925-04DUP - 1	TP-1SDUP	-31	-0.0170	-7470A	PPB	15 Nov 2018 10:05:24	U	SMPL	
J5925-04MS - 1	TP-1SMS	12939	3.4392	-7470A	PPB	15 Nov 2018 10:07:33	U	SMPL	
J5925-04MSD - 1	TP-1SMSD	12987	3.4520	-7470A	PPB	15 Nov 2018 10:09:41	U	SMPL	
J5925-08 - 1	TP-1D	-94	-0.0338	-7470A	PPB	15 Nov 2018 10:11:48	U	SMPL	
CCV12 - 1	CCV12	18883	5.0232	-7470A	PPB	15 Nov 2018 10:13:54	U	SMPL	
CCB12 - 1	CCB12	-107	-0.0373	-7470A	PPB	15 Nov 2018 10:16:01	U	SMPL	
J5925-12 - 1	TP-5S	-46	-0.0210	-7470A	PPB	15 Nov 2018 10:18:07	U	SMPL	
J5925-16 - 1	TP-6S	-25	-0.0154	-7470A	PPB	15 Nov 2018 10:20:14	U	SMPL	
J5929-01 - 1	207-WEST-TRENCH	20	-0.0035	-7470A	PPB	15 Nov 2018 10:22:21	U	SMPL	
J5930-02 - 1	32727-45914-21716	-1	-0.0090	-7470A	PPB	15 Nov 2018 10:24:28	U	SMPL	
J5932-04 - 1	CG-SOIL	-10	-0.0114	-7470A	PPB	15 Nov 2018 10:26:35	U	SMPL	
J5940-04 - 1	B-25(0-10)	-5	-0.0101	-7470A	PPB	15 Nov 2018 10:28:43	U	SMPL	
PB114824BL - 1	PBW	4	-0.0077	-7470A	PPB	15 Nov 2018 10:30:51	U	SMPL	
PB114824BS - 1	LCSW	13989	3.7190	-7470A	PPB	15 Nov 2018 10:33:00	U	SMPL	
J5939-01 - 1	OILY-WATER	9378	2.4903	-7470A	PPB	15 Nov 2018 10:35:08	U	SMPL	
J5939-01DUP - 1	OILY-WATERDUP	9142	2.4274	-7470A	PPB	15 Nov 2018 10:37:15	U	SMPL	
CCV13 - 1	CCV13	18754	4.9888	-7470A	PPB	15 Nov 2018 10:39:24	U	SMPL	
CCB13 - 1	CCB13	-88	-0.0322	-7470A	PPB	15 Nov 2018 10:41:31	U	SMPL	
J5939-01MS - 1	OILY-WATERMS	23464	6.2440	-7470A	PPB	15 Nov 2018 10:43:39	U	SMPL	
J5939-01MSD - 1	OILY-WATERMSD	23196	6.1725	-7470A	PPB	15 Nov 2018 10:45:46	U	SMPL	
PB114823TB - 1	PB114786TB	-147	-0.0480	-7470A	PPB	15 Nov 2018 10:47:55	U	SMPL	
J5925-04LX5 - 1		-14	-0.0125	-7470A	PPB	15 Nov 2018 10:50:03	U	SMPL	
J5925-04A - 1		15003	3.9893	-7470A	PPB	15 Nov 2018 10:52:11	U	SMPL	
J5939-01LX5 - 1		1641	0.4285	-7470A	PPB	15 Nov 2018 10:54:19	U	SMPL	
J5939-01A - 1		24765	6.5907	-7470A	PPB	15 Nov 2018 10:56:27	U	SMPL	
CCV14 - 1	CCV14	18756	4.9894	-7470A	PPB	15 Nov 2018 10:58:36	U	SMPL	
CCB14 - 1	CCB14	-61	-0.0250	-7470A	PPB	15 Nov 2018 11:00:46	U	SMPL	