

245.1


LB99244

INSTRUMENT ID: CV2

Linear

 μ Abs.:

37214



A= 0.0000e+000

B= 2.6622e-004

C= 4.0536e-003

Rho= 0.9998342

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.016	0.016	43	0.000	43				
0.05	0.050	0.046	-0.004	158	0.0 %	158				
0.2	0.200	0.178	-0.022	655	0.0 %	655				
2.5	2.500	2.420	-0.080	9075	0.0 %	9075				
5.0	5.000	5.112	0.112	19187	0.0 %	19187				
7.5	7.500	7.567	0.067	28407	0.0 %	28407				
10.0	10.000	9.911	-0.089	37214	0.0 %	37214				

LB99244 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 15 Nov 2018 11:11:28

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	43	-	0.0000	245.1	PPB	15 Nov 2018 11:12:47	S	Std
0.05 - 1	50.05	158	-	0.0500	245.1	PPB	15 Nov 2018 11:20:21	S	Std
0.2 - 1	50.2	655	-	0.2000	245.1	PPB	15 Nov 2018 11:22:39	S	Std
2.5 - 1	52.5	9075	-	2.5000	245.1	PPB	15 Nov 2018 11:24:57	S	Std
5.0 - 1	55	19187	-	5.0000	245.1	PPB	15 Nov 2018 11:27:14	S	Std
7.5 - 1	57.5	28407	-	7.5000	245.1	PPB	15 Nov 2018 11:29:31	S	Std
10.0 - 1	57.8	37214	-	10.0000	245.1	PPB	15 Nov 2018 11:34:26	S	Std
ICV76 - 1	ICV76	14899	3.9705	-245.1		PPB	15 Nov 2018 11:39:44	U	SMPL
ICB76 - 1	ICB76	-51	-0.0095	-245.1		PPB	15 Nov 2018 11:42:00	U	SMPL
CCV15 - 1	CCV15	19084	5.0846	-245.1		PPB	15 Nov 2018 11:44:16	U	SMPL
CCB15 - 1	CCB15	-52	-0.0098	-245.1		PPB	15 Nov 2018 11:46:32	U	SMPL
CRA - 1	CRA	147	0.0432	-245.1		PPB	15 Nov 2018 11:48:49	U	SMPL
HighStd - 1	HighStd	37210	9.9102	-245.1		PPB	15 Nov 2018 11:51:05	U	SMPL
ChkStd - 1	ChkStd	26498	7.0584	-245.1		PPB	15 Nov 2018 11:53:21	U	SMPL
PB114825BS - 1	LCSW	14750	3.9308	-245.1		PPB	15 Nov 2018 12:00:35	U	SMPL
J5916-01 - 1	110918-SIDI-F-D-B	482	0.1324	-245.1		PPB	15 Nov 2018 12:02:53	U	SMPL
J5916-02 - 1	110918-SIDI-F-D-M	510	0.1398	-245.1		PPB	15 Nov 2018 12:05:10	U	SMPL
J5916-03 - 1	110918-SIDI-F-D-S	492	0.1350	-245.1		PPB	15 Nov 2018 12:07:27	U	SMPL
J5916-04 - 1	110918-SIDI-F-U-B	467	0.1284	-245.1		PPB	15 Nov 2018 12:09:45	U	SMPL
J5916-05 - 1	110918-SIDI-F-U-M	419	0.1156	-245.1		PPB	15 Nov 2018 12:12:01	U	SMPL
CCV16 - 1	CCV16	19337	5.1520	-245.1		PPB	15 Nov 2018 12:14:16	U	SMPL
CCB16 - 1	CCB16	-27	-0.0031	-245.1		PPB	15 Nov 2018 12:16:32	U	SMPL
J5916-06 - 1	110918-SIDI-F-U-S	378	0.1047	-245.1		PPB	15 Nov 2018 12:18:48	U	SMPL
J5916-07 - 1	110918-DBRI-F-D-B	375	0.1039	-245.1		PPB	15 Nov 2018 12:21:03	U	SMPL
J5916-08 - 1	110918-DBRI-F-D-M	346	0.0962	-245.1		PPB	15 Nov 2018 12:23:20	U	SMPL
J5916-09 - 1	110918-DBRI-F-D-S	264	0.0743	-245.1		PPB	15 Nov 2018 12:25:36	U	SMPL
J5916-10 - 1	110918-DBRI-F-U-B	46	0.0163	-245.1		PPB	15 Nov 2018 12:27:53	U	SMPL
J5916-11 - 1	110918-DBRI-F-U-M	419	0.1156	-245.1		PPB	15 Nov 2018 12:30:10	U	SMPL
J5916-12 - 1	110918-DBRI-F-U-S	399	0.1103	-245.1		PPB	15 Nov 2018 12:32:27	U	SMPL
J5916-13 - 1	110918-TIPC-F-D-B	415	0.1145	-245.1		PPB	15 Nov 2018 12:34:44	U	SMPL
J5916-14 - 1	110918-TIPC-F-D-S	419	0.1156	-245.1		PPB	15 Nov 2018 12:37:03	U	SMPL
J5916-15 - 1	110918-TIPC-F-U-B	385	0.1065	-245.1		PPB	15 Nov 2018 12:39:19	U	SMPL
CCV17 - 1	CCV17	19387	5.1653	-245.1		PPB	15 Nov 2018 12:41:35	U	SMPL
CCB17 - 1	CCB17	-26	-0.0029	-245.1		PPB	15 Nov 2018 12:43:52	U	SMPL
J5916-16 - 1	110918-TIPC-F-U-S	403	0.1113	-245.1		PPB	15 Nov 2018 12:46:09	U	SMPL
J5916-17 - 1	110918-JETT-F-D-B	417	0.1151	-245.1		PPB	15 Nov 2018 12:48:25	U	SMPL
J5916-18 - 1	110918-JETT-F-D-M	429	0.1183	-245.1		PPB	15 Nov 2018 12:50:42	U	SMPL
J5916-19 - 1	110918-JETT-F-D-S	412	0.1137	-245.1		PPB	15 Nov 2018 12:52:58	U	SMPL
J5916-20 - 1	110918-JETT-F-U-B	393	0.1087	-245.1		PPB	15 Nov 2018 12:55:15	U	SMPL
J5916-21 - 1	110918-JETT-F-U-M	428	0.1180	-245.1		PPB	15 Nov 2018 12:57:32	U	SMPL
J5916-22 - 1	110918-JETT-F-U-S	423	0.1167	-245.1		PPB	15 Nov 2018 12:59:49	U	SMPL
J5916-22DUP - 1	110918-JETT-F-U-SDUP	431	0.1188	-245.1		PPB	15 Nov 2018 13:02:07	U	SMPL
J5916-22MS - 1	110918-JETT-F-U-SMS	14129	3.7655	-245.1		PPB	15 Nov 2018 13:04:24	U	SMPL
J5916-22MSD - 1	110918-JETT-F-U-SMSD	14121	3.7634	-245.1		PPB	15 Nov 2018 13:06:41	U	SMPL
CCV18 - 1	CCV18	19713	5.2521	-245.1		PPB	15 Nov 2018 13:08:57	U	SMPL
CCB18 - 1	CCB18	-23	-0.0021	-245.1		PPB	15 Nov 2018 13:11:14	U	SMPL
J5916-22LX5 - 1		40	0.0147	-245.1		PPB	15 Nov 2018 13:13:30	U	SMPL
J5916-22A - 1		14886	3.9670	-245.1		PPB	15 Nov 2018 13:15:47	U	SMPL
PB114825BL - 1	PBW	-69	-0.0143	-245.1		PPB	15 Nov 2018 13:18:03	U	SMPL
CCV19 - 1	CCV19	18907	5.0375	-245.1		PPB	15 Nov 2018 13:23:21	U	SMPL
CCB19 - 1	CCB19	-46	-0.0082	-245.1		PPB	15 Nov 2018 13:25:38	U	SMPL