



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : JIGNESH

Run Number: LB99263

Slope : 99.0

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.1	4.01	11/15/2018	14:40
2	CAL2	1	Water			20.1	7.00	11/15/2018	14:42
3	CAL3	1	Water			20.2	10.04	11/15/2018	14:44
4	ICV	1	Water			20.3	7.01	11/15/2018	14:46
5	CCV1	1	Water			20.1	2.00	11/15/2018	14:48
6	J5954-01	1	Solid	20.02	20	16.2	6.36	11/15/2018	14:50
7	J5954-01DUP	1	Solid	20.04	20	16.3	6.37	11/15/2018	14:52
8	J5963-02	1	Solid	20.03	20	16.7	5.10	11/15/2018	14:54
9	J5964-03	1	Solid	20.02	20	16.7	7.53	11/15/2018	14:56
10	J5964-06	1	Solid	20.02	20	16.9	10.30	11/15/2018	14:58
11	CCV2	1	Water			20.0	12.00	11/15/2018	15:00

WORKLIST(Hardcopy Internal Chain)

LB99263

WorkList Name : CORROSIONITY111518

WorkList ID : 119302

Date : 11/15/2018 2:28:06 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/13/2018	Solid	J5954-01	Corrosivity	Cool 4 deg C	CLEA11		OWS-1	11/13/2018	9045D
11/15/2018	Solid	J5963-02	Corrosivity	Cool 4 deg C	PSEG01	L21	FS-OILY-STONE	11/15/2018	9045D
11/15/2018	Solid	J5964-03	Corrosivity	Cool 4 deg C	PSEG01	L21	LR-RBR-200036	11/15/2018	9045D
11/15/2018	Solid	J5964-06	Corrosivity	Cool 4 deg C	PSEG01	L21	LR-RBR-200029	11/15/2018	9045D

Date/Time 11/15/18 2:30

Received by: SP

Relinquished by: CP

Date/Time 11/15/18 3:05

Received by: CP

Relinquished by: SP