



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : sweta

Parameter: pH

Supervisor Review By : JIGNESH

Run Number: LB99272

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			19.5	4.01	11/16/2018	14:00
2	CAL2	1	Water			19.6	7.00	11/16/2018	14:02
3	CAL3	1	Water			19.8	10.04	11/16/2018	14:04
4	ICV	1	Water			19.5	7.01	11/16/2018	14:06
5	CCV1	1	Water			19.3	2.00	11/16/2018	14:08
6	J5978-01	1	Water			15.3	4.77	11/16/2018	14:10
7	J5978-01DUP	1	Water			15.4	4.78	11/16/2018	14:12
8	J5978-02	1	Water			15.5	4.50	11/16/2018	14:14
9	J5978-03	1	Water			15.7	4.63	11/16/2018	14:16
10	CCV2	1	Water			19.5	12.00	11/16/2018	14:18

WORKLIST(Hardcopy Internal Chain)

LB 99272

WorkList Name : pH111618-1

WorkList ID : 119351

Date : 11/16/2018 1:46:11 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/14/2018	Water	J5978-01 A	pH	Cool 4 deg C	STER13	L22	CELL-A-POOL-WATER	11/14/2018	SM 4500-H B
11/14/2018	Water	J5978-02 A	pH	Cool 4 deg C	STER13	L22	CELL-B-POOL-WATER	11/14/2018	SM 4500-H B
11/14/2018	Water	J5978-03 A	pH	Cool 4 deg C	STER13	L22	CELL-C-POOL-WATER	11/14/2018	SM 4500-H B

Date/Time 11/16/18 1:50
 Received by: SP
 Relinquished by: CP

Date/Time 11/16/18 2:30
 Received by: CP
 Relinquished by: SP