

CHMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

11/16/2018 15:05

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	96.846	0.0	0.084	
ICB1	0.129	0.0	0.004	
CCV1	243.110	0.0	0.206	
CCB1	0.148	0.0	0.004	
PB114886BS	97.387	0.0	0.085	
LOWPB114886	10.095	0.0	0.012	
HIGHPB114886	497.023	0.0	0.416	
J5762-05	1.409	0.0	0.005	
J5952-01	29.084	0.0	0.028	
J5952-01DUP	29.002	0.0	0.028	
J5952-02MS	68.758	0.0	0.061	
J5952-03MSD	70.251	0.0	0.062	
J5952-04	641.393	0.0	0.536	Test limit high
CCV2	242.069	0.0	0.205	
CCB2	0.097	0.0	0.004	
J5952-05	57.001	0.0	0.051	
J5952-06	72.204	0.0	0.064	
J5963-01	2.401	0.0	0.006	
J5964-04	3.966	0.0	0.007	
CCV3	248.050	0.0	0.210	
CCB3	-0.403	0.0	0.003	
PB114886BL	0.342	0.0	0.004	
J5964-01	0.678	0.0	0.004	
J5952-04DLX2	323.179	0.0	0.272	
CCV4	254.724	0.0	0.215	
CCB4	-0.366	0.0	0.003	

N	26
Mean	114.945
SD	167.6013
CV%	145.81

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

11/16/2018 9:47


Test Total CN

Accepted 11/16/2018 9:47

Factor 1205
Bias 0.004

Coeff. of det. 0.999980

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-0.4622	0.0000	
2	5.0PPBCN	0.007	3.9986	5.0000	
3	10PPBCN	0.012	9.6995	10.0000	
4	50PPBCN	0.046	51.0897	50.0000	
5	100PPBCN	0.087	100.2956	100.0000	
6	250PPBCN	0.212	251.0618	250.0000	
7	500PPBCN	0.418	499.3170	500.0000	

Aquakem v. 7.2AQ1

Results from time period:

Fri Nov 16 13:14:33 2018

Fri Nov 16 15:04:36 2018

149

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	-0.4622	µg/l	11/16/2018 9:14:05	
5.OPPBCN	A	Total CN	P	3.9986	µg/l	11/16/2018 9:14:06	
10PPBCN	A	Total CN	P	9.6995	µg/l	11/16/2018 9:14:07	
50PPBCN	A	Total CN	P	51.0897	µg/l	11/16/2018 9:14:08	
100PPBCN	A	Total CN	P	100.2956	µg/l	11/16/2018 9:14:09	
250PPBCN	A	Total CN	P	251.0618	µg/l	11/16/2018 9:14:10	
500PPBCN	A	Total CN	P	499.317	µg/l	11/16/2018 9:14:11	
ICV1	S	Total CN	P	96.8459	µg/l	11/16/2018 14:09:16	
ICB1	S	Total CN	P	0.1288	µg/l	11/16/2018 14:09:17	
CCV1	S	Total CN	P	243.1102	µg/l	11/16/2018 14:09:18	
CCB1	S	Total CN	P	0.1481	µg/l	11/16/2018 14:09:19	
PB114886BS	S	Total CN	P	97.3867	µg/l	11/16/2018 14:09:21	
LOWPB114886	S	Total CN	P	10.095	µg/l	11/16/2018 14:09:22	
HIGHPB114886	S	Total CN	P	497.0234	µg/l	11/16/2018 14:09:23	
J5762-05	S	Total CN	P	1.4089	µg/l	11/16/2018 14:09:24	
J5952-01	S	Total CN	P	29.084	µg/l	11/16/2018 14:09:25	
J5952-01DUP	S	Total CN	P	29.002	µg/l	11/16/2018 14:09:26	
J5952-02MS	S	Total CN	P	68.7576	µg/l	11/16/2018 14:16:51	
J5952-03MSD	S	Total CN	P	70.2515	µg/l	11/16/2018 14:16:52	
J5952-04	S	Total CN	P	641.3929	µg/l	11/16/2018 14:16:53	
CCV2	S	Total CN	P	242.0692	µg/l	11/16/2018 14:16:54	
CCB2	S	Total CN	P	0.0968	µg/l	11/16/2018 14:16:55	
J5952-05	S	Total CN	P	57.0014	µg/l	11/16/2018 14:16:56	
J5952-06	S	Total CN	P	72.204	µg/l	11/16/2018 14:16:57	
J5963-01	S	Total CN	P	2.4009	µg/l	11/16/2018 14:16:58	
J5964-04	S	Total CN	P	3.9664	µg/l	11/16/2018 14:17:00	
CCV3	S	Total CN	P	248.0497	µg/l	11/16/2018 14:17:01	
CCB3	S	Total CN	P	-0.4035	µg/l	11/16/2018 14:21:38	
PB114886BL	S	Total CN	P	0.3417	µg/l	11/16/2018 14:42:31	
J5964-01	S	Total CN	P	0.6777	µg/l	11/16/2018 14:42:33	
J5952-04DLX2	S	Total CN	P	323.1794	µg/l	11/16/2018 15:04:34	
CCV4	S	Total CN	P	254.7241	µg/l	11/16/2018 15:04:35	
CCB4	S	Total CN	P	-0.3664	µg/l	11/16/2018 15:04:36	