


7471B

LB99281
INSTRUMENT ID: CV1

 μ Abs.:

5255E



Linear

A= 0.0000e+000

B= 1.8992e-004

C= -3.0165e-002

Rho= 0.9999620

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.011	-0.011	102	0.000	102				
0.2	0.200	0.191	-0.009	1163	0.0 %	1163				
2.5	2.500	2.485	-0.015	13244	0.0 %	13244				
5.0	5.000	5.043	0.043	26714	0.0 %	26714				
7.5	7.500	7.540	0.040	39858	0.0 %	39858				
10.0	10.000	9.952	-0.048	52558	0.0 %	52558				

LB99281 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 16 Nov 2018 17:29:46

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	102	-	0.0000	7471B	PPB	16 Nov 2018 17:31:50	S	Std
0.2 - 1	51.2	1163	-	0.2000	7471B	PPB	16 Nov 2018 17:34:08	S	Std
2.5 - 1	52.5	13244	-	2.5000	7471B	PPB	16 Nov 2018 17:36:26	S	Std
5.0 - 1	55	26714	-	5.0000	7471B	PPB	16 Nov 2018 17:38:44	S	Std
7.5 - 1	57.5	39858	-	7.5000	7471B	PPB	16 Nov 2018 17:41:01	S	Std
10.0 - 1	510	52558	-	10.0000	7471B	PPB	16 Nov 2018 17:43:19	S	Std
ICV80 - 1	ICV80	20603	3.8828		7471B	PPB	16 Nov 2018 17:46:05	U	SMPL
ICB80 - 1	ICB80	-25	-0.0349		7471B	PPB	16 Nov 2018 17:48:21	U	SMPL
CCV38 - 1	CCV38	26995	5.0968		7471B	PPB	16 Nov 2018 17:50:38	U	SMPL
CCB38 - 1	CCB38	-59	-0.0414		7471B	PPB	16 Nov 2018 17:52:53	U	SMPL
CRA - 1	CRA	1142	0.1867		7471B	PPB	16 Nov 2018 17:55:10	U	SMPL
HighStd - 1	HighStd	53048	10.0448		7471B	PPB	16 Nov 2018 17:57:26	U	SMPL
ChkStd - 1	ChkStd	37961	7.1795		7471B	PPB	16 Nov 2018 17:59:43	U	SMPL
PB114888BL - 1	PBS	52	-0.0203		7471B	PPB	16 Nov 2018 18:04:35	U	SMPL
PB114888BS - 1	LCSS	19062	3.5901		7471B	PPB	16 Nov 2018 18:06:52	U	SMPL
J5762-05 - 1	BU-03-111518	1873	0.3256		7471B	PPB	16 Nov 2018 18:09:08	U	SMPL
J5807-12 - 1	PT-MET-SOIL	800172	HIGH		7471B	PPB	16 Nov 2018 18:11:26	X	
J5952-01 - 1	VTW-01(7-10)	39292	7.4323		7471B	PPB	16 Nov 2018 18:16:37	U	SMPL
J5952-01DUP - 1	VTW-01(7-10)DUP	40519	7.6653		7471B	PPB	16 Nov 2018 18:21:44	U	SMPL
J5952-02 - 1	J5952-01MS	56918	10.7798		7471B	PPB	16 Nov 2018 18:23:59	U	SMPL
CCV39 - 1	CCV39	26782	5.0563		7471B	PPB	16 Nov 2018 18:26:17	U	SMPL
CCB39 - 1	CCB39	-94	-0.0480		7471B	PPB	16 Nov 2018 18:28:35	U	SMPL
J5952-03 - 1	J5952-01MSD	55789	10.5654		7471B	PPB	16 Nov 2018 18:30:53	U	SMPL
J5952-04 - 1	VTW-02(4-6)	35906	6.7892		7471B	PPB	16 Nov 2018 18:33:08	U	SMPL
J5952-05 - 1	VTW-03(4-6)	17313	3.2580		7471B	PPB	16 Nov 2018 18:35:27	U	SMPL
J5952-06 - 1	DUP-01	35372	6.6878		7471B	PPB	16 Nov 2018 18:37:45	U	SMPL
J5963-01 - 1	FS-OILY-STONE	2430	0.4313		7471B	PPB	16 Nov 2018 18:40:03	U	SMPL
J5964-01 - 1	LR-RBR-200036	734	0.1092		7471B	PPB	16 Nov 2018 18:42:22	U	SMPL
J5964-04 - 1	LR-RBR-200029	2998	0.5392		7471B	PPB	16 Nov 2018 18:44:40	U	SMPL
J5952-01LX5 - 1		7833	1.4575		7471B	PPB	16 Nov 2018 18:46:57	U	SMPL
J5952-01A - 1		62708	11.8795		7471B	PPB	16 Nov 2018 18:49:15	U	SMPL
CCV40 - 1	CCV40	25520	4.8166		7471B	PPB	16 Nov 2018 18:58:13	U	SMPL
CCB40 - 1	CCB40	0	-0.0302		7471B	PPB	16 Nov 2018 19:03:47	U	SMPL
J5952-01DLX2 - 1		19622	3.6965		7471B	PPB	16 Nov 2018 19:06:02	U	SMPL
J5952-01DUPDLX2 - 1		20221	3.8102		7471B	PPB	16 Nov 2018 19:10:46	U	SMPL
J5952-02DLX2 - 1		28475	5.3779		7471B	PPB	16 Nov 2018 19:13:03	U	SMPL
J5952-03DLX2 - 1		27080	5.1129		7471B	PPB	16 Nov 2018 19:22:06	U	SMPL
J5952-01LDLX10 - 1		4000	0.7295		7471B	PPB	16 Nov 2018 19:24:25	U	SMPL
J5952-01ADLX2 - 1		31861	6.0209		7471B	PPB	16 Nov 2018 19:26:44	U	SMPL
J5807-12DLX10 - 1	PT-MET-SOIL	175574	33.3152		7471B	PPB	16 Nov 2018 19:36:05	U	SMPL
CCV41 - 1	CCV41	26183	4.9426		7471B	PPB	16 Nov 2018 19:38:33	U	SMPL
CCB41 - 1	CCB41	-100	-0.0492		7471B	PPB	16 Nov 2018 19:40:57	U	SMPL
J5807-12DL2X50 - 1	PT-MET-SOIL	36873	6.9728		7471B	PPB	16 Nov 2018 19:43:15	U	SMPL
CCV42 - 1	CCV42	26776	5.0552		7471B	PPB	16 Nov 2018 19:50:11	U	SMPL
CCB42 - 1	CCB42	-47	-0.0391		7471B	PPB	16 Nov 2018 19:52:30	U	SMPL