



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: Corrosivity

Supervisor Review By : JIGNESH

Run Number: LB99289

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			17.2	4.01	11/16/2018	16:40
2	CAL2	1	Water			17.4	7.00	11/16/2018	16:42
3	CAL3	1	Water			17.1	10.04	11/16/2018	16:44
4	ICV	1	Water			17.3	7.01	11/16/2018	16:46
5	CCV1	1	Water			17.2	2.00	11/16/2018	16:48
6	J5975-04	1	Water			15.6	7.30	11/16/2018	16:50
7	J5975-04DUP	1	Water			15.7	7.31	11/16/2018	16:52
8	J5976-02	1	Water			14.8	7.33	11/16/2018	16:54
9	J5980-02	1	Water			11.2	6.53	11/16/2018	16:56
10	CCV2	1	Water			17.2	12.00	11/16/2018	16:58

WORKLIST(Hardcopy Internal Chain)

LB 09289

WorkList Name : CORROSIONITY111618-1

WorkList ID : 119331

Date : 11/16/2018 10:08:07 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/15/2018	Water	J5975-04	Corrosivity	Cool 4 deg C	PSEG01	J31	VB44949	11/15/2018	9040C
11/15/2018	Water	J5976-02	Corrosivity	Cool 4 deg C	PSEG01	L31	OILY-WATER	11/15/2018	9040C
11/16/2018	Water	J5980-02	Corrosivity	Cool 4 deg C	PSEG01	L21	TS	11/16/2018	9040C

Date/Time 11/16/18 4:20
 Received by: SP
 Relinquished by: CP

Date/Time 11/16/18 5:20
 Received by: CP
 Relinquished by: SP