



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: sweta

Run Number: LB99301

Ambient Barometric Pressure (mmHg): 745

Barometric Scale ID: 55971

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	81.5	11/19/2018	10:05
2	J5975-04		1	8	>100.0	>212.0	>212.0	11/19/2018	10:40
3	J5975-04DUP		1	8	>100.0	>212.0	>212.0	11/19/2018	11:02
4	J5976-02		1	9	>100.0	>212.0	>212.0	11/19/2018	11:35
5	J5980-02		1	7	>100.0	>212.0	>212.0	11/19/2018	11:55

$$\text{Result} = (\text{Celsius} * 1.8) + 32$$

$$\text{Final Result} = \text{Result} + (760 - \text{Ambient Barometric Pressure}) * 0.06$$

LB 99301

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp111918 WorkList ID : 119389 Date : 11/19/2018 9:10:28 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/20/2018	Water	J5975-04 B	Flash Point	Cool 4 deg C	PSEG01	J31	VB44949	11/15/2018	1010A
11/20/2018	Water	J5976-02 B	Flash Point	Cool 4 deg C	PSEG01	L31	OILY-WATER	11/15/2018	1010A
11/21/2018	Water	J5980-02 B	Flash Point	Cool 4 deg C	PSEG01	L21	TS	11/16/2018	1010A

Date/Time 11/19/18 9:30AM
 Received by: KS
 Relinquished by: CP

Date/Time 11/19/18 12:05PM
 Received by: CP
 Relinquished by: KS