



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB99305

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP68911
COD calibration std. 100 ppm	WP68908
COD calibration std. 10 ppm	WP68907
COD calibration std. 150 ppm	WP68910
COD calibration std. 50 ppm	WP68909
COD calibration std. 0 ppm	WP68906
COD Digestion Vials Low Level 0-150Mg/L	W2421
COD ICV-LCS std, 50ppm	WP69668
COD CCV std, 50ppm	WP69669

Temp In (C): 150	Date In: 11/19/2018	Time In: 09:30
Temp Out (C): 150	Date Out: 11/19/2018	Time Out: 11:30

Intercept: -0.7935

Slope: 1.0096

Regression: 1

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	10/20/2018	13:05
2	CAL2	10	1	Water	9.000	10/20/2018	13:05
3	CAL3	50	1	Water	49.000	10/20/2018	13:06
4	CAL4	100	1	Water	100.000	10/20/2018	13:06
5	CAL5	150	1	Water	151.000	10/20/2018	13:07



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB99305

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	49.000	49.320	10/20/2018	13:07
2	ICB				1	Water	0.000	0.786	10/20/2018	13:08
3	CCV1	50			1	Water	50.000	50.311	11/19/2018	12:30
4	CCB1				1	Water	0.000	0.786	11/19/2018	12:30
5	LB99305BL				1	Water	0.000	0.786	11/19/2018	12:31
6	LB99305BS	50			1	Water	49.000	49.320	11/19/2018	12:31
7	J5975-03				10	Water	82.000	82.006	11/19/2018	12:32
8	J5976-01				1	Water	27.000	27.529	11/19/2018	12:32
9	J5980-01				1	Water	5.000	5.738	11/19/2018	12:33
10	J5980-01DUP				1	Water	4.000	4.748	11/19/2018	12:33
11	J5980-01MS	50			1	Water	55.000	55.263	11/19/2018	12:34
12	J5980-01MSD	50			1	Water	53.000	53.282	11/19/2018	12:34
13	CCV2	50			1	Water	50.000	50.311	11/19/2018	12:35
14	CCB2				1	Water	0.000	0.786	11/19/2018	12:35