

7471B


LB 99306

INSTRUMENT ID: CV1

Linear

 μ Abs.:

54176



A= 0.0000e+000

B= 1.8341e-004

C= -3.0252e-002

Rho= 0.9998649

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.022	-0.022	47	0.000	47				
0.2	0.200	0.186	-0.014	1181	0.0 %	1181				
2.5	2.500	2.474	-0.026	13656	0.0 %	13656				
5.0	5.000	5.071	0.071	27813	0.0 %	27813				
7.5	7.500	7.583	0.083	41511	0.0 %	41511				
10.0	10.000	9.906	-0.094	54176	0.0 %	54176				

LB99306 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 19 Nov 2018 16:09:36

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc/Method	Units	Date	Type	Type
0.0 - 1	50	47	-	0.00007471B	PPB	19 Nov 2018 16:11:20	S	Std
0.2 - 1	50.2	1181	-	0.20007471B	PPB	19 Nov 2018 16:13:38	S	Std
2.5 - 1	52.5	13656	-	2.50007471B	PPB	19 Nov 2018 16:15:56	S	Std
5.0 - 1	55	27813	-	5.00007471B	PPB	19 Nov 2018 16:18:14	S	Std
7.5 - 1	57.5	41511	-	7.50007471B	PPB	19 Nov 2018 16:20:31	S	Std
10.0 - 1	510	54176	-	10.00007471B	PPB	19 Nov 2018 16:22:48	S	Std
ICV81 - 1	ICV81	21850	3.9773	-7471B	PPB	19 Nov 2018 16:25:38	U	SMPL
ICB81 - 1	ICB81	-89	-0.0466	-7471B	PPB	19 Nov 2018 16:27:54	U	SMPL
CCV47 - 1	CCV47	27789	5.0666	-7471B	PPB	19 Nov 2018 16:30:11	U	SMPL
CCB47 - 1	CCB47	-77	-0.0444	-7471B	PPB	19 Nov 2018 16:32:27	U	SMPL
CRA - 1	CRA	1188	0.1876	-7471B	PPB	19 Nov 2018 16:34:45	U	SMPL
HighStd - 1	HighStd	53795	9.8365	-7471B	PPB	19 Nov 2018 16:37:01	U	SMPL
ChkStd - 1	ChkStd	38562	7.0425	-7471B	PPB	19 Nov 2018 16:39:17	U	SMPL
PB114913BL - 1	PBS	-123	-0.0528	-7471B	PPB	19 Nov 2018 16:41:35	U	SMPL
PB114913BS - 1	LCSS	20568	3.7422	-7471B	PPB	19 Nov 2018 16:46:46	U	SMPL
J5774-01 - 1	SU-04-111618	1922	0.3223	-7471B	PPB	19 Nov 2018 16:49:02	U	SMPL
J5970-01 - 1	FK-GF-02-028-A	4429	0.7821	-7471B	PPB	19 Nov 2018 16:51:21	U	SMPL
J5970-03 - 1	FK-GF-02-028-B	3386	0.5908	-7471B	PPB	19 Nov 2018 16:53:39	U	SMPL
J5975-01 - 1	VB44949	14685	2.6632	-7471B	PPB	19 Nov 2018 16:55:56	U	SMPL
J5977-08 - 1	MMTW-03(8-10)	97636	17.8775	-7471B	PPB	19 Nov 2018 16:58:12	U	SMPL
CCV48 - 1	CCV48	27510	5.0155	-7471B	PPB	19 Nov 2018 17:00:29	U	SMPL
CCB48 - 1	CCB48	-130	-0.0541	-7471B	PPB	19 Nov 2018 17:02:48	U	SMPL
J5977-08DUP - 1	MMTW-03(8-10)DUP	98765	18.0846	-7471B	PPB	19 Nov 2018 17:07:37	U	SMPL
J5977-09 - 1	J5977-08MS	120124	22.0021	-7471B	PPB	19 Nov 2018 17:12:31	U	SMPL
J5977-10 - 1	J5977-08MSD	114779	21.0218	-7471B	PPB	19 Nov 2018 17:14:50	U	SMPL
J5977-11 - 1	MMTW-02(10-12)	28440	5.1860	-7471B	PPB	19 Nov 2018 17:17:10	U	SMPL
J5977-12 - 1	MMTW-01(3-5)	50445	9.2220	-7471B	PPB	19 Nov 2018 17:19:29	U	SMPL
J5977-13 - 1	DUP-01	25930	4.7257	-7471B	PPB	19 Nov 2018 17:21:48	U	SMPL
J5981-01 - 1	DTW-01(6-10)	19657	3.5751	-7471B	PPB	19 Nov 2018 17:24:07	U	SMPL
J5981-02 - 1	DTW-02(11-13)	27798	5.0683	-7471B	PPB	19 Nov 2018 17:26:26	U	SMPL
J5981-03 - 1	J5981-02MS	49101	8.9755	-7471B	PPB	19 Nov 2018 17:33:00	U	SMPL
J5981-04 - 1	J5981-02MSD	49060	8.9680	-7471B	PPB	19 Nov 2018 17:35:18	U	SMPL
CCV49 - 1	CCV49	28025	5.1099	-7471B	PPB	19 Nov 2018 17:37:36	U	SMPL
CCB49 - 1	CCB49	-127	-0.0535	-7471B	PPB	19 Nov 2018 17:39:54	U	SMPL
J5981-05 - 1	DTW-03(8-10)	126695	23.2073	-7471B	PPB	19 Nov 2018 17:42:11	U	SMPL
J5981-06 - 1	DUP-01	23989	4.3697	-7471B	PPB	19 Nov 2018 17:44:28	U	SMPL
J5989-01 - 1	ES-RO-343	10105	1.8231	-7471B	PPB	19 Nov 2018 17:46:48	U	SMPL
J5991-01 - 1	WS-RO-358	11093	2.0044	-7471B	PPB	19 Nov 2018 17:49:06	U	SMPL
J5977-08LX5 - 1		19961	3.6309	-7471B	PPB	19 Nov 2018 17:51:22	U	SMPL
J5977-08A - 1		116358	21.3114	-7471B	PPB	19 Nov 2018 17:53:39	U	SMPL
CCV50 - 1	CCV50	28046	5.1138	-7471B	PPB	19 Nov 2018 17:55:56	U	SMPL
CCB50 - 1	CCB50	-136	-0.0552	-7471B	PPB	19 Nov 2018 17:58:17	U	SMPL
J5977-08DLX5 - 1		19220	3.4950	-7471B	PPB	19 Nov 2018 18:00:36	U	SMPL
J5977-08DUPDLX5 - 1		19120	3.4766	-7471B	PPB	19 Nov 2018 18:02:52	U	SMPL
J5977-09DLX5 - 1		22935	4.1763	-7471B	PPB	19 Nov 2018 18:05:09	U	SMPL
J5977-10DLX5 - 1		22747	4.1419	-7471B	PPB	19 Nov 2018 18:07:26	U	SMPL
J5977-08LDLX25 - 1		3817	0.6698	-7471B	PPB	19 Nov 2018 18:09:44	U	SMPL
J5977-08ADLX5 - 1		24548	4.4722	-7471B	PPB	19 Nov 2018 18:12:02	U	SMPL
J5981-05DLX5 - 1		26324	4.7979	-7471B	PPB	19 Nov 2018 18:14:18	U	SMPL
CCV51 - 1	CCV51	27710	5.0521	-7471B	PPB	19 Nov 2018 18:16:36	U	SMPL
CCB51 - 1	CCB51	-103	-0.0491	-7471B	PPB	19 Nov 2018 18:18:54	U	SMPL