



PERCENT SOLID

Supervisor: apatel  
Analyst: JIGNESH  
Date: 12/11/2018

OVENTEMP IN Celsius(°C): 107  
Time IN: 17:35  
In Date: 11/19/2018  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M Oven-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:20  
Out Date: 11/20/2018  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M Sc-1

QC:LB99312

| Lab ID   | Client Sample ID   | Dish# | Dish Wt(g)<br>(A) | Dish +<br>Sample Wt(g)<br>(B) | Dish + Dry<br>Sample Wt(g)<br>(C) | % Solid |
|----------|--------------------|-------|-------------------|-------------------------------|-----------------------------------|---------|
| J5763-01 | HR-06-111918-A     | 1     | 1.13              | 9.73                          | 8.77                              | 88.8    |
| J5763-03 | HR-06-111918-B     | 2     | 1.19              | 9.79                          | 8.71                              | 87.4    |
| J5775-01 | TR-05-111918       | 3     | 1.12              | 9.9                           | 9.03                              | 90.1    |
| J5996-01 | AOC2-01R           | 4     | 1.12              | 9.55                          | 8.65                              | 89.3    |
| J5996-03 | NON-UST-01R-A      | 5     | 1.12              | 9.58                          | 8.61                              | 88.5    |
| J5996-05 | AOC8-01R-A         | 6     | 1.12              | 9.58                          | 7.78                              | 78.7    |
| J5996-07 | AOC8-01R-C         | 7     | 1.13              | 9.97                          | 8.26                              | 80.7    |
| J5996-08 | AOC8-04A           | 8     | 1.12              | 9.6                           | 7.87                              | 79.6    |
| J5996-10 | AOC8-05A           | 9     | 1.12              | 9.9                           | 9.00                              | 89.7    |
| J5996-12 | AOC1-05-A          | 10    | 1.12              | 9.72                          | 8.73                              | 88.5    |
| J5996-14 | AOC1-05-B          | 11    | 1.12              | 9.72                          | 8.73                              | 88.5    |
| J5996-15 | AOC1-06-A          | 12    | 1.12              | 9.72                          | 9.01                              | 91.7    |
| J5996-17 | AOC1-06-B          | 13    | 1.12              | 9.96                          | 8.98                              | 88.9    |
| J5996-18 | NON-UST-06-A       | 14    | 1.12              | 9.6                           | 8.68                              | 89.2    |
| J5996-25 | AOC1-06-A          | 15    | 1.12              | 9.72                          | 9.01                              | 91.7    |
| J5999-01 | SSSI-1114-S-TCLP-2 | 16    | 1.12              | 9.72                          | 3.96                              | 33.0    |
| J5999-02 | SSSI-1114-S-TCLP-3 | 17    | 1.12              | 9.96                          | 1.7                               | 6.6     |
| J5999-03 | SSSI-1114-S-TCLP-4 | 18    | 1.1               | 9.73                          | 2.39                              | 14.9    |
| J5999-04 | SSSI-1115-S-AB-1   | 19    | 1.14              | 9.69                          | 7.54                              | 74.9    |
| J5999-05 | SSSI-1115-S-AB-1   | 20    | 1.11              | 9.87                          | 8.11                              | 79.9    |
| J5999-06 | SSSI-1115-S-AB-2   | 21    | 1.13              | 9.81                          | 7.68                              | 75.5    |
| J5999-07 | SSSI-1115-S-AB-2   | 22    | 1.12              | 9.88                          | 7.61                              | 74.1    |
| J5999-08 | SSSI-1116-S-AB-4   | 23    | 1.12              | 9.97                          | 7.62                              | 73.4    |
| J5999-09 | SSSI-1116-S-AB-4   | 24    | 1.12              | 9.66                          | 7.92                              | 79.6    |
| J5999-10 | SSSI-1116-S-AB-5   | 25    | 1.12              | 9.52                          | 8.28                              | 85.2    |
| J5999-11 | SSSI-1116-S-AB-5   | 26    | 1.14              | 9.69                          | 8.66                              | 88.0    |
| J6000-01 | FS-12072-11929     | 27    | 1.12              | 9.73                          | 8.61                              | 87.0    |
| J6003-03 | NON-UST-03R-A      | 28    | 1.15              | 9.92                          | 9.02                              | 89.7    |
| J6003-04 | J6003-03MS         | 29    | 1.15              | 9.92                          | 9.02                              | 89.7    |
| J6003-05 | J6003-03MSD        | 30    | 1.15              | 9.92                          | 9.02                              | 89.7    |
| J6003-07 | NON-UST-03R-C      | 31    | 1.12              | 9.96                          | 8.98                              | 88.9    |
| J6003-08 | AOC2-03A           | 32    | 1.13              | 9.98                          | 9.07                              | 89.7    |
| J6003-09 | AOC2-03B           | 33    | 1.13              | 9.98                          | 9.07                              | 89.7    |
| J6003-10 | AOC2-04            | 34    | 1.00              | 2.00                          | 2.00                              | 100.0   |
| J6003-11 | AOC1-03R           | 35    | 1.00              | 2.00                          | 2.00                              | 100.0   |
| J6003-12 | AOC2-05            | 36    | 1.14              | 9.74                          | 8.76                              | 88.6    |
| J6003-13 | NON-UST-02R        | 37    | 1.13              | 9.98                          | 9.21                              | 91.3    |
| J6003-14 | NON-UST-04         | 38    | 1.12              | 9.64                          | 8.66                              | 88.5    |
| J6003-15 | NON-UST-05A        | 39    | 1.12              | 9.91                          | 8.8                               | 87.4    |



PERCENT SOLID

Supervisor: apatel  
Analyst: JIGNESH  
Date: 12/11/2018

OVENTEMP IN Celsius(°C): 107  
Time IN: 17:35  
In Date: 11/19/2018  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M Oven-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:20  
Out Date: 11/20/2018  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M Sc-1

QC:LB99312

| Lab ID   | Client Sample ID     | Dish# | Dish Wt(g)<br>(A) | Dish +<br>Sample Wt(g)<br>(B) | Dish + Dry<br>Sample Wt(g)<br>(C) | % Solid |
|----------|----------------------|-------|-------------------|-------------------------------|-----------------------------------|---------|
| J6003-19 | AOC2-04              | 40    | 1.13              | 9.55                          | 8.66                              | 89.4    |
| J6003-20 | AOC2-03A             | 41    | 1.00              | 2.00                          | 2.00                              | 100.0   |
| J6006-01 | RS-CON-WALL-1        | 42    | 1.00              | 2.00                          | 2.00                              | 100.0   |
| J6006-03 | RS-CON-WALL-2        | 43    | 1.00              | 2.00                          | 2.00                              | 100.0   |
| J6006-05 | RS-CON-FLOOR         | 44    | 1.00              | 2.00                          | 2.00                              | 100.0   |
| J6010-01 | FS-RT3786            | 45    | 1.13              | 9.94                          | 8.66                              | 85.5    |
| J6010-02 | FS-RT3786            | 46    | 1.13              | 9.94                          | 8.69                              | 85.8    |
| J6013-01 | DISPOSAL-WASTE-CLASS | 47    | 1.13              | 9.96                          | 8.69                              | 85.6    |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

# WORKLIST(Hardcopy Internal Chain)

LB99312

WorkList Name : %1-111918

WorkList ID : 119362

Date : 11/19/2018 7:58:39 AM

| Due Date   | Matrix | Sample   | Test           | Preservative | Customer | Storage Location | Customer Sample    | Collect Date | Method       |
|------------|--------|----------|----------------|--------------|----------|------------------|--------------------|--------------|--------------|
| 11/22/2018 | Solid  | J5763-01 | Percent Solids | Cool 4 deg C | PSEG05   | L31              | HR-06-111918-A     | 11/19/2018   | Chemtech -SO |
| 11/22/2018 | Solid  | J5763-03 | Percent Solids | Cool 4 deg C | PSEG05   | L31              | HR-06-111918-B     | 11/19/2018   | Chemtech -SO |
| 11/22/2018 | Solid  | J5775-01 | Percent Solids | Cool 4 deg C | PSEG05   | L31              | TR-05-111918       | 11/19/2018   | Chemtech -SO |
| 11/27/2018 | Solid  | J5996-01 | Percent Solids | Cool 4 deg C | BEMS01   | L53              | AOC2-01R           | 11/17/2018   | Chemtech -SO |
| 11/27/2018 | Solid  | J5996-03 | Percent Solids | Cool 4 deg C | BEMS01   | L53              | NON-UST-01R-A      | 11/17/2018   | Chemtech -SO |
| 11/26/2018 | Solid  | J5996-05 | Percent Solids | Cool 4 deg C | BEMS01   | L53              | AOC8-01R-A         | 11/17/2018   | Chemtech -SO |
| 11/26/2018 | Solid  | J5996-07 | Percent Solids | Cool 4 deg C | BEMS01   | L53              | AOC8-01R-C         | 11/17/2018   | Chemtech -SO |
| 11/26/2018 | Solid  | J5996-08 | Percent Solids | Cool 4 deg C | BEMS01   | L53              | AOC8-04A           | 11/17/2018   | Chemtech -SO |
| 11/26/2018 | Solid  | J5996-10 | Percent Solids | Cool 4 deg C | BEMS01   | L53              | AOC8-05A           | 11/17/2018   | Chemtech -SO |
| 11/27/2018 | Solid  | J5996-12 | Percent Solids | Cool 4 deg C | BEMS01   | L53              | AOC1-05R-A         | 11/17/2018   | Chemtech -SO |
| 11/27/2018 | Solid  | J5996-14 | Percent Solids | Cool 4 deg C | BEMS01   | L53              | AOC1-05R-B         | 11/17/2018   | Chemtech -SO |
| 11/27/2018 | Solid  | J5996-15 | Percent Solids | Cool 4 deg C | BEMS01   | L53              | AOC1-06R-A         | 11/17/2018   | Chemtech -SO |
| 11/27/2018 | Solid  | J5996-17 | Percent Solids | Cool 4 deg C | BEMS01   | L53              | AOC1-06R-B         | 11/17/2018   | Chemtech -SO |
| 11/27/2018 | Solid  | J5996-18 | Percent Solids | Cool 4 deg C | BEMS01   | L53              | NON-UST-06-A       | 11/17/2018   | Chemtech -SO |
| 11/27/2018 | Solid  | J5996-25 | Percent Solids | Cool 4 deg C | BEMS01   | L53              | AOC1-06R-A         | 11/17/2018   | Chemtech -SO |
| 11/24/2018 | Solid  | J5999-01 | Percent Solids | Cool 4 deg C | USAR03   | L51              | SSSI-1114-S-TCLP-2 | 11/14/2018   | Chemtech -SO |
| 11/24/2018 | Solid  | J5999-02 | Percent Solids | Cool 4 deg C | USAR03   | L51              | SSSI-1114-S-TCLP-3 | 11/14/2018   | Chemtech -SO |
| 11/24/2018 | Solid  | J5999-03 | Percent Solids | Cool 4 deg C | USAR03   | L51              | SSSI-1114-S-TCLP-4 | 11/14/2018   | Chemtech -SO |
| 11/25/2018 | Solid  | J5999-04 | Percent Solids | Cool 4 deg C | USAR03   | L51              | SSSI-1115-S-AB-1   | 11/15/2018   | Chemtech -SO |
| 11/25/2018 | Solid  | J5999-05 | Percent Solids | Cool 4 deg C | USAR03   | L51              | SSSI-1115-S-AB-1   | 11/15/2018   | Chemtech -SO |
| 11/25/2018 | Solid  | J5999-06 | Percent Solids | Cool 4 deg C | USAR03   | L51              | SSSI-1115-S-AB-2   | 11/15/2018   | Chemtech -SO |

Date/Time 11/19/18 4:00 PM Date/Time 11-19-18 5:40 PM  
 Received by: JI Received by: 68  
 Relinquished by: 68 Relinquished by: JD

LB 44312

# WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-111918

WorkList ID : 119362

Date : 11/19/2018 7:58:39 AM

| Due Date   | Matrix | Sample   | Test           | Preservative | Customer | Storage Location | Customer Sample  | Collect Date | Method       |
|------------|--------|----------|----------------|--------------|----------|------------------|------------------|--------------|--------------|
| 11/25/2018 | Solid  | J5999-07 | Percent Solids | Cool 4 deg C | USAR03   | L51              | SSSI-1115-S-AB-2 | 11/15/2018   | Chemtech -SO |
| 11/26/2018 | Solid  | J5999-08 | Percent Solids | Cool 4 deg C | USAR03   | L51              | SSSI-1116-S-AB-4 | 11/16/2018   | Chemtech -SO |
| 11/26/2018 | Solid  | J5999-09 | Percent Solids | Cool 4 deg C | USAR03   | L51              | SSSI-1116-S-AB-4 | 11/16/2018   | Chemtech -SO |
| 11/26/2018 | Solid  | J5999-10 | Percent Solids | Cool 4 deg C | USAR03   | L51              | SSSI-1116-S-AB-5 | 11/16/2018   | Chemtech -SO |
| 11/26/2018 | Solid  | J5999-11 | Percent Solids | Cool 4 deg C | USAR03   | L51              | SSSI-1116-S-AB-5 | 11/16/2018   | Chemtech -SO |
| 11/26/2018 | Solid  | J6000-01 | Percent Solids | Cool 4 deg C | PSEG01   | L31              | FS-12072-11929   | 11/19/2018   | Chemtech -SO |
| 11/26/2018 | Solid  | J6003-03 | Percent Solids | Cool 4 deg C | BEMS01   | L61              | NON-UST-03R-A    | 11/17/2018   | Chemtech -SO |
| 11/27/2018 | Solid  | J6003-04 | Percent Solids | Cool 4 deg C | BEMS01   | L61              | J6003-03MS       | 11/17/2018   | Chemtech -SO |
| 11/27/2018 | Solid  | J6003-05 | Percent Solids | Cool 4 deg C | BEMS01   | L61              | J6003-03MSD      | 11/17/2018   | Chemtech -SO |
| 11/26/2018 | Solid  | J6003-07 | Percent Solids | Cool 4 deg C | BEMS01   | L61              | NON-UST-03R-C    | 11/17/2018   | Chemtech -SO |
| 11/27/2018 | Solid  | J6003-08 | Percent Solids | Cool 4 deg C | BEMS01   | L61              | AOC2-03A         | 11/17/2018   | Chemtech -SO |
| 11/27/2018 | Solid  | J6003-09 | Percent Solids | Cool 4 deg C | BEMS01   | L61              | AOC2-03B         | 11/17/2018   | Chemtech -SO |
| 11/27/2018 | Solid  | J6003-10 | Percent Solids | Cool 4 deg C | BEMS01   | L61              | AOC2-04          | 11/17/2018   | Chemtech -SO |
| 11/27/2018 | Solid  | J6003-11 | Percent Solids | Cool 4 deg C | BEMS01   | L61              | AOC1-03R         | 11/17/2018   | Chemtech -SO |
| 11/27/2018 | Solid  | J6003-12 | Percent Solids | Cool 4 deg C | BEMS01   | L61              | AOC2-05          | 11/17/2018   | Chemtech -SO |
| 11/27/2018 | Solid  | J6003-13 | Percent Solids | Cool 4 deg C | BEMS01   | L61              | NON-UST-02R      | 11/17/2018   | Chemtech -SO |
| 11/26/2018 | Solid  | J6003-14 | Percent Solids | Cool 4 deg C | BEMS01   | L61              | NON-UST-04       | 11/17/2018   | Chemtech -SO |
| 11/26/2018 | Solid  | J6003-15 | Percent Solids | Cool 4 deg C | BEMS01   | L61              | NON-UST-05A      | 11/17/2018   | Chemtech -SO |
| 11/26/2018 | Solid  | J6003-19 | Percent Solids | Cool 4 deg C | BEMS01   | L61              | AOC2-04          | 11/17/2018   | Chemtech -SO |
| 11/26/2018 | Solid  | J6003-20 | Percent Solids | Cool 4 deg C | BEMS01   | L61              | AOC2-03A         | 11/17/2018   | Chemtech -SO |
| 11/26/2018 | Solid  | J6006-01 | Percent Solids | Cool 4 deg C | PSEG01   | L31              | RS-CON-WALL-1    | 11/19/2018   | Chemtech -SO |

Date/Time 11-19-18 7:00 PM Date/Time 11-19-18 5:40 PM  
 Received by: [Signature] Received by: [Signature]  
 Relinquished by: [Signature] Relinquished by: [Signature]

# WORKLIST(Hardcopy Internal Chain)

2B 99312

WorkList Name : %1-111918

WorkList ID : 119362

Date : 11/19/2018 7:58:39 AM

| Due Date   | Matrix | Sample   | Test           | Preservative | Customer | Storage Location | Customer Sample      | Collect Date | Method       |
|------------|--------|----------|----------------|--------------|----------|------------------|----------------------|--------------|--------------|
| 11/26/2018 | Solid  | J6006-03 | Percent Solids | Cool 4 deg C | PSEG01   | L31              | RS-CON-WALL-2        | 11/19/2018   | Chemtech -SO |
| 11/26/2018 | Solid  | J6006-05 | Percent Solids | Cool 4 deg C | PSEG01   | L31              | RS-CON-FLOOR         | 11/19/2018   | Chemtech -SO |
| 11/22/2018 | Solid  | J6010-01 | Percent Solids | Cool 4 deg C | PSEG01   | L31              | FS-RT3786            | 11/19/2018   | Chemtech -SO |
| 11/22/2018 | Solid  | J6010-02 | Percent Solids | Cool 4 deg C | PSEG01   | L31              | FS-RT3786            | 11/19/2018   | Chemtech -SO |
| 11/29/2018 | Solid  | J6013-01 | Percent Solids | Cool 4 deg C | YANN01   | L33              | DISPOSAL-WASTE-CLASS | 11/19/2018   | Chemtech -SO |

Date/Time 11/19/18 11:00 AM  
 Received by: Jo  
 Relinquished by: Jo

Date/Time 11/19/18 5:40 PM  
 Received by: Jo  
 Relinquished by: Jo