



## Analytical Summary Report

Analysis Method: 9045D  
Parameter: Corrosivity  
Run Number: LB99315  
BalanceID: WC SC-4

Analyst By : Sweta  
Supervisor Review By : JIGNESH  
Slope : 99.3  
pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			19.3	4.01	11/19/2018	17:30
2	CAL2	1	Water			19.2	7.00	11/19/2018	17:32
3	CAL3	1	Water			19.2	10.04	11/19/2018	17:34
4	ICV	1	Water			19.3	7.01	11/19/2018	17:36
5	CCV1	1	Water			19.4	2.00	11/19/2018	17:38
6	J5763-02	1	Solid	20.02	20	18.5	9.06	11/19/2018	17:40
7	J5763-02DUP	1	Solid	20.03	20	18.6	9.07	11/19/2018	17:42
8	J5763-04	1	Solid	20.02	20	18.2	8.72	11/19/2018	17:44
9	J5775-02	1	Solid	20.03	20	18.5	8.19	11/19/2018	17:46
10	J5999-01	1	Solid	20.03	20	16.3	9.35	11/19/2018	17:48
11	J5999-02	1	Solid	20.02	20	16.5	9.42	11/19/2018	17:50
12	J5999-03	1	Solid	20.04	20	16.7	10.24	11/19/2018	17:52
13	J6000-02	1	Solid	20.02	20	18.5	6.22	11/19/2018	17:54
14	J6006-02	1	Solid	20.03	20	18.5	8.82	11/19/2018	17:56
15	J6006-04	1	Solid	20.03	20	18.5	9.47	11/19/2018	17:58
16	CCV2	1	Water			19.2	12.00	11/19/2018	18:00
17	J6006-06	1	Solid	20.04	20	18.5	10.47	11/19/2018	18:02
18	J6010-03	1	Solid	20.02	20	18.5	7.48	11/19/2018	18:04
19	CCV3	1	Water			19.4	2.00	11/19/2018	18:06

LB99315

# WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity111918

WorkList ID : 119408

Date : 11/19/2018 4:26:32 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/19/2018	Solid	J5763-02	Corrosivity	Cool 4 deg C	PSEG05	L31	HR-06-111918-A	11/19/2018	9045D
11/19/2018	Solid	J5763-04	Corrosivity	Cool 4 deg C	PSEG05	L31	HR-06-111918-B	11/19/2018	9045D
11/19/2018	Solid	J5775-02	Corrosivity	Cool 4 deg C	PSEG05	L31	TR-05-111918	11/19/2018	9045D
11/14/2018	Solid	J5999-01	Corrosivity	Cool 4 deg C	USAR03	L51	SSSI-1114-S-TCLP-2	11/14/2018	9045D
11/14/2018	Solid	J5999-02	Corrosivity	Cool 4 deg C	USAR03	L51	SSSI-1114-S-TCLP-3	11/14/2018	9045D
11/14/2018	Solid	J5999-03	Corrosivity	Cool 4 deg C	USAR03	L51	SSSI-1114-S-TCLP-4	11/14/2018	9045D
11/19/2018	Solid	J6000-02	Corrosivity	Cool 4 deg C	PSEG01	L31	FS-12072-11929	11/19/2018	9045D
11/19/2018	Solid	J6006-02	Corrosivity	Cool 4 deg C	PSEG01	L31	RS-CON-WALL-1	11/19/2018	9045D
11/19/2018	Solid	J6006-04	Corrosivity	Cool 4 deg C	PSEG01	L31	RS-CON-WALL-2	11/19/2018	9045D
11/19/2018	Solid	J6006-06	Corrosivity	Cool 4 deg C	PSEG01	L31	RS-CON-FLOOR	11/19/2018	9045D
11/19/2018	Solid	J6010-03	Corrosivity	Cool 4 deg C	PSEG01	L31	FS-RT3786	11/19/2018	9045D

Date/Time 11/19/18 5:15  
 Received by: sp  
 Relinquished by: [Signature]

Date/Time 11/19/18 6:10  
 Received by: [Signature]  
 Relinquished by: [Signature]