

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB99317

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP65054
calibration std. phosphate 1 ppm	WP69620
calibration std. phosphate 0.5 ppm	WP69619
calibration std. phosphate 0.3 ppm	WP69618
calibration std. 0 ppm	WP69615
phosphate ICV std.	WP69622
phosphate CCV std.	WP69621
calibration std. phosphate 0.1 ppm	WP69617
calibration std. phosphate 0.05 ppm	WP69616
Combined reagent	WP69625
11 N sulfuric acid	WP66905
Phenolphthalein indicator	WP66907
Sodium hydroxide, 1N	WP66924

Intercept: -0.0007Slope: 0.6495Regression: 0.999827

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	11/16/2018	15:00
2	CAL2	0.05	1	50	50	0.033	11/16/2018	15:01
3	CAL3	0.10	1	50	50	0.065	11/16/2018	15:01
4	CAL4	0.30	1	50	50	0.186	11/16/2018	15:02
5	CAL5	0.50	1	50	50	0.330	11/16/2018	15:02
6	CAL6	1.00	1	50	50	0.648	11/16/2018	15:03



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB99317

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.325	0.501	11/16/2018	15:03
2	ICB		1	50	50	0.000	0.001	11/16/2018	15:04
3	CCV1	0.50	1	50	50	0.327	0.505	11/16/2018	15:04
4	CCB1		1	50	50	0.000	0.001	11/16/2018	15:05
5	PB114926BL		1	50	50	0.000	0.001	11/16/2018	15:05
6	PB114926BS	0.50	1	50	50	0.322	0.497	11/16/2018	13:06
7	J5982-01		1	50	50	1.254	1.932	11/16/2018	15:06
8	J5982-01DUP		1	50	50	1.174	1.809	11/16/2018	15:07
9	J5982-01MS	0.50	1	50	50	1.810	2.788	11/16/2018	15:07
10	J5982-01MSD	0.50	1	50	50	1.828	2.816	11/16/2018	15:08
11	J5982-01		10	50	50	0.138	0.214	11/16/2018	15:16
12	J5982-01DUP		10	50	50	0.145	0.224	11/16/2018	15:16
13	J5982-01MS	0.50	10	50	50	0.166	0.257	11/16/2018	15:17
14	J5982-01MSD	0.50	10	50	50	0.164	0.254	11/16/2018	15:17
15	CCV2	0.50	1	50	50	0.328	0.506	11/16/2018	15:18
16	CCB2		1	50	50	0.000	0.001	11/16/2018	15:18