

LB99327


7471B

INSTRUMENT ID: CVZ

Linear

 μ Abs.:

5184E



A= 0.0000e+000

B= 1.9207e-004

C= 3.0042e-002

Rho= 0.9999556

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.032	0.032	9	0.000	9				
0.2	0.200	0.210	0.010	939	0.0 %	939				
2.5	2.500	2.443	-0.057	12564	0.0 %	12564				
5.0	5.000	4.979	-0.021	25765	0.0 %	25765				
7.5	7.500	7.548	0.048	39142	0.0 %	39142				
10.0	10.000	9.988	-0.012	51845	0.0 %	51845				

LB99327 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 20 Nov 2018 14:44:13

Sample ID	Extended ID	μ Abs	Conc	Strd CondMethod	Units	Date	Type	Type
0.0 - 1	50	9	-	0.00007471B	PPB	20 Nov 2018 15:06:20	S	Std
0.2 - 1	50-2	939	-	0.20007471B	PPB	20 Nov 2018 15:08:38	S	Std
2.5 - 1	52.5	12564	-	2.50007471B	PPB	20 Nov 2018 15:10:56	S	Std
5.0 - 1	55	25765	-	5.00007471B	PPB	20 Nov 2018 15:13:13	S	Std
7.5 - 1	57.5	39142	-	7.50007471B	PPB	20 Nov 2018 15:15:30	S	Std
10.0 - 1	57.5	51845	-	10.00007471B	PPB	20 Nov 2018 15:17:48	S	Std
ICV83 - 1	ICV83	20708	4.0074	-7471B	PPB	20 Nov 2018 15:20:29	U	SMPL
ICB83 - 1	ICB83	-88	0.0131	-7471B	PPB	20 Nov 2018 15:22:45	U	SMPL
CCV59 - 1	CCV59	25782	4.9820	-7471B	PPB	20 Nov 2018 15:25:02	U	SMPL
CCB59 - 1	CCB59	-106	0.0097	-7471B	PPB	20 Nov 2018 15:27:18	U	SMPL
CRA - 1	CRA	930	0.2087	-7471B	PPB	20 Nov 2018 15:29:35	U	SMPL
HighStd - 1	HighStd	50879	9.8023	-7471B	PPB	20 Nov 2018 15:31:51	U	SMPL
PB114957BL - 1	PBS	-7	0.0287	-7471B	PPB	20 Nov 2018 15:36:26	U	SMPL
PB114957BS - 1	LCSS	18537	3.5904	-7471B	PPB	20 Nov 2018 15:38:43	U	SMPL
J5993-01 - 1	20180925-WCC-1-COMPOSITE	32209	6.2164	-7471B	PPB	20 Nov 2018 15:41:00	U	SMPL
J5993-02 - 1	20180926-WCC-2-COMPOSITE	88960	17.1165	-7471B	PPB	20 Nov 2018 15:43:18	U	SMPL
J5993-03 - 1	20180927-WCC-3-COMPOSITE	79512	15.3019	-7471B	PPB	20 Nov 2018 15:45:37	U	SMPL
J5993-04 - 1	20180928-AMENDED-WCC-1-COMPOSITE	22505	4.3526	-7471B	PPB	20 Nov 2018 15:47:58	U	SMPL
J5993-05 - 1	20180929-AMENDED-WCC-2-COMPOSITE	111296	21.4066	-7471B	PPB	20 Nov 2018 15:50:16	U	SMPL
CCV60 - 1	CCV60	25131	4.8569	-7471B	PPB	20 Nov 2018 15:52:34	U	SMPL
CCB60 - 1	CCB60	-144	0.0024	-7471B	PPB	20 Nov 2018 15:54:53	U	SMPL
J5993-06 - 1	20180930-AMENDED-WCC-3-COMPOSITE	115120	22.1411	-7471B	PPB	20 Nov 2018 15:57:11	U	SMPL
J5993-06DUP - 1	20180930-AMENDED-WCC-3-COMPOSITEDUP	113789	21.8854	-7471B	PPB	20 Nov 2018 15:59:26	U	SMPL
J5993-06MS - 1	20180930-AMENDED-WCC-3-COMPOSITEMS	113784	21.8845	-7471B	PPB	20 Nov 2018 16:01:46	U	SMPL
J5993-06MSD - 1	20180930-AMENDED-WCC-3-COMPOSITEMSD	114641	22.0491	-7471B	PPB	20 Nov 2018 16:04:07	U	SMPL
PB114958BL - 1	PBS	-48	0.0208	-7471B	PPB	20 Nov 2018 16:09:13	U	SMPL
PB114958BS - 1	LCSS	18675	3.6169	-7471B	PPB	20 Nov 2018 16:11:30	U	SMPL
J5763-01 - 1	HR-06-111918-A	5041	0.9983	-7471B	PPB	20 Nov 2018 16:13:47	U	SMPL
J5763-03 - 1	HR-06-111918-B	2498	0.5098	-7471B	PPB	20 Nov 2018 16:16:04	U	SMPL
J5775-01 - 1	TR-05-111918	2173	0.4474	-7471B	PPB	20 Nov 2018 16:18:23	U	SMPL
J5996-25 - 1	AOC1-06-A	1879	0.3909	-7471B	PPB	20 Nov 2018 16:20:39	U	SMPL
CCV61 - 1	CCV61	25636	4.9539	-7471B	PPB	20 Nov 2018 16:22:55	U	SMPL
CCB61 - 1	CCB61	-108	0.0093	-7471B	PPB	20 Nov 2018 16:25:12	U	SMPL
J5999-05 - 1	SSSI-1115-S-AB-1	4003	0.7989	-7471B	PPB	20 Nov 2018 16:27:29	U	SMPL
J5999-07 - 1	SSSI-1115-S-AB-2	14580	2.8304	-7471B	PPB	20 Nov 2018 16:29:46	U	SMPL
J5999-09 - 1	SSSI-1116-S-AB-4	5608	1.1072	-7471B	PPB	20 Nov 2018 16:32:02	U	SMPL
J5999-10 - 1	SSSI-1116-S-AB-5	1318	0.2832	-7471B	PPB	20 Nov 2018 16:34:20	U	SMPL
J6000-01 - 1	FS-12072-11929	3210	0.6466	-7471B	PPB	20 Nov 2018 16:36:37	U	SMPL
J6003-03 - 1	NON-UST-03R-A	348	0.0969	-7471B	PPB	20 Nov 2018 16:38:54	U	SMPL
J6003-03DUP - 1	NON-UST-03R-ADUP	379	0.1028	-7471B	PPB	20 Nov 2018 16:41:11	U	SMPL
J6003-04 - 1	J6003-03MS	19077	3.6942	-7471B	PPB	20 Nov 2018 16:43:28	U	SMPL
J6003-05 - 1	J6003-03MSD	18614	3.6052	-7471B	PPB	20 Nov 2018 16:45:46	U	SMPL
J6003-07 - 1	NON-UST-03R-C	698	0.1641	-7471B	PPB	20 Nov 2018 16:48:03	U	SMPL
CCV62 - 1	CCV62	25713	4.9687	-7471B	PPB	20 Nov 2018 16:50:20	U	SMPL
CCB62 - 1	CCB62	-101	0.0106	-7471B	PPB	20 Nov 2018 16:52:37	U	SMPL
J6003-15 - 1	NON-UST-05A	462	0.1188	-7471B	PPB	20 Nov 2018 16:54:55	U	SMPL
J6006-01 - 1	RS-CON-WALL-1	96	0.0485	-7471B	PPB	20 Nov 2018 16:57:11	U	SMPL
J6006-03 - 1	RS-CON-WALL-2	75	0.0444	-7471B	PPB	20 Nov 2018 16:59:27	U	SMPL
J6006-05 - 1	RS-CON-FLOOR	4733	0.9391	-7471B	PPB	20 Nov 2018 17:01:44	U	SMPL
J6010-01 - 1	FS-RT3786	3087	0.6230	-7471B	PPB	20 Nov 2018 17:04:01	U	SMPL
J6013-01 - 1	DISPOSAL-WASTE-CLASS	3188	0.6424	-7471B	PPB	20 Nov 2018 17:06:18	U	SMPL
J5993-06LX5 - 1		23581	4.5592	-7471B	PPB	20 Nov 2018 17:08:36	U	SMPL
J5993-06A - 1		132775	25.5321	-7471B	PPB	20 Nov 2018 17:10:53	U	SMPL
J6003-03LX5 - 1		-431	-0.0527	-7471B	PPB	20 Nov 2018 17:13:12	U	SMPL
J6003-03A - 1		20452	3.9582	-7471B	PPB	20 Nov 2018 17:15:32	U	SMPL
CCV63 - 1	CCV63	25228	4.8756	-7471B	PPB	20 Nov 2018 17:17:48	U	SMPL
CCB63 - 1	CCB63	-101	0.0106	-7471B	PPB	20 Nov 2018 17:20:06	U	SMPL
J5993-06DLX5 - 1		20214	3.9125	-7471B	PPB	20 Nov 2018 17:22:23	U	SMPL
J5993-06DUPDLX5 - 1		19415	3.7591	-7471B	PPB	20 Nov 2018 17:24:39	U	SMPL
J5993-06MSDLX5 - 1		21029	4.0691	-7471B	PPB	20 Nov 2018 17:26:57	U	SMPL
J5993-06MSDDLX5 - 1		20824	4.0297	-7471B	PPB	20 Nov 2018 17:29:15	U	SMPL
J5993-06LDLX25 - 1		4310	0.8579	-7471B	PPB	20 Nov 2018 17:31:34	U	SMPL
J5993-06ADLX5 - 1		23355	4.5158	-7471B	PPB	20 Nov 2018 17:33:53	U	SMPL
ChkStd - 1	ChkStd	35734	6.8935	-7471B	PPB	20 Nov 2018 17:48:22	U	SMPL
CCV64 - 1	CCV64	25002	4.8322	-7471B	PPB	20 Nov 2018 17:50:56	U	SMPL
CCB64 - 1	CCB64	-102	0.0105	-7471B	PPB	20 Nov 2018 17:53:14	U	SMPL