



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 11/21/2018

OVENTEMP IN Celsius(°C): 107
Time IN: 17:25
In Date: 11/20/2018
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M Oven-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 11/21/2018
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M Sc-1

QC:LB99332

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
J5765-03	NB-08-112018	1	1.14	9.96	9.4	93.7
J5768-05	OK-03-112018	2	1.17	9.98	9.3	92.3
J5772-03	PL-01-112018-A	3	1.13	9.59	8.49	87.0
J5772-05	PL-01-112018-B	4	1.13	9.96	9.07	89.9
J5772-07	PL-01-112018-C	5	1.1	9.95	8.41	82.6
J6012-01	NS-18	6	1.15	9.67	6.34	60.9
J6012-02	NS-18	7	1.15	9.67	6.34	60.9
J6012-03	NS-19	8	1.18	9.85	8.23	81.3
J6012-04	NS-19	9	1.18	9.85	8.23	81.3
J6012-05	NS-20	10	1.11	9.97	5.43	48.8
J6012-06	NS-20	11	1.11	9.97	5.43	48.8
J6012-07	NS-21	12	1.13	9.63	6.12	58.7
J6012-08	NS-21	13	1.13	9.63	6.12	58.7
J6012-09	NS-22	14	1.12	9.87	6.25	58.6
J6012-10	NS-22	15	1.12	9.87	6.25	58.6
J6012-11	NS-23	16	1.12	9.59	5.56	52.4
J6012-12	NS-23	17	1.12	9.59	5.56	52.4
J6012-13	NS-24	18	1.11	9.85	5.31	48.1
J6012-14	NS-24	19	1.11	9.85	5.31	48.1
J6012-15	NS-25	20	1.13	9.8	4.91	43.6
J6012-16	NS-25	21	1.13	9.8	4.91	43.6
J6012-17	NS-26	22	1.16	9.63	5.24	48.2
J6012-18	NS-26	23	1.16	9.63	5.24	48.2
J6012-19	PLATING-11	24	1.13	9.55	7.25	72.7
J6012-20	PLATING-11	25	1.13	9.55	7.25	72.7
J6014-01	TS-BUCKETS	26	1.14	9.88	8.3	81.9
J6014-02	TS-BUCKETS	27	1.14	9.88	8.3	81.9
J6017-01	VW-RO-294-CON	28	1.14	9.9	8.89	88.5
J6024-01	PS-SOIL	29	1.12	9.97	8.78	86.6
J6024-02	PS-SOIL	30	1.12	9.97	8.78	86.6
J6025-01	A366336	31	1.00	2.00	2.00	100.0
J6025-02	B368415	32	1.00	2.00	2.00	100.0

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$