

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 11/21/2018

TEMP1 IN: 103 11/21/2018 07:10 TEMP OUT: 104 11/21/2018 08:20

TEMP2 IN: 104 11/21/2018 09:00 TEMP OUT: 104 11/21/2018 10:05

TEMP3 IN: 105 11/21/2018 11:20 TEMP OUT: 105 11/21/2018 13:00

TEMP4 IN: 105 11/21/2018 13:35 TEMP OUT: 105 11/21/2018 15:10

Run Number: LB99337

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB99337BL	LB99337BL	1.4103	1.4103	100	1.4103	1.4103	1.4103	0.0000	0
2	LB99337BS	LB99337BS	1.3711	1.3711	100	1.4220	1.4220	1.4220	0.0509	509
3	J6023-01	111918-JETT-E-D-B	1.3699	1.3699	250	1.3794	1.3794	1.3794	0.0095	38
4	J6023-01DUP	111918-JETT-E-D-BDUP	1.4008	1.4008	250	1.4105	1.4105	1.4105	0.0097	38.8
5	J6023-02	111918-JETT-E-D-M	1.3736	1.3736	500	1.3815	1.3815	1.3815	0.0079	15.8
6	J6023-03	111918-JETT-E-D-S	1.4005	1.4005	500	1.4053	1.4053	1.4053	0.0048	9.6
7	J6023-04	111918-JETT-E-U-B	1.3929	1.3929	500	1.4071	1.4071	1.4071	0.0142	28.4
8	J6023-05	111918-JETT-E-U-M	1.3790	1.3790	500	1.3870	1.3870	1.3870	0.0080	16
9	J6023-06	111918-JETT-E-U-S	1.3713	1.3713	500	1.3763	1.3763	1.3763	0.0050	10
10	J6023-07	111918-DBRI-E-D-M	1.3913	1.3913	500	1.4056	1.4056	1.4056	0.0143	28.6
11	J6023-08	111918-DBRI-E-U-M	1.4034	1.4034	450	1.4098	1.4098	1.4098	0.0064	14.2
12	J6023-09	111918-TIPC-E-D-B	1.3985	1.3985	500	1.4057	1.4057	1.4057	0.0072	14.4
13	J6023-10	111918-TIPC-E-D-S	1.4074	1.4074	500	1.4145	1.4145	1.4145	0.0071	14.2
14	J6023-11	111918-TIPC-E-U-B	1.4124	1.4124	500	1.4215	1.4215	1.4215	0.0091	18.2
15	J6023-12	111918-TIPC-E-U-S	1.4011	1.4011	500	1.4088	1.4088	1.4088	0.0077	15.4

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 11/21/2018

TEMP1 IN: 103 11/21/2018 07:10 TEMP OUT: 104 11/21/2018 08:20

TEMP2 IN: 104 11/21/2018 09:00 TEMP OUT: 104 11/21/2018 10:05

TEMP3 IN: 105 11/21/2018 11:20 TEMP OUT: 105 11/21/2018 13:00

TEMP4 IN: 105 11/21/2018 13:35 TEMP OUT: 105 11/21/2018 15:10

Run Number: LB99337

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
16	J6023-13	111918-CHDE-E-D-B	1.4093	1.4093	450	1.4203	1.4203	1.4203	0.0110	24.4
17	J6023-14	111918-CHDE-E-D-S	1.3907	1.3907	500	1.3997	1.3997	1.3997	0.0090	18
18	J6023-15	111918-CHDE-E-U-B	1.3707	1.3707	500	1.3814	1.3814	1.3814	0.0107	21.4
19	J6023-16	111918-CHDE-E-U-S	1.3831	1.3831	250	1.3883	1.3883	1.3883	0.0052	20.8
20	J6023-16DUP	111918-CHDE-E-U-SDUP	1.4057	1.4057	250	1.4111	1.4111	1.4111	0.0054	21.6

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS112118

WorkList ID : 119473

Date : 11/21/2018 7:52:47 AM

1399337

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/26/2018	Water	J6023-01 C	TSS	Cool 4 deg C	HDR103	L62	111918-JETT-E-D-B	11/19/2018	SM2540 D
11/26/2018	Water	J6023-02 C	TSS	Cool 4 deg C	HDR103	L62	111918-JETT-E-D-M	11/19/2018	SM2540 D
11/26/2018	Water	J6023-03 C	TSS	Cool 4 deg C	HDR103	L62	111918-JETT-E-D-S	11/19/2018	SM2540 D
11/26/2018	Water	J6023-04 C	TSS	Cool 4 deg C	HDR103	L62	111918-JETT-E-U-B	11/19/2018	SM2540 D
11/26/2018	Water	J6023-05 C	TSS	Cool 4 deg C	HDR103	L62	111918-JETT-E-U-M	11/19/2018	SM2540 D
11/26/2018	Water	J6023-06 C	TSS	Cool 4 deg C	HDR103	L62	111918-JETT-E-U-S	11/19/2018	SM2540 D
11/26/2018	Water	J6023-07 C	TSS	Cool 4 deg C	HDR103	L62	111918-DBRI-E-D-M	11/19/2018	SM2540 D
11/26/2018	Water	J6023-08 C	TSS	Cool 4 deg C	HDR103	L62	111918-DBRI-E-U-M	11/19/2018	SM2540 D
11/22/2018	Water	J6023-09 C	TSS	Cool 4 deg C	HDR103	L62	111918-TIPC-E-D-B	11/19/2018	SM2540 D
11/22/2018	Water	J6023-10 C	TSS	Cool 4 deg C	HDR103	L62	111918-TIPC-E-D-S	11/19/2018	SM2540 D
11/22/2018	Water	J6023-11 C	TSS	Cool 4 deg C	HDR103	L62	111918-TIPC-E-U-B	11/19/2018	SM2540 D
11/22/2018	Water	J6023-12 C	TSS	Cool 4 deg C	HDR103	L62	111918-TIPC-E-U-S	11/19/2018	SM2540 D
11/26/2018	Water	J6023-13 C	TSS	Cool 4 deg C	HDR103	L62	111918-CHDE-E-D-B	11/19/2018	SM2540 D
11/26/2018	Water	J6023-14 C	TSS	Cool 4 deg C	HDR103	L62	111918-CHDE-E-D-S	11/19/2018	SM2540 D
11/26/2018	Water	J6023-15 C	TSS	Cool 4 deg C	HDR103	L62	111918-CHDE-E-U-B	11/19/2018	SM2540 D
11/26/2018	Water	J6023-16 C	TSS	Cool 4 deg C	HDR103	L62	111918-CHDE-E-U-S	11/19/2018	SM2540 D

Date/Time

Received by:

Relinquished by:

11/21/18, 08:00

AS

AS

Date/Time

Received by:

Relinquished by:

11/21/18, 12:30

AS

AS