



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: pH

Supervisor Review By : JIGNESH

Run Number: LB99343

Slope : 99.0

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.3	4.01	11/21/2018	10:50
2	CAL2	1	Water			20.4	7.00	11/21/2018	10:52
3	CAL3	1	Water			20.3	10.04	11/21/2018	10:54
4	ICV	1	Water			20.2	7.01	11/21/2018	10:56
5	CCV1	1	Water			20.5	2.00	11/21/2018	10:58
6	J6011-01	1	Solid	20.02	20	20.9	6.67	11/21/2018	11:00
7	J6011-02	1	Solid	20.02	20	20.9	6.71	11/21/2018	11:02
8	J6011-03	1	Solid	20.04	20	20.9	8.22	11/21/2018	11:04
9	J6011-04	1	Solid	20.02	20	20.8	6.81	11/21/2018	11:06
10	J6011-05	1	Solid	20.04	20	20.9	7.01	11/21/2018	11:08
11	J6011-06	1	Solid	20.02	20	21.0	7.20	11/21/2018	11:10
12	J6011-06DUP	1	Solid	20.03	20	21.1	7.21	11/21/2018	11:12
13	CCV2	1	Water			20.2	12.00	11/21/2018	11:14

WORKLIST(Hardcopy Internal Chain)

LB 99343

WorkList Name : pH112118

WorkList ID : 119489

Date : 11/21/2018 10:37:09 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/15/2018	Solid	J6011-01	pH	Cool 4 deg C	USEP01	A11	MYAJN5	11/15/2018	9045D
11/15/2018	Solid	J6011-02	pH	Cool 4 deg C	USEP01	A11	MYAJN6	11/15/2018	9045D
11/14/2018	Solid	J6011-03	pH	Cool 4 deg C	USEP01	A11	MYAJN7	11/14/2018	9045D
11/15/2018	Solid	J6011-04	pH	Cool 4 deg C	USEP01	A11	MYAJN8	11/15/2018	9045D
11/15/2018	Solid	J6011-05	pH	Cool 4 deg C	USEP01	A11	MYAJN9	11/15/2018	9045D
11/15/2018	Solid	J6011-06	pH	Cool 4 deg C	USEP01	A11	MYAJP0	11/15/2018	9045D

Date/Time 11/21/18 10:40

Received by: SP

Relinquished by: CP

Date/Time 11/21/18 11:30

Received by: CP

Relinquished by: SP