



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : sweta

Parameter: pH

Supervisor Review By : JIGNESH

Run Number: LB99344

Slope : 99.0

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.3	4.01	11/21/2018	10:50
2	CAL2	1	Water			20.4	7.00	11/21/2018	10:52
3	CAL3	1	Water			20.3	10.04	11/21/2018	10:54
4	ICV	1	Water			20.2	7.01	11/21/2018	10:56
5	CCV1	1	Water			20.2	2.00	11/21/2018	11:16
6	J6037-01	1	Water			17.6	5.46	11/21/2018	11:18
7	J6037-01DUP	1	Water			17.7	5.47	11/21/2018	11:20
8	CCV2	1	Water			20.3	12.00	11/21/2018	11:22

WORKLIST(Hardcopy Internal Chain)

B99344

WorkList Name : pH112118-1

WorkList ID : 119491

Date : 11/21/2018 11:21:37 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/19/2018	Water	J6037-01	pH	Cool 4 deg C	STER06	L43	POOL-WATER	11/19/2018	SM 4500-H B

Date/Time 11/21/18 11:25
 Received by: SP
 Relinquished by: CP

Date/Time 11/21/18 11:36
 Received by: CP
 Relinquished by: SP