


7471B

LB99345
INSTRUMENT ID: CV1

Linear

 μ Abs.:

52017



A= 0.0000e+000

B= 1.9186e-004

C= 1.4482e-002

Rho= 0.9999735

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.016	0.016	10	0.000	10				
0.2	0.200	0.204	0.004	986	0.0 %	986				
2.5	2.500	2.492	-0.008	12913	0.0 %	12913				
5.0	5.000	4.952	-0.048	25737	0.0 %	25737				
7.5	7.500	7.541	0.041	39230	0.0 %	39230				
10.0	10.000	9.994	-0.006	52017	0.0 %	52017				

LB99345 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 21 Nov 2018 11:50:19

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	10	-	0.0000	7471B	PPB	21 Nov 2018 11:51:07	S	Std
0.2 - 1	50.2	986	-	0.2000	7471B	PPB	21 Nov 2018 11:53:26	S	Std
2.5 - 1	52.5	12913	-	2.5000	7471B	PPB	21 Nov 2018 11:55:43	S	Std
5.0 - 1	55	25737	-	5.0000	7471B	PPB	21 Nov 2018 11:58:01	S	Std
7.5 - 1	57.5	39230	-	7.5000	7471B	PPB	21 Nov 2018 12:00:18	S	Std
10.0 - 1	510	52017	-	10.0000	7471B	PPB	21 Nov 2018 12:05:23	S	Std
ICV85 - 1	ICV85	20514	3.9503	-	7471B	PPB	21 Nov 2018 12:08:02	U	SMPL
ICB85 - 1	ICB85	-111	-0.0068	-	7471B	PPB	21 Nov 2018 12:10:18	U	SMPL
CCV69 - 1	CCV69	25961	4.9954	-	7471B	PPB	21 Nov 2018 12:12:35	U	SMPL
CCB69 - 1	CCB69	-104	-0.0055	-	7471B	PPB	21 Nov 2018 12:14:50	U	SMPL
CRA - 1	CRA	969	0.2004	-	7471B	PPB	21 Nov 2018 12:17:07	U	SMPL
HighStd - 1	HighStd	51644	9.9229	-	7471B	PPB	21 Nov 2018 12:19:23	U	SMPL
ChkStd - 1	ChkStd	36723	7.0602	-	7471B	PPB	21 Nov 2018 12:21:39	U	SMPL
PB114996BL - 1	PBS	-156	-0.0154	-	7471B	PPB	21 Nov 2018 12:23:58	U	SMPL
PB114996BS - 1	LCSS	21139	4.0702	-	7471B	PPB	21 Nov 2018 12:28:56	U	SMPL
J5765-03 - 1	NB-08-112018	504	0.1112	-	7471B	PPB	21 Nov 2018 12:31:13	U	SMPL
J5768-05 - 1	OK-03-112018	1124	0.2301	-	7471B	PPB	21 Nov 2018 12:33:30	U	SMPL
J5772-03 - 1	PL-01-112018-A	722	0.1530	-	7471B	PPB	21 Nov 2018 12:35:48	U	SMPL
J5772-05 - 1	PL-01-112018-B	987	0.2038	-	7471B	PPB	21 Nov 2018 12:38:06	U	SMPL
J5772-07 - 1	PL-01-112018-C	690	0.1469	-	7471B	PPB	21 Nov 2018 12:40:23	U	SMPL
CCV70 - 1	CCV70	25582	4.9226	-	7471B	PPB	21 Nov 2018 12:42:39	U	SMPL
CCB70 - 1	CCB70	-99	-0.0045	-	7471B	PPB	21 Nov 2018 12:44:55	U	SMPL
J6012-01 - 1	NS-18	2906	0.5720	-	7471B	PPB	21 Nov 2018 12:47:12	U	SMPL
J6012-03 - 1	NS-19	21638	4.1659	-	7471B	PPB	21 Nov 2018 12:49:28	U	SMPL
J6012-03DUP - 1	NS-19DUP	21663	4.1707	-	7471B	PPB	21 Nov 2018 12:51:44	U	SMPL
J6012-03MS - 1	NS-19MS	40923	7.8660	-	7471B	PPB	21 Nov 2018 12:59:27	U	SMPL
J6012-03MSD - 1	NS-19MSD	39530	7.5987	-	7471B	PPB	21 Nov 2018 13:01:43	U	SMPL
J6012-05 - 1	NS-20	-12	0.0122	-	7471B	PPB	21 Nov 2018 13:04:02	U	SMPL
J6012-07 - 1	NS-21	213	0.0553	-	7471B	PPB	21 Nov 2018 13:06:22	U	SMPL
J6012-09 - 1	NS-22	601	0.1298	-	7471B	PPB	21 Nov 2018 13:08:39	U	SMPL
J6012-11 - 1	NS-23	3379	0.6628	-	7471B	PPB	21 Nov 2018 13:10:57	U	SMPL
J6012-13 - 1	NS-24	1067	0.2192	-	7471B	PPB	21 Nov 2018 13:13:14	U	SMPL
CCV71 - 1	CCV71	26193	5.0399	-	7471B	PPB	21 Nov 2018 13:15:30	U	SMPL
CCB71 - 1	CCB71	-82	-0.0013	-	7471B	PPB	21 Nov 2018 13:17:46	U	SMPL
J6012-15 - 1	NS-25	657	0.1405	-	7471B	PPB	21 Nov 2018 13:20:03	U	SMPL
J6012-17 - 1	NS-26	1441	0.2910	-	7471B	PPB	21 Nov 2018 13:22:20	U	SMPL
J6012-19 - 1	PLATING-11	79067	15.1843	-	7471B	PPB	21 Nov 2018 13:24:36	U	SMPL
J6014-01 - 1	TS-BUCKETS	12403	2.3941	-	7471B	PPB	21 Nov 2018 13:26:53	U	SMPL
J6024-01 - 1	PS-SOIL	62716	12.0472	-	7471B	PPB	21 Nov 2018 13:29:13	U	SMPL
J6012-03LX5 - 1		4326	0.8445	-	7471B	PPB	21 Nov 2018 13:31:30	U	SMPL
J6012-03A - 1		42336	8.1371	-	7471B	PPB	21 Nov 2018 13:33:49	U	SMPL
J5993-02DLX5 - 1		16446	3.1698	-	7471B	PPB	21 Nov 2018 13:36:07	U	SMPL
J5993-03DLX5 - 1		14162	2.7316	-	7471B	PPB	21 Nov 2018 13:38:27	U	SMPL
J5993-05DLX5 - 1		20350	3.9188	-	7471B	PPB	21 Nov 2018 13:40:44	U	SMPL
CCV72 - 1	CCV72	25942	4.9917	-	7471B	PPB	21 Nov 2018 13:43:01	U	SMPL
CCB72 - 1	CCB72	-103	-0.0053	-	7471B	PPB	21 Nov 2018 13:45:17	U	SMPL
J6012-19DLX5 - 1		15817	3.0491	-	7471B	PPB	21 Nov 2018 13:47:34	U	SMPL
J6024-01DLX2 - 1		30264	5.8209	-	7471B	PPB	21 Nov 2018 13:49:51	U	SMPL
CCV73 - 1	CCV73	26201	5.0414	-	7471B	PPB	21 Nov 2018 13:52:07	U	SMPL
CCB73 - 1	CCB73	-106	-0.0059	-	7471B	PPB	21 Nov 2018 13:54:25	U	SMPL