



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 11/26/2018

OVENTEMP IN Celsius(°C): 107
Time IN: 15:35
In Date: 11/21/2018
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M Oven-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:24
Out Date: 11/22/2018
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M Sc-1

QC:LB99348

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
J6039-01	CB4M7	1	1.14	9.94	8.03	78.3
J6039-02	CB4M8	2	1.16	9.68	7.19	70.8
J6039-03	CB4M9	3	1.19	9.6	8.48	86.7
J6039-04	CB4N0	4	1.1	9.82	8.76	87.8
J6039-05	CB4N1	5	1.14	9.88	8.29	81.8
J6039-06	CB4N2	6	1.15	9.97	8.54	83.8
J6039-07	CB4N3	7	1.16	9.61	8.24	83.8
J6039-08	CB4N4	8	1.18	9.9	7.87	76.7
J6039-09	CB4N5	9	1.13	9.59	7.91	80.1
J6039-10	CB4N6	10	1.14	9.88	8.54	84.7
J6039-11	CB4N7	11	1.18	9.71	8.68	87.9
J6039-12	CB4N8	12	1.15	9.68	8.12	81.7
J6039-13	CB4N9	13	1.16	9.57	8.3	84.9
J6039-14	CB4P0	14	1.13	9.66	8.47	86.0
J6039-15	CB4P1	15	1.18	9.58	7.63	76.8
J6039-16	CB4P2	16	1.14	9.66	8.09	81.6
J6039-17	CB4P3	17	1.18	9.82	8.25	81.8
J6039-18	CB4P4	18	1.19	9.63	8.32	84.5
J6039-19	CB4P4MS	19	1.19	9.63	8.32	84.5
J6039-20	CB4P4MSD	20	1.19	9.63	8.32	84.5
J6039-21	CB4P5	21	1.18	9.86	8.56	85.0
J6039-22	CB4P6	22	1.1	9.76	8.12	81.1

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$