



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 11/26/2018

OVENTEMP IN Celsius(°C): 107
Time IN: 16:25
In Date: 11/21/2018
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M Oven-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:30
Out Date: 11/22/2018
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M Sc-1

QC:LB99354

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
J6041-01	CB4K5	1	1.13	9.68	7.57	75.3
J6041-02	CB4Z9	2	1.16	9.54	8.6	88.8
J6041-03	CB500	3	1.1	9.89	8.19	80.7
J6041-04	CB501	4	1.17	9.53	6.6	65.0
J6041-05	CB4P7	5	1.19	9.78	8.63	86.6
J6041-06	CB4Z5	6	1.14	9.94	8.05	78.5
J6041-07	CB4M6	7	1.14	9.82	8.83	88.6
J6041-08	CB4Z6	8	1.11	9.7	8.37	84.5
J6041-09	CB4Z7	9	1.18	9.96	8.58	84.3
J6041-10	CB4Z8	10	1.1	9.82	8.47	84.5
J6041-11	CB4H4	11	1.19	9.96	8.71	85.7
J6041-12	CB4H4MS	12	1.19	9.96	8.71	85.7
J6041-13	CB4H4MSD	13	1.19	9.96	8.71	85.7
J6041-14	CB4P8	14	1.12	9.75	8.15	81.5
J6041-15	CB4P9	15	1.16	9.57	8.5	87.3
J6041-16	CB4Q0	16	1.18	9.97	8.46	82.8
J6041-17	CB4Q1	17	1.18	9.97	8.4	82.1
J6041-18	CB4Q2	18	1.12	9.51	8.32	85.8
J6041-19	CB4Q3	19	1.16	9.89	9.2	92.1
J6041-20	CB4Q4	20	1.14	9.97	8.94	88.3
J6041-21	CB4Y1	21	1.18	9.95	5.16	45.4
J6041-22	CB4Y2	22	1.15	9.87	8.27	81.7

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$