



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 11/26/2018

OVENTEMP IN Celsius(°C): 107
Time IN: 16:35
In Date: 11/21/2018
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M Oven-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:36
Out Date: 11/22/2018
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M Sc-1

QC:LB99360

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
J6042-01	CB4Y3	1	1.12	9.91	7.69	74.7
J6042-02	CB4Y4	2	1.11	9.64	8.4	85.5
J6042-03	CB4Y5	3	1.13	9.97	9.27	92.1
J6042-04	CB4Y6	4	1.11	9.71	8.6	87.1
J6042-05	CB4Y7	5	1.13	9.76	8.55	86.0
J6042-06	CB4Y8	6	1.12	9.85	9.1	91.4
J6042-07	CB4Y9	7	1.11	9.55	8.15	83.4
J6042-08	CB4Z0	8	1.12	9.61	7.95	80.4
J6042-09	CB4Z1	9	1.13	9.51	8.18	84.1
J6042-10	CB4Z2	10	1.13	9.53	8.1	83.0
J6042-11	CB4Z3	11	1.11	9.64	8.01	80.9
J6042-12	CB4Z4	12	1.13	9.97	9.41	93.7
J6042-13	CB4Z4MS	13	1.13	9.97	9.41	93.7
J6042-14	CB4Z4MSD	14	1.13	9.97	9.41	93.7
J6042-15	CB507	15	1.12	9.98	8.75	86.1
J6042-16	CB4W0	16	1.12	9.95	6.01	55.4
J6042-17	CB4W1	17	1.12	9.85	8.52	84.8
J6042-18	CB4W2	18	1.11	9.77	8.58	86.3
J6042-19	CB4W3	19	1.12	9.56	8.81	91.1
J6042-20	CB4W4	20	1.11	9.94	9.14	90.9
J6042-21	CB4W5	21	1.12	9.86	8.5	84.4
J6042-22	CB4W6	22	1.12	9.6	8.08	82.1

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$