



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 11/26/2018

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 11/21/2018
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M Oven-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:45
Out Date: 11/22/2018
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M Sc-1

QC:LB99364

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
J6045-01	CB4S2	1	1.14	9.98	8.74	86.0
J6045-02	CB4S3	2	1.12	9.75	8.51	85.6
J6045-03	CB4S4	3	1.12	9.7	8.42	85.1
J6045-04	CB4S5	4	1.13	9.98	7.74	74.7
J6045-05	CB4S6	5	1.12	9.73	8.21	82.3
J6045-06	CB4S7	6	1.11	9.98	8.52	83.5
J6045-07	CB4S8	7	1.12	9.55	8.39	86.2
J6045-08	CB4S9	8	1.12	9.75	8.55	86.1
J6045-09	CB4T0	9	1.13	9.65	8.4	85.3
J6045-10	CB4T1	10	1.12	9.75	8.2	82.0
J6045-11	CB4T2	11	1.12	9.9	7.33	70.7
J6045-12	CB4T3	12	1.12	9.6	8.24	84.0
J6045-13	CB4T4	13	1.11	9.95	7.07	67.4
J6045-14	CB4T4MS	14	1.11	9.95	7.07	67.4
J6045-15	CB4T4MSD	15	1.11	9.95	7.07	67.4
J6045-16	CB4T5	16	1.13	9.98	8.83	87.0
J6045-17	CB4T6	17	1.14	9.54	8.55	88.2
J6045-18	CB4T7	18	1.13	9.85	8.83	88.3
J6045-19	CB4T8	19	1.13	9.77	8.21	81.9
J6045-20	CB4T9	20	1.12	9.97	7.8	75.5
J6045-21	CB502	21	1.14	9.66	8.08	81.5
J6045-22	CB503	22	1.13	9.93	8.15	79.8

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$