



PERCENT SOLID

Supervisor: apatel  
Analyst: JIGNESH  
Date: 11/26/2018

OVENTEMP IN Celsius(°C): 107  
Time IN: 17:00  
In Date: 11/21/2018  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M Oven-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 07:45  
Out Date: 11/22/2018  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M Sc-1

QC:LB99364

| Lab ID   | Client Sample ID | Dish# | Dish Wt(g)<br>(A) | Dish +<br>Sample Wt(g)<br>(B) | Dish + Dry<br>Sample Wt(g)<br>(C) | % Solid |
|----------|------------------|-------|-------------------|-------------------------------|-----------------------------------|---------|
| J6045-01 | CB4S2            | 1     | 1.14              | 9.98                          | 8.74                              | 86.0    |
| J6045-02 | CB4S3            | 2     | 1.12              | 9.75                          | 8.51                              | 85.6    |
| J6045-03 | CB4S4            | 3     | 1.12              | 9.7                           | 8.42                              | 85.1    |
| J6045-04 | CB4S5            | 4     | 1.13              | 9.98                          | 7.74                              | 74.7    |
| J6045-05 | CB4S6            | 5     | 1.12              | 9.73                          | 8.21                              | 82.3    |
| J6045-06 | CB4S7            | 6     | 1.11              | 9.98                          | 8.52                              | 83.5    |
| J6045-07 | CB4S8            | 7     | 1.12              | 9.55                          | 8.39                              | 86.2    |
| J6045-08 | CB4S9            | 8     | 1.12              | 9.75                          | 8.55                              | 86.1    |
| J6045-09 | CB4T0            | 9     | 1.13              | 9.65                          | 8.4                               | 85.3    |
| J6045-10 | CB4T1            | 10    | 1.12              | 9.75                          | 8.2                               | 82.0    |
| J6045-11 | CB4T2            | 11    | 1.12              | 9.9                           | 7.33                              | 70.7    |
| J6045-12 | CB4T3            | 12    | 1.12              | 9.6                           | 8.24                              | 84.0    |
| J6045-13 | CB4T4            | 13    | 1.11              | 9.95                          | 7.07                              | 67.4    |
| J6045-14 | CB4T4MS          | 14    | 1.11              | 9.95                          | 7.07                              | 67.4    |
| J6045-15 | CB4T4MSD         | 15    | 1.11              | 9.95                          | 7.07                              | 67.4    |
| J6045-16 | CB4T5            | 16    | 1.13              | 9.98                          | 8.83                              | 87.0    |
| J6045-17 | CB4T6            | 17    | 1.14              | 9.54                          | 8.55                              | 88.2    |
| J6045-18 | CB4T7            | 18    | 1.13              | 9.85                          | 8.83                              | 88.3    |
| J6045-19 | CB4T8            | 19    | 1.13              | 9.77                          | 8.21                              | 81.9    |
| J6045-20 | CB4T9            | 20    | 1.12              | 9.97                          | 7.8                               | 75.5    |
| J6045-21 | CB502            | 21    | 1.14              | 9.66                          | 8.08                              | 81.5    |
| J6045-22 | CB503            | 22    | 1.13              | 9.93                          | 8.15                              | 79.8    |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

✓ 1309364

# WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-J6045

WorkList ID : 119523

Date : 11/21/2018 3:35:33 PM

| Due Date   | Matrix | Sample   | Test           | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method       |
|------------|--------|----------|----------------|--------------|----------|------------------|-----------------|--------------|--------------|
| 11/28/2018 | Solid  | J6045-01 | Percent Solids | Cool 4 deg C | USEP04   | A41              | CB4S2           | 11/19/2018   | Chemtech -SO |
| 11/28/2018 | Solid  | J6045-02 | Percent Solids | Cool 4 deg C | USEP04   | A41              | CB4S3           | 11/19/2018   | Chemtech -SO |
| 11/28/2018 | Solid  | J6045-03 | Percent Solids | Cool 4 deg C | USEP04   | A41              | CB4S4           | 11/19/2018   | Chemtech -SO |
| 11/28/2018 | Solid  | J6045-04 | Percent Solids | Cool 4 deg C | USEP04   | A41              | CB4S5           | 11/19/2018   | Chemtech -SO |
| 11/28/2018 | Solid  | J6045-05 | Percent Solids | Cool 4 deg C | USEP04   | A41              | CB4S6           | 11/19/2018   | Chemtech -SO |
| 11/28/2018 | Solid  | J6045-06 | Percent Solids | Cool 4 deg C | USEP04   | A41              | CB4S7           | 11/19/2018   | Chemtech -SO |
| 11/28/2018 | Solid  | J6045-07 | Percent Solids | Cool 4 deg C | USEP04   | A41              | CB4S8           | 11/19/2018   | Chemtech -SO |
| 11/28/2018 | Solid  | J6045-08 | Percent Solids | Cool 4 deg C | USEP04   | A41              | CB4S9           | 11/19/2018   | Chemtech -SO |
| 11/28/2018 | Solid  | J6045-09 | Percent Solids | Cool 4 deg C | USEP04   | A41              | CB4T0           | 11/19/2018   | Chemtech -SO |
| 11/28/2018 | Solid  | J6045-10 | Percent Solids | Cool 4 deg C | USEP04   | A41              | CB4T1           | 11/19/2018   | Chemtech -SO |
| 11/28/2018 | Solid  | J6045-11 | Percent Solids | Cool 4 deg C | USEP04   | A41              | CB4T2           | 11/19/2018   | Chemtech -SO |
| 11/28/2018 | Solid  | J6045-12 | Percent Solids | Cool 4 deg C | USEP04   | A41              | CB4T3           | 11/19/2018   | Chemtech -SO |
| 11/28/2018 | Solid  | J6045-13 | Percent Solids | Cool 4 deg C | USEP04   | A41              | CB4T4           | 11/19/2018   | Chemtech -SO |
| 11/28/2018 | Solid  | J6045-14 | Percent Solids | Cool 4 deg C | USEP04   | A41              | CB4T4MS         | 11/19/2018   | Chemtech -SO |
| 11/28/2018 | Solid  | J6045-15 | Percent Solids | Cool 4 deg C | USEP04   | A41              | CB4T4MSD        | 11/19/2018   | Chemtech -SO |
| 11/28/2018 | Solid  | J6045-16 | Percent Solids | Cool 4 deg C | USEP04   | A41              | CB4T5           | 11/19/2018   | Chemtech -SO |
| 11/28/2018 | Solid  | J6045-17 | Percent Solids | Cool 4 deg C | USEP04   | A41              | CB4T6           | 11/19/2018   | Chemtech -SO |
| 11/28/2018 | Solid  | J6045-18 | Percent Solids | Cool 4 deg C | USEP04   | A41              | CB4T7           | 11/19/2018   | Chemtech -SO |
| 11/28/2018 | Solid  | J6045-19 | Percent Solids | Cool 4 deg C | USEP04   | A41              | CB4T8           | 11/19/2018   | Chemtech -SO |
| 11/28/2018 | Solid  | J6045-20 | Percent Solids | Cool 4 deg C | USEP04   | A41              | CB4T9           | 11/19/2018   | Chemtech -SO |
| 11/28/2018 | Solid  | J6045-21 | Percent Solids | Cool 4 deg C | USEP04   | A41              | CB502           | 11/19/2018   | Chemtech -SO |

Date/Time 11-21-18 3:40 PM      Date/Time 11-21-18      5:00 PM  
 Received by: JA      Received by: CS  
 Relinquished by: CS      Relinquished by: JO

1329364

# WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-J6045

WorkList ID : 119523

Date : 11/21/2018 3:35:33 PM

| Due Date   | Matrix | Sample   | Test           | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method       |
|------------|--------|----------|----------------|--------------|----------|------------------|-----------------|--------------|--------------|
| 11/28/2018 | Solid  | J6045-22 | Percent Solids | Cool 4 deg C | USEP04   | A41              | CB503           | 11/19/2018   | Chemtech -SO |

Date/Time 11-21-18 3:40 PM  
 Received by: JA  
 Relinquished by: [Signature]

Date/Time 11-21-18 5:00 PM  
 Received by:  
 Relinquished by: JA