



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : JIGNESH

Run Number: LB99365

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.6	4.01	11/21/2018	16:30
2	CAL2	1	Water			20.4	7.00	11/21/2018	16:32
3	CAL3	1	Water			20.3	10.04	11/21/2018	16:34
4	ICV	1	Water			20.6	7.01	11/21/2018	16:36
5	CCV1	1	Water			20.4	2.00	11/21/2018	16:38
6	J5770-06	1	Solid	20.02	20	20.5	7.90	11/21/2018	16:40
7	J5770-06DUP	1	Solid	20.03	20	20.6	7.91	11/21/2018	16:42
8	J5770-08	1	Solid	20.03	20	20.5	7.80	11/21/2018	16:44
9	J5774-04	1	Solid	20.04	20	20.6	7.89	11/21/2018	16:46
10	J5774-06	1	Solid	20.02	20	20.9	9.49	11/21/2018	16:48
11	J5774-08	1	Solid	20.02	20	20.5	8.93	11/21/2018	16:50
12	J6028-05	1	Solid	20.02	20	21.0	9.79	11/21/2018	16:52
13	J6028-07	1	Solid	20.04	20	21.0	9.53	11/21/2018	16:54
14	J6028-09	1	Solid	20.02	20	21.2	9.58	11/21/2018	16:56
15	CCV2	1	Water			20.5	12.00	11/21/2018	16:58

WORKLIST(Hardcopy Internal Chain)

LB 99365

WorkList Name : corrosivity112118

WorkList ID : 119524

Date : 11/21/2018 4:02:45 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/21/2018	Solid	J5770-06	Corrosivity	Cool 4 deg C	PSEG05		OR-02-112118-A	11/21/2018	9045D
11/21/2018	Solid	J5770-08	Corrosivity	Cool 4 deg C	PSEG05		OR-02-112118-B	11/21/2018	9045D
11/21/2018	Solid	J5774-04	Corrosivity	Cool 4 deg C	PSEG05		SU-02-112118	11/21/2018	9045D
11/21/2018	Solid	J5774-06	Corrosivity	Cool 4 deg C	PSEG05		SU-03-112118-A	11/21/2018	9045D
11/21/2018	Solid	J5774-08	Corrosivity	Cool 4 deg C	PSEG05		SU-03-112118-B	11/21/2018	9045D
11/20/2018	Solid	J6028-05	Corrosivity	Cool 4 deg C	TECT01	L22	COMP-1	11/20/2018	9045D
11/20/2018	Solid	J6028-07	Corrosivity	Cool 4 deg C	TECT01	L22	COMP-2	11/20/2018	9045D
11/20/2018	Solid	J6028-09	Corrosivity	Cool 4 deg C	TECT01	L22	COMP-3	11/20/2018	9045D

Date/Time 11/21/18 4:20
 Received by: SP
 Relinquished by: SP

Date/Time 11/21/18 5:20
 Received by: CP
 Relinquished by: SP