



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: pH

Supervisor Review By : JIGNESH

Run Number: LB99366

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.6	4.01	11/21/2018	16:30
2	CAL2	1	Water			20.4	7.00	11/21/2018	16:32
3	CAL3	1	Water			20.3	10.04	11/21/2018	16:34
4	ICV	1	Water			20.6	7.01	11/21/2018	16:36
5	CCV1	1	Water			20.4	2.00	11/21/2018	17:02
6	J6044-01	1	Solid	20.02	20	20.2	6.68	11/21/2018	17:04
7	J6044-02	1	Solid	20.02	20	20.8	6.07	11/21/2018	17:06
8	J6044-02DUP	1	Solid	20.04	20	20.9	6.08	11/21/2018	17:08
9	CCV2	1	Water			20.1	12.00	11/21/2018	17:16

WORKLIST(Hardcopy Internal Chain)

LB 99366

WorkList Name : pH112118-3

WorkList ID : 119526

Date : 11/21/2018 4:03:09 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/21/2018	Solid	J6044-01	pH	Cool 4 deg C	PSEG01	L31	30055	11/21/2018	9045D
11/21/2018	Solid	J6044-02	pH	Cool 4 deg C	PSEG01	L31	30107	11/21/2018	9045D

Date/Time 11/21/18 4:20
 Received by: SP
 Relinquished by: CP

Date/Time 11/21/18 5:20
 Received by: CP
 Relinquished by: SP