



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: pH

Supervisor Review By : JIGNESH

Run Number: LB99367

Slope : 99.1

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.6	4.01	11/21/2018	16:30
2	CAL2	1	Water			20.4	7.00	11/21/2018	16:32
3	CAL3	1	Water			20.3	10.04	11/21/2018	16:34
4	ICV	1	Water			20.6	7.01	11/21/2018	16:36
5	CCV1	1	Water			20.4	2.00	11/21/2018	17:02
6	J6044-03	1	Water			13.6	5.08	11/21/2018	17:10
7	J6044-03DUP	1	Water			13.7	5.09	11/21/2018	17:12
8	J6044-04	1	Water			13.8	5.36	11/21/2018	17:14
9	CCV2	1	Water			20.1	12.00	11/21/2018	17:16

WORKLIST(Hardcopy Internal Chain)

WorkList Name : pH112118-2

WorkList ID : 119525

Date : 11/21/2018 4:02:56 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/21/2018	Water	J6044-03	pH	Cool 4 deg C	PSEG01	L31	30003	11/21/2018	9040C
11/21/2018	Water	J6044-04	pH	Cool 4 deg C	PSEG01	L31	30002	11/21/2018	9040C

Date/Time 11/21/18 4:20
 Received by: SP
 Relinquished by: CP

Date/Time 11/21/18 5:20
 Received by: CP
 Relinquished by: SP