


7470A

LB99271
INSTRUMENT ID: CV1

Linear

 μ Abs.:

35519



A= 0.0000e+000

B= 2.7888e-004

C= -1.0157e-002

Rho= 0.9998334

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.024	-0.024	-51	0.000	-51				
0.2	0.200	0.179	-0.021	680	0.0 %	680				
2.5	2.500	2.478	-0.022	8921	0.0 %	8921				
5.0	5.000	5.084	0.084	18266	0.0 %	18266				
7.5	7.500	7.588	0.088	27245	0.0 %	27245				
10.0	10.000	9.895	-0.105	35519	0.0 %	35519				

LB99371 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 21 Nov 2018 17:38:01

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc.	Method	Units	Date	Type	Type
0.0 - 1	50	-51	-	0.0000	7470A	PPB	21 Nov 2018 17:39:44	S	Std
0.2 - 1	50.2	680	-	0.2000	7470A	PPB	21 Nov 2018 17:41:53	S	Std
2.5 - 1	52.5	8921	-	2.5000	7470A	PPB	21 Nov 2018 17:44:01	S	Std
5.0 - 1	55	18266	-	5.0000	7470A	PPB	21 Nov 2018 17:46:09	S	Std
7.5 - 1	57.5	27245	-	7.5000	7470A	PPB	21 Nov 2018 17:48:17	S	Std
10.0 - 1	510	35519	-	10.0000	7470A	PPB	21 Nov 2018 17:52:51	S	Std
ICV87 - 1	ICV87	14071	3.9140	-	7470A	PPB	21 Nov 2018 17:55:22	U	SMPL
ICB87 - 1	ICB87	-71	-0.0300	-	7470A	PPB	21 Nov 2018 17:57:29	U	SMPL
CCV80 - 1	CCV80	18461	5.1382	-	7470A	PPB	21 Nov 2018 17:59:35	U	SMPL
CCB80 - 1	CCB80	-78	-0.0319	-	7470A	PPB	21 Nov 2018 18:01:42	U	SMPL
CRA - 1	CRA	669	0.1764	-	7470A	PPB	21 Nov 2018 18:03:49	U	SMPL
HighStd - 1	HighStd	35634	9.9275	-	7470A	PPB	21 Nov 2018 18:05:56	U	SMPL
ChkStd - 1	ChkStd	25331	7.0542	-	7470A	PPB	21 Nov 2018 18:08:02	U	SMPL
PB115019BL - 1	PBW	-120	-0.0436	-	7470A	PPB	21 Nov 2018 18:10:10	U	SMPL
PB115019BS - 1	LCSW	14293	3.9759	-	7470A	PPB	21 Nov 2018 18:12:19	U	SMPL
J5765-04 - 1	NB-08-112018	-86	-0.0341	-	7470A	PPB	21 Nov 2018 18:14:26	U	SMPL
J5765-04DUP - 1	NB-08-112018DUP	-33	-0.0194	-	7470A	PPB	21 Nov 2018 18:16:34	U	SMPL
J5765-04MS - 1	NB-08-112018MS	14297	3.9770	-	7470A	PPB	21 Nov 2018 18:18:43	U	SMPL
J5765-04MSD - 1	NB-08-112018MSD	14260	3.9667	-	7470A	PPB	21 Nov 2018 18:20:51	U	SMPL
J5768-06 - 1	OK-03-112018	-78	-0.0319	-	7470A	PPB	21 Nov 2018 18:22:59	U	SMPL
CCV81 - 1	CCV81	18431	5.1299	-	7470A	PPB	21 Nov 2018 18:25:06	U	SMPL
CCB81 - 1	CCB81	-66	-0.0286	-	7470A	PPB	21 Nov 2018 18:27:13	U	SMPL
J5772-04 - 1	PL-01-112018-A	-23	-0.0166	-	7470A	PPB	21 Nov 2018 18:29:19	U	SMPL
J5772-06 - 1	PL-01-112018-B	-22	-0.0163	-	7470A	PPB	21 Nov 2018 18:31:26	U	SMPL
J5772-08 - 1	PL-01-112018-C	-24	-0.0169	-	7470A	PPB	21 Nov 2018 18:33:33	U	SMPL
J6014-02 - 1	TS-BUCKETS	-14	-0.0141	-	7470A	PPB	21 Nov 2018 18:35:40	U	SMPL
J6024-02 - 1	PS-SOIL	-20	-0.0157	-	7470A	PPB	21 Nov 2018 18:37:47	U	SMPL
PB115019TB - 1	PB114991TB	-3	-0.0110	-	7470A	PPB	21 Nov 2018 18:39:55	U	SMPL
J5765-04LX5 - 1		-13	-0.0138	-	7470A	PPB	21 Nov 2018 18:42:03	U	SMPL
J5765-04A - 1		13655	3.7979	-	7470A	PPB	21 Nov 2018 18:44:12	U	SMPL
CCV82 - 1	CCV82	18033	5.0189	-	7470A	PPB	21 Nov 2018 18:46:21	U	SMPL
CCB82 - 1	CCB82	-58	-0.0263	-	7470A	PPB	21 Nov 2018 18:48:28	U	SMPL