

LB99384


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

49066



A= 0.0000e+000

B= 2.0249e-004

C= 6.1741e-003

Rho= 0.9998988

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.028	0.028	109	0.000	109				
0.2	0.200	0.193	-0.007	922	0.0 %	922				
2.5	2.500	2.431	-0.069	11975	0.0 %	11975				
5.0	5.000	5.022	0.022	24772	0.0 %	24772				
7.5	7.500	7.584	0.084	37422	0.0 %	37422				
10.0	10.000	9.942	-0.058	49066	0.0 %	49066				

LB99384 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 26 Nov 2018 14:08:38

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	109	-	0.00007471B	PPB		26 Nov 2018 14:12:03	S	Std
0.2 - 1	50.2	922	-	0.20007471B	PPB		26 Nov 2018 14:14:21	S	Std
2.5 - 1	52.5	11975	-	2.50007471B	PPB		26 Nov 2018 14:16:39	S	Std
5.0 - 1	55	24772	-	5.00007471B	PPB		26 Nov 2018 14:18:57	S	Std
7.5 - 1	57.5	37422	-	7.50007471B	PPB		26 Nov 2018 14:23:54	S	Std
10.0 - 1	57.0	49066	-	10.00007471B	PPB		26 Nov 2018 14:28:47	S	Std
ICV88 - 1	ICV88	20069	4.0700	-7471B	PPB		26 Nov 2018 14:31:28	U	SMPL
ICB88 - 1	ICB88	-49	-0.0037	-7471B	PPB		26 Nov 2018 14:33:43	U	SMPL
CCV83 - 1	CCV83	24860	5.0401	-7471B	PPB		26 Nov 2018 14:36:00	U	SMPL
CCB83 - 1	CCB83	-49	-0.0037	-7471B	PPB		26 Nov 2018 14:38:15	U	SMPL
CRA - 1	CRA	885	0.1854	-7471B	PPB		26 Nov 2018 14:40:32	U	SMPL
HighStd - 1	HighStd	49211	9.9710	-7471B	PPB		26 Nov 2018 14:42:48	U	SMPL
ChkStd - 1	ChkStd	34799	7.0527	-7471B	PPB		26 Nov 2018 14:45:04	U	SMPL
PB115040BL - 1	PBS	-82	-0.0104	-7471B	PPB		26 Nov 2018 14:47:23	U	SMPL
PB115040BS - 1	LCSS	22596	4.5817	-7471B	PPB		26 Nov 2018 14:49:41	U	SMPL
J5770-05 - 1	OR-02-112118-A	2094	0.4302	-7471B	PPB		26 Nov 2018 14:51:58	U	SMPL
J5770-07 - 1	OR-02-112118-B	2467	0.5057	-7471B	PPB		26 Nov 2018 14:54:16	U	SMPL
J5774-03 - 1	SU-02-112118	2323	0.4766	-7471B	PPB		26 Nov 2018 14:56:34	U	SMPL
J5774-03DUP - 1	SU-02-112118DUP	2310	0.4739	-7471B	PPB		26 Nov 2018 14:58:51	U	SMPL
J5774-03MS - 1	SU-02-112118MS	20436	4.1443	-7471B	PPB		26 Nov 2018 15:01:07	U	SMPL
CCV84 - 1	CCV84	24794	5.0268	-7471B	PPB		26 Nov 2018 15:03:23	U	SMPL
CCB84 - 1	CCB84	-80	-0.0100	-7471B	PPB		26 Nov 2018 15:05:39	U	SMPL
J5774-03MSD - 1	SU-02-112118MSD	20737	4.2053	-7471B	PPB		26 Nov 2018 15:07:56	U	SMPL
J5774-05 - 1	SU-03-112118-A	1622	0.3346	-7471B	PPB		26 Nov 2018 15:10:11	U	SMPL
J5774-07 - 1	SU-03-112118-B	2596	0.5318	-7471B	PPB		26 Nov 2018 15:12:27	U	SMPL
J6028-04 - 1	COMP-1	7050	1.4337	-7471B	PPB		26 Nov 2018 15:14:44	U	SMPL
J6028-06 - 1	COMP-2	15070	3.0577	-7471B	PPB		26 Nov 2018 15:17:00	U	SMPL
J6028-08 - 1	COMP-3	7838	1.5933	-7471B	PPB		26 Nov 2018 15:19:17	U	SMPL
J6055-01 - 1	RICH-REC-ROAD-STONE-7	881	0.1846	-7471B	PPB		26 Nov 2018 15:21:35	U	SMPL
J5774-03LX5 - 1		495	0.1064	-7471B	PPB		26 Nov 2018 15:23:53	U	SMPL
J5774-03A - 1		23717	4.8087	-7471B	PPB		26 Nov 2018 15:26:11	U	SMPL
CCV85 - 1	CCV85	25169	5.1027	-7471B	PPB		26 Nov 2018 15:28:27	U	SMPL
CCB85 - 1	CCB85	-63	-0.0066	-7471B	PPB		26 Nov 2018 15:30:44	U	SMPL