



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: sweta

Run Number: LB99390

Ambient Barometric Pressure (mmHg): 745

Barometric Scale ID: 55971

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.0	80.6	81.5	11/26/2018	11:25
2	J6044-03		1	9	>100.0	>212.0	>212.0	11/26/2018	11:55
3	J6044-03DUP		1	9	>100.0	>212.0	>212.0	11/26/2018	12:20
4	J6044-04		1	7	>100.0	>212.0	>212.0	11/26/2018	12:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 99390

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp112618

WorkList ID : 119562

Date : 11/26/2018 10:59:19 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/28/2018	Water	J6044-03	Flash Point	Cool 4 deg C	PSEG01	L31	30003	11/21/2018	1010A
11/28/2018	Water	J6044-04	Flash Point	Cool 4 deg C	PSEG01	L31	30002	11/21/2018	1010A

Date/Time 11/26/18 11:00 AM
 Received by: KS
 Relinquished by: CP

Date/Time 11/26/18 12:55 PM
 Received by: CP
 Relinquished by: KS