


7470A

LB99401
INSTRUMENT ID: CX1

Linear

 μ Abs.:

37007



A= 0.0000e+000

B= 2.6649e-004

C= 3.5189e-002

Rho= 0.9996504

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.033	0.033	-10	0.000	-10				
0.2	0.200	0.197	-0.003	608	0.0 %	608				
2.5	2.500	2.452	-0.048	9069	0.0 %	9069				
5.0	5.000	4.927	-0.073	18355	0.0 %	18355				
7.5	7.500	7.694	0.194	28741	0.0 %	28741				
10.0	10.000	9.897	-0.103	37007	0.0 %	37007				

LB99401 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 27 Nov 2018 09:13:14

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-10	-	0.00007470A	PPB		27 Nov 2018 09:16:40	S	Std
0.2 - 1	50.2	608	-	0.20007470A	PPB		27 Nov 2018 09:18:49	S	Std
2.5 - 1	52.5	9069	-	2.50007470A	PPB		27 Nov 2018 09:20:58	S	Std
5.0 - 1	55	18355	-	5.00007470A	PPB		27 Nov 2018 09:23:06	S	Std
7.5 - 1	57.5	28741	-	7.50007470A	PPB		27 Nov 2018 09:25:14	S	Std
10.0 - 1	50	37007	-	10.00007470A	PPB		27 Nov 2018 09:29:49	S	Std
ICV89 - 1	ICV89	14153	3.8068	-7470A	PPB		27 Nov 2018 09:32:31	U	SMPL
ICB89 - 1	ICB89	-58	0.0197	-7470A	PPB		27 Nov 2018 09:37:00	U	SMPL
CCV86 - 1	CCV86	18175	4.8787	-7470A	PPB		27 Nov 2018 09:39:06	U	SMPL
CCB86 - 1	CCB86	-120	0.0032	-7470A	PPB		27 Nov 2018 09:41:13	U	SMPL
CRA - 1	CRA	586	0.1914	-7470A	PPB		27 Nov 2018 09:43:19	U	SMPL
HighStd - 1	HighStd	36823	9.8482	-7470A	PPB		27 Nov 2018 09:45:26	U	SMPL
ChkStd - 1	ChkStd	26328	7.0514	-7470A	PPB		27 Nov 2018 09:47:33	U	SMPL
PB115058BL - 1	PBW	-18	0.0304	-7470A	PPB		27 Nov 2018 09:52:05	U	SMPL
PB115058BS - 1	LCSW	13693	3.6843	-7470A	PPB		27 Nov 2018 09:54:12	U	SMPL
J5770-06 - 1	OR-02-112118-A	430	0.1498	-7470A	PPB		27 Nov 2018 09:56:20	U	SMPL
J5770-06DUP - 1	OR-02-112118-ADUP	419	0.1468	-7470A	PPB		27 Nov 2018 09:58:28	U	SMPL
J5770-06MS - 1	OR-02-112118-AMS	12697	3.4188	-7470A	PPB		27 Nov 2018 10:00:37	U	SMPL
J5770-06MSD - 1	OR-02-112118-AMSD	13751	3.6997	-7470A	PPB		27 Nov 2018 10:02:46	U	SMPL
J5770-08 - 1	OR-02-112118-B	380	0.1365	-7470A	PPB		27 Nov 2018 10:04:52	U	SMPL
CCV87 - 1	CCV87	18256	4.9002	-7470A	PPB		27 Nov 2018 10:06:59	U	SMPL
CCB87 - 1	CCB87	-161	-0.0077	-7470A	PPB		27 Nov 2018 10:09:05	U	SMPL
J5774-04 - 1	SU-02-112118	516	0.1727	-7470A	PPB		27 Nov 2018 10:11:12	U	SMPL
J5774-06 - 1	SU-03-112118-A	479	0.1628	-7470A	PPB		27 Nov 2018 10:13:19	U	SMPL
J5774-08 - 1	SU-03-112118-B	379	0.1362	-7470A	PPB		27 Nov 2018 10:15:26	U	SMPL
J6028-05 - 1	COMP-1	132	0.0704	-7470A	PPB		27 Nov 2018 10:17:33	U	SMPL
J6028-07 - 1	COMP-2	395	0.1405	-7470A	PPB		27 Nov 2018 10:19:40	U	SMPL
J6028-09 - 1	COMP-3	447	0.1543	-7470A	PPB		27 Nov 2018 10:21:48	U	SMPL
J6044-01 - 1	30055	446	0.1540	-7470A	PPB		27 Nov 2018 10:23:56	U	SMPL
J6044-02 - 1	30107	391	0.1394	-7470A	PPB		27 Nov 2018 10:26:04	U	SMPL
PB115059BL - 1	PBW	-28	0.0277	-7470A	PPB		27 Nov 2018 10:28:13	U	SMPL
PB115059BS - 1	LCSW	14093	3.7908	-7470A	PPB		27 Nov 2018 10:30:20	U	SMPL
CCV88 - 1	CCV88	18348	4.9248	-7470A	PPB		27 Nov 2018 10:32:27	U	SMPL
CCB88 - 1	CCB88	-82	0.0133	-7470A	PPB		27 Nov 2018 10:34:34	U	SMPL
J6044-03 - 1	30003	521	0.1740	-7470A	PPB		27 Nov 2018 10:36:42	U	SMPL
J6044-04 - 1	30002	461	0.1580	-7470A	PPB		27 Nov 2018 10:38:48	U	SMPL
J6044-04DUP - 1	30002DUP	471	0.1607	-7470A	PPB		27 Nov 2018 10:40:55	U	SMPL
J6044-04MS - 1	30002MS	14057	3.7813	-7470A	PPB		27 Nov 2018 10:43:02	U	SMPL
J6044-04MSD - 1	30002MSD	13766	3.7037	-7470A	PPB		27 Nov 2018 10:45:10	U	SMPL
PB115060BL - 1	PBW	-92	0.0107	-7470A	PPB		27 Nov 2018 10:47:18	U	SMPL
PB115060BS - 1	LCSW	14030	3.7741	-7470A	PPB		27 Nov 2018 10:52:01	U	SMPL
J6030-01 - 1	DTW-01	13	0.0387	-7470A	PPB		27 Nov 2018 10:54:10	U	SMPL
J6030-02 - 1	DTW-03	331	0.1234	-7470A	PPB		27 Nov 2018 11:00:54	U	SMPL
J6030-02DUP - 1	DTW-03DUP	277	0.1090	-7470A	PPB		27 Nov 2018 11:03:01	U	SMPL
CCV89 - 1	CCV89	18317	4.9165	-7470A	PPB		27 Nov 2018 11:05:08	U	SMPL
CCB89 - 1	CCB89	-116	0.0043	-7470A	PPB		27 Nov 2018 11:07:15	U	SMPL
J6030-03 - 1	J6030-02MS	14384	3.8684	-7470A	PPB		27 Nov 2018 11:11:51	U	SMPL
J6030-04 - 1	J6030-02MSD	14040	3.7767	-7470A	PPB		27 Nov 2018 11:13:58	U	SMPL
J6030-05 - 1	DUP-01	-108	0.0064	-7470A	PPB		27 Nov 2018 11:16:06	U	SMPL
J6031-01 - 1	VTW-01	-76	0.0149	-7470A	PPB		27 Nov 2018 11:18:14	U	SMPL
J6031-02 - 1	VTW-02	-119	0.0035	-7470A	PPB		27 Nov 2018 11:20:21	U	SMPL
J6031-03 - 1	J6031-02MS	14037	3.7759	-7470A	PPB		27 Nov 2018 11:25:11	U	SMPL
J6031-04 - 1	J6031-02MSD	14189	3.8164	-7470A	PPB		27 Nov 2018 11:27:19	U	SMPL
J6031-05 - 1	VTW-03	-149	-0.0045	-7470A	PPB		27 Nov 2018 11:29:27	U	SMPL
J6031-06 - 1	DUP-01	-112	0.0053	-7470A	PPB		27 Nov 2018 11:31:36	U	SMPL
J6032-01 - 1	MMTW-01	-33	0.0264	-7470A	PPB		27 Nov 2018 11:33:43	U	SMPL
CCV90 - 1	CCV90	18113	4.8621	-7470A	PPB		27 Nov 2018 11:35:50	U	SMPL
CCB90 - 1	CCB90	-150	-0.0048	-7470A	PPB		27 Nov 2018 11:37:58	U	SMPL
J6032-02 - 1	MMTW-02	33	0.0440	-7470A	PPB		27 Nov 2018 11:40:05	U	SMPL
J6032-03 - 1	MMTW-03	203	0.0893	-7470A	PPB		27 Nov 2018 11:42:12	U	SMPL
J6032-04 - 1	MMTW-DUP	646	0.2073	-7470A	PPB		27 Nov 2018 11:44:19	U	SMPL
J6032-05 - 1	J6032-01MS	13345	3.5915	-7470A	PPB		27 Nov 2018 11:49:15	U	SMPL
J6032-06 - 1	J6032-01MSD	13614	3.6632	-7470A	PPB		27 Nov 2018 11:51:23	U	SMPL
J6047-01 - 1	EP2-MW-101	-187	-0.0146	-7470A	PPB		27 Nov 2018 11:53:30	U	SMPL
J6047-02 - 1	J6047-01MS	14345	3.8580	-7470A	PPB		27 Nov 2018 11:55:38	U	SMPL
J6047-03 - 1	J6047-01MSD	13982	3.7613	-7470A	PPB		27 Nov 2018 11:57:47	U	SMPL
J6047-04 - 1	EP2-MW-102	-166	-0.0090	-7470A	PPB		27 Nov 2018 11:59:55	U	SMPL
J6047-05 - 1	EP2-FD-02	-127	0.0013	-7470A	PPB		27 Nov 2018 12:02:03	U	SMPL
CCV91 - 1	CCV91	18126	4.8656	-7470A	PPB		27 Nov 2018 12:04:10	U	SMPL
CCB91 - 1	CCB91	-135	-0.0008	-7470A	PPB		27 Nov 2018 12:06:18	U	SMPL
J6050-01 - 1	DTW-02	336	0.1247	-7470A	PPB		27 Nov 2018 12:08:25	U	SMPL
PB115058TB - 1	PB115022TB	-13	0.0317	-7470A	PPB		27 Nov 2018 12:12:57	U	SMPL
PB115059TB - 1	PB115039TB	-14	0.0315	-7470A	PPB		27 Nov 2018 12:15:04	U	SMPL
J5770-06LX5 - 1		72	0.0544	-7470A	PPB		27 Nov 2018 12:17:12	U	SMPL
J5770-06A - 1		15292	4.1104	-7470A	PPB		27 Nov 2018 12:19:19	U	SMPL
J6044-04LX5 - 1		5	0.0365	-7470A	PPB		27 Nov 2018 12:21:26	U	SMPL
J6044-04A - 1		15168	4.0773	-7470A	PPB		27 Nov 2018 12:23:34	U	SMPL
J6030-02LX5 - 1		-88	0.0117	-7470A	PPB		27 Nov 2018 12:28:08	U	SMPL
J6030-02A - 1		13189	3.5499	-7470A	PPB		27 Nov 2018 12:30:16	U	SMPL
CCV92 - 1	CCV92	18223	4.8915	-7470A	PPB		27 Nov 2018 12:32:24	U	SMPL
CCB92 - 1	CCB92	-110	0.0059	-7470A	PPB		27 Nov 2018 12:34:33	U	SMPL