



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : JIGNESH

Run Number: LB99403

Slope : 97.4

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.3	4.01	11/26/2018	16:10
2	CAL2	1	Water			20.3	7.00	11/26/2018	16:12
3	CAL3	1	Water			20.4	10.04	11/26/2018	16:14
4	ICV	1	Water			20.5	7.01	11/26/2018	16:16
5	CCV1	1	Water			20.4	2.00	11/26/2018	16:18
6	J6062-03	1	Solid	20.02	20	21.4	7.30	11/26/2018	16:20
7	J6062-03DUP	1	Solid	20.03	20	21.5	7.31	11/26/2018	16:22
8	J6064-02	1	Solid	20.02	20	20.3	6.36	11/26/2018	16:24
9	J6065-02	1	Solid	20.04	20	20.8	8.27	11/26/2018	16:26
10	CCV2	1	Water			20.3	12.00	11/26/2018	16:32

WORKLIST(Hardcopy Internal Chain)

1399403

WorkList Name : CORROSIVITY112618

WorkList ID : 119568

Date : 11/26/2018 3:23:41 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/26/2018	Solid	J6064-02	Corrosivity	Cool 4 deg C	PSEG01	L31	OILY-SOIL	11/26/2018	9045D
11/26/2018	Solid	J6065-02	Corrosivity	Cool 4 deg C	PSEG01	L31	USTS-1-2-3	11/26/2018	9045D
11/20/2018	Solid	J6062-03	Corrosivity	Cool 4 deg C	WSPO01	L63	B-20(1-10)	11/20/2018	9045D

Date/Time 11/26/18 4:00
 Received by: SP
 Relinquished by: SP

Date/Time 11/26/18 4:40
 Received by: SP
 Relinquished by: SP