

245.1


LB99409

INSTRUMENT ID: CY1

Linear

 μ Abs.:

34683



A= 0.0000e+000

B= 2.8941e-004

C= 2.4816e-002

Rho= 0.9998135

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.009	0.009	-54	0.000	-54				
0.05	0.050	0.060	0.010	121	0.0 %	121				
0.2	0.200	0.187	-0.013	560	0.0 %	560				
2.5	2.500	2.480	-0.020	8483	0.0 %	8483				
5.0	5.000	5.099	0.099	17531	0.0 %	17531				
7.5	7.500	7.353	-0.147	25321	0.0 %	25321				
10.0	10.000	10.063	0.063	34683	0.0 %	34683				

LB99409 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 27 Nov 2018 12:53:22

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD	-54	-	0.0000	245.1	PPB	27 Nov 2018 12:54:21	S	Std
0.05 - 1	50.05	121	-	0.0500	245.1	PPB	27 Nov 2018 12:59:25	S	Std
0.2 - 1	50.2	560	-	0.2000	245.1	PPB	27 Nov 2018 13:01:43	S	Std
2.5 - 1	52.5	8483	-	2.5000	245.1	PPB	27 Nov 2018 13:04:01	S	Std
5.0 - 1	55	17531	-	5.0000	245.1	PPB	27 Nov 2018 13:06:19	S	Std
7.5 - 1	57.5	25321	-	7.5000	245.1	PPB	27 Nov 2018 13:08:35	S	Std
10.0 - 1	51.0	34683	-	10.0000	245.1	PPB	27 Nov 2018 13:18:35	S	Std
ICV90 - 1	ICV90	14247	4.1481	-245.1		PPB	27 Nov 2018 13:23:53	U	SMPL
ICB90 - 1	ICB90	-125	-0.0114	-245.1		PPB	27 Nov 2018 13:26:09	U	SMPL
CCV93 - 1	CCV93	17385	5.0563	-245.1		PPB	27 Nov 2018 13:28:25	U	SMPL
CCB93 - 1	CCB93	-147	-0.0177	-245.1		PPB	27 Nov 2018 13:30:41	U	SMPL
CRA - 1	CRA	97	0.0529	-245.1		PPB	27 Nov 2018 13:32:58	U	SMPL
HighStd - 1	HighStd	33957	9.8525	-245.1		PPB	27 Nov 2018 13:35:14	U	SMPL
ChkStd - 1	ChkStd	23699	6.8836	-245.1		PPB	27 Nov 2018 13:37:30	U	SMPL
PB115061BL - 1	PBW	-177	-0.0264	-245.1		PPB	27 Nov 2018 13:39:48	U	SMPL
PB115061BS - 1	LCSW	13210	3.8480	-245.1		PPB	27 Nov 2018 13:42:07	U	SMPL
J6061-01 - 1	112018-JETT-E-D-B	-182	-0.0279	-245.1		PPB	27 Nov 2018 13:44:24	U	SMPL
J6061-02 - 1	112018-JETT-E-D-M	-111	-0.0073	-245.1		PPB	27 Nov 2018 13:46:41	U	SMPL
J6061-03 - 1	112018-JETT-E-D-S	-114	-0.0082	-245.1		PPB	27 Nov 2018 13:48:58	U	SMPL
J6061-04 - 1	112018-JETT-E-U-B	-99	-0.0038	-245.1		PPB	27 Nov 2018 13:51:16	U	SMPL
J6061-05 - 1	112018-JETT-E-U-M	-78	0.0022	-245.1		PPB	27 Nov 2018 13:53:32	U	SMPL
CCV94 - 1	CCV94	17046	4.9582	-245.1		PPB	27 Nov 2018 13:55:48	U	SMPL
CCB94 - 1	CCB94	-143	-0.0166	-245.1		PPB	27 Nov 2018 13:58:05	U	SMPL
J6061-06 - 1	112018-JETT-E-U-S	-148	-0.0180	-245.1		PPB	27 Nov 2018 14:00:22	U	SMPL
J6061-07 - 1	112018-CHDE-E-D-B	-106	-0.0059	-245.1		PPB	27 Nov 2018 14:02:38	U	SMPL
J6061-08 - 1	112018-CHDE-E-D-S	-105	-0.0056	-245.1		PPB	27 Nov 2018 14:04:54	U	SMPL
J6061-09 - 1	112018-CHDE-E-U-B	-127	-0.0119	-245.1		PPB	27 Nov 2018 14:07:10	U	SMPL
J6061-10 - 1	112018-CHDE-E-U-S	-116	-0.0088	-245.1		PPB	27 Nov 2018 14:09:27	U	SMPL
J6061-11 - 1	112018-TIPC-E-D-M	-92	-0.0018	-245.1		PPB	27 Nov 2018 14:11:44	U	SMPL
J6061-12 - 1	112018-TIPC-E-U-M	-87	-0.0004	-245.1		PPB	27 Nov 2018 14:14:02	U	SMPL
J6061-13 - 1	112018-DBRI-E-D-B	-90	-0.0012	-245.1		PPB	27 Nov 2018 14:16:20	U	SMPL
J6061-14 - 1	112018-DBRI-E-D-S	-104	-0.0053	-245.1		PPB	27 Nov 2018 14:18:38	U	SMPL
J6061-15 - 1	112018-DBRI-E-U-B	-109	-0.0067	-245.1		PPB	27 Nov 2018 14:20:54	U	SMPL
CCV95 - 1	CCV95	17323	5.0383	-245.1		PPB	27 Nov 2018 14:23:11	U	SMPL
CCB95 - 1	CCB95	-143	-0.0166	-245.1		PPB	27 Nov 2018 14:25:27	U	SMPL
J6061-16 - 1	112018-DBRI-E-U-S	-123	-0.0108	-245.1		PPB	27 Nov 2018 14:27:44	U	SMPL
J6061-16DUP - 1	112018-DBRI-E-U-SDUP	-53	0.0095	-245.1		PPB	27 Nov 2018 14:30:01	U	SMPL
J6061-16MS - 1	112018-DBRI-E-U-SMS	13758	4.0066	-245.1		PPB	27 Nov 2018 14:32:17	U	SMPL
J6061-16MSD - 1	112018-DBRI-E-U-SMSD	13892	4.0454	-245.1		PPB	27 Nov 2018 14:34:33	U	SMPL
J6061-16LX5 - 1		-128	-0.0122	-245.1		PPB	27 Nov 2018 14:36:50	U	SMPL
J6061-16A - 1		14525	4.2286	-245.1		PPB	27 Nov 2018 14:42:22	U	SMPL
CCV96 - 1	CCV96	17438	5.0716	-245.1		PPB	27 Nov 2018 14:44:39	U	SMPL
CCB96 - 1	CCB96	-159	-0.0212	-245.1		PPB	27 Nov 2018 14:46:56	U	SMPL