



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: sweta

Run Number: LB99412

Ambient Barometric Pressure (mmHg): 750

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.0	80.6	81.2	11/27/2018	10:05
2	J6064-04		1	9	>100.0	>212.0	>212.0	11/27/2018	10:40
3	J6064-04DUP		1	9	>100.0	>212.0	>212.0	11/27/2018	11:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 99412

WorkList Name : fp112718

WorkList ID : 119605

Date : 11/27/2018 9:17:39 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/30/2018	Water	J6064-04	Flash Point	Cool 4 deg C	PSEG01	L31	OILY-WATER	11/26/2018	1010A

Date/Time

Received by:

Relinquished by:

11/27/18 9:25 AM
KS
CP

Date/Time

Received by:

Relinquished by:

11/27/18 11:15 AM
CP
KS