


7471B

LB99421
INSTRUMENT ID: CV1

Linear

 μ Abs.:

53299



A= 0.0000e+000

B= 1.8687e-004

C= -2.1817e-002

Rho= 0.9999039

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.018	-0.018	19	0.000	19				
0.2	0.200	0.176	-0.024	1058	0.0 %	1058				
2.5	2.500	2.551	0.051	13769	0.0 %	13769				
5.0	5.000	4.967	-0.033	26698	0.0 %	26698				
7.5	7.500	7.585	0.085	40708	0.0 %	40708				
10.0	10.000	9.938	-0.062	53299	0.0 %	53299				

LB99421 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 27 Nov 2018 15:35:33

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	19	-	0.0000	7471B	PPB	27 Nov 2018 15:55:15	S	Std
0.2 - 1	50.2	1058	-	0.2000	7471B	PPB	27 Nov 2018 15:57:34	S	Std
2.5 - 1	52.5	13769	-	2.5000	7471B	PPB	27 Nov 2018 15:59:52	S	Std
5.0 - 1	55	26698	-	5.0000	7471B	PPB	27 Nov 2018 16:02:09	S	Std
7.5 - 1	57.5	40708	-	7.5000	7471B	PPB	27 Nov 2018 16:04:26	S	Std
10.0 - 1	510	53299	-	10.0000	7471B	PPB	27 Nov 2018 16:06:45	S	Std
ICV91 - 1	ICV91	20820	3.8689	-	7471B	PPB	27 Nov 2018 16:09:49	U	SMPL
ICB91 - 1	ICB91	-98	-0.0401	-	7471B	PPB	27 Nov 2018 16:12:06	U	SMPL
CCV01 - 1	CCV01	26252	4.8840	-	7471B	PPB	27 Nov 2018 16:14:21	U	SMPL
CCB01 - 1	CCB01	-103	-0.0411	-	7471B	PPB	27 Nov 2018 16:16:37	U	SMPL
CRA - 1	CRA	999	0.1649	-	7471B	PPB	27 Nov 2018 16:18:55	U	SMPL
HighStd - 1	HighStd	52390	9.7685	-	7471B	PPB	27 Nov 2018 16:21:10	U	SMPL
ChkStd - 1	ChkStd	37562	6.9975	-	7471B	PPB	27 Nov 2018 16:23:26	U	SMPL
PB115076BL - 1	PBS	-162	-0.0521	-	7471B	PPB	27 Nov 2018 16:25:45	U	SMPL
PB115076BS - 1	LCSS	18320	3.4017	-	7471B	PPB	27 Nov 2018 16:28:03	U	SMPL
J6051-01 - 1	SS-01	3394	0.6124	-	7471B	PPB	27 Nov 2018 16:35:10	U	SMPL
J6051-01DUP - 1	SS-01DUP	3512	0.6345	-	7471B	PPB	27 Nov 2018 16:37:27	U	SMPL
J6051-01MS - 1	SS-01MS	24489	4.5545	-	7471B	PPB	27 Nov 2018 16:42:25	U	SMPL
J6051-01MSD - 1	SS-01MSD	24290	4.5173	-	7471B	PPB	27 Nov 2018 16:44:42	U	SMPL
J6051-02 - 1	SS-02	24353	4.5291	-	7471B	PPB	27 Nov 2018 16:47:00	U	SMPL
CCV02 - 1	CCV02	24812	4.6149	-	7471B	PPB	27 Nov 2018 16:49:18	U	SMPL
CCB02 - 1	CCB02	-147	-0.0493	-	7471B	PPB	27 Nov 2018 16:51:35	U	SMPL
J6060-01 - 1	20180942-COMPOSITE-1	2618	0.4674	-	7471B	PPB	27 Nov 2018 16:53:52	U	SMPL
J6060-02 - 1	20180943-COMPOSITE-2	3309	0.5965	-	7471B	PPB	27 Nov 2018 16:56:09	U	SMPL
J6060-03 - 1	20180944-COMPOSITE-3	3695	0.6687	-	7471B	PPB	27 Nov 2018 16:58:24	U	SMPL
J6060-04 - 1	20180945-COMPOSITE-4	2467	0.4392	-	7471B	PPB	27 Nov 2018 17:00:41	U	SMPL
J6060-05 - 1	20180946-COMPOSITE-5	3084	0.5545	-	7471B	PPB	27 Nov 2018 17:02:58	U	SMPL
J6060-06 - 1	20180947-COMPOSITE-6	10908	2.0166	-	7471B	PPB	27 Nov 2018 17:05:16	U	SMPL
J6062-02 - 1	B-20(1-10)	9355	1.7264	-	7471B	PPB	27 Nov 2018 17:07:33	U	SMPL
J6064-01 - 1	OILY-SOIL	7453	1.3710	-	7471B	PPB	27 Nov 2018 17:09:51	U	SMPL
J6065-01 - 1	USTS-1-2-3	1370	0.2342	-	7471B	PPB	27 Nov 2018 17:12:09	U	SMPL
CCV03 - 1	CCV03	24411	4.5400	-	7471B	PPB	27 Nov 2018 17:14:25	U	SMPL
CCB03 - 1	CCB03	-114	-0.0431	-	7471B	PPB	27 Nov 2018 17:16:41	U	SMPL
J6051-01LX5 - 1		762	0.1206	-	7471B	PPB	27 Nov 2018 17:18:59	U	SMPL
J6051-01A - 1		24891	4.6297	-	7471B	PPB	27 Nov 2018 17:21:15	U	SMPL
CCV04 - 1	CCV04	24733	4.6001	-	7471B	PPB	27 Nov 2018 17:23:32	U	SMPL
CCB04 - 1	CCB04	-118	-0.0439	-	7471B	PPB	27 Nov 2018 17:25:50	U	SMPL