



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 11/28/2018

TEMP1 IN: 103 11/28/2018 07:00 TEMP OUT: 104 11/28/2018 08:05

TEMP2 IN: 104 11/28/2018 08:40 TEMP OUT: 104 11/28/2018 09:50

TEMP3 IN: 105 11/28/2018 11:00 TEMP OUT: 105 11/28/2018 12:40

TEMP4 IN: 105 11/28/2018 13:30 TEMP OUT: 105 11/28/2018 15:00

Run Number: LB99426

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB99426BL	LB99426BL	1.3664	1.3664	100	1.3664	1.3664	1.3664	0.0000	0
2	LB99426BS	LB99426BS	1.3951	1.3951	100	1.4460	1.4460	1.4460	0.0509	509
3	J6078-01	112618-JETT-F-D-B	1.3882	1.3882	200	1.4067	1.4067	1.4067	0.0185	92.5
4	J6078-01DUP	112618-JETT-F-D-BDUP	1.3943	1.3943	200	1.4125	1.4125	1.4125	0.0182	91
5	J6078-02	112618-JETT-F-D-M	1.4057	1.4057	300	1.4174	1.4174	1.4174	0.0117	39
6	J6078-03	112618-JETT-F-D-S	1.3677	1.3677	300	1.3785	1.3785	1.3785	0.0108	36
7	J6078-04	112618-JETT-F-U-B	1.3940	1.3940	200	1.4171	1.4171	1.4171	0.0231	115.5
8	J6078-05	112618-JETT-F-U-M	1.3776	1.3776	200	1.3918	1.3918	1.3918	0.0142	71
9	J6078-06	112618-JETT-F-U-S	1.4010	1.4010	400	1.4145	1.4145	1.4145	0.0135	33.8
10	J6078-07	112618-TIPC-F-D-B	1.3844	1.3844	200	1.4000	1.4000	1.4000	0.0156	78
11	J6078-08	112618-TIPC-F-D-S	1.3808	1.3808	200	1.3970	1.3970	1.3970	0.0162	81
12	J6078-09	112618-TIPC-F-U-B	1.3961	1.3961	200	1.4172	1.4172	1.4172	0.0211	105.5
13	J6078-10	112618-TIPC-F-U-S	1.3947	1.3947	200	1.4093	1.4093	1.4093	0.0146	73
14	J6078-11	112618-DBRI-F-D-B	1.3839	1.3839	200	1.4080	1.4080	1.4080	0.0241	120.5
15	J6078-12	112618-DBRI-F-D-S	1.3664	1.3664	200	1.3898	1.3898	1.3898	0.0234	117

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 11/28/2018

TEMP1 IN: 103 11/28/2018 07:00 TEMP OUT: 104 11/28/2018 08:05

TEMP2 IN: 104 11/28/2018 08:40 TEMP OUT: 104 11/28/2018 09:50

TEMP3 IN: 105 11/28/2018 11:00 TEMP OUT: 105 11/28/2018 12:40

TEMP4 IN: 105 11/28/2018 13:30 TEMP OUT: 105 11/28/2018 15:00

Run Number: LB99426

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
16	J6078-13	112618-DBRI-F-U-B	1.3693	1.3693	200	1.3899	1.3899	1.3899	0.0206	103
17	J6078-14	112618-DBRI-F-U-S	1.3775	1.3775	200	1.3954	1.3954	1.3954	0.0179	89.5
18	J6078-15	112618-CHDE-F-D-B	1.3752	1.3752	200	1.3945	1.3945	1.3945	0.0193	96.5
19	J6078-16	112618-CHDE-F-D-M	1.3976	1.3976	300	1.4141	1.4141	1.4141	0.0165	55
20	J6078-17	112618-CHDE-F-D-S	1.3912	1.3912	300	1.4103	1.4103	1.4103	0.0191	63.7
21	J6078-18	112618-CHDE-F-U-B	1.3984	1.3984	250	1.4188	1.4188	1.4188	0.0204	81.6
22	J6078-19	112618-CHDE-F-U-M	1.4042	1.4042	300	1.4201	1.4201	1.4201	0.0159	53
23	J6078-20	112618-CHDE-F-U-S	1.3560	1.3560	200	1.3646	1.3646	1.3646	0.0086	43
24	J6078-20DUP	112618-CHDE-F-U-SDUP	1.3906	1.3906	200	1.3994	1.3994	1.3994	0.0088	44

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS112818

WorkList ID : 119638

Date : 11/28/2018 7:36:58 AM

2399426

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/03/2018	Water	J6078-01	TSS	Cool 4 deg C	HDR103	L42	112618-JETT-F-D-B	11/26/2018	SM2540 D
12/03/2018	Water	J6078-02	TSS	Cool 4 deg C	HDR103	L42	112618-JETT-F-D-M	11/26/2018	SM2540 D
12/03/2018	Water	J6078-03	TSS	Cool 4 deg C	HDR103	L42	112618-JETT-F-D-S	11/26/2018	SM2540 D
12/03/2018	Water	J6078-04	TSS	Cool 4 deg C	HDR103	L42	112618-JETT-F-U-B	11/26/2018	SM2540 D
12/03/2018	Water	J6078-05	TSS	Cool 4 deg C	HDR103	L42	112618-JETT-F-U-M	11/26/2018	SM2540 D
12/03/2018	Water	J6078-06	TSS	Cool 4 deg C	HDR103	L42	112618-JETT-F-U-S	11/26/2018	SM2540 D
11/29/2018	Water	J6078-07	TSS	Cool 4 deg C	HDR103	L42	112618-TIPC-F-D-B	11/26/2018	SM2540 D
11/29/2018	Water	J6078-08	TSS	Cool 4 deg C	HDR103	L42	112618-TIPC-F-D-S	11/26/2018	SM2540 D
11/29/2018	Water	J6078-09	TSS	Cool 4 deg C	HDR103	L42	112618-TIPC-F-U-B	11/26/2018	SM2540 D
11/29/2018	Water	J6078-10	TSS	Cool 4 deg C	HDR103	L42	112618-TIPC-F-U-S	11/26/2018	SM2540 D
12/03/2018	Water	J6078-11	TSS	Cool 4 deg C	HDR103	L42	112618-DBRI-F-D-B	11/26/2018	SM2540 D
12/03/2018	Water	J6078-12	TSS	Cool 4 deg C	HDR103	L42	112618-DBRI-F-D-S	11/26/2018	SM2540 D
12/03/2018	Water	J6078-13	TSS	Cool 4 deg C	HDR103	L42	112618-DBRI-F-U-B	11/26/2018	SM2540 D
12/03/2018	Water	J6078-14	TSS	Cool 4 deg C	HDR103	L42	112618-DBRI-F-U-S	11/26/2018	SM2540 D
12/03/2018	Water	J6078-15	TSS	Cool 4 deg C	HDR103	L42	112618-CHDE-F-D-B	11/26/2018	SM2540 D
12/03/2018	Water	J6078-16	TSS	Cool 4 deg C	HDR103	L42	112618-CHDE-F-D-M	11/26/2018	SM2540 D
12/03/2018	Water	J6078-17	TSS	Cool 4 deg C	HDR103	L42	112618-CHDE-F-D-S	11/26/2018	SM2540 D
12/03/2018	Water	J6078-18	TSS	Cool 4 deg C	HDR103	L42	112618-CHDE-F-U-B	11/26/2018	SM2540 D
12/03/2018	Water	J6078-19	TSS	Cool 4 deg C	HDR103	L42	112618-CHDE-F-U-M	11/26/2018	SM2540 D
12/03/2018	Water	J6078-20	TSS	Cool 4 deg C	HDR103	L42	112618-CHDE-F-U-S	11/26/2018	SM2540 D

fp-7042602

Date/Time 11/28/18, 08:00
 Received by: AS
 Relinquished by: CJ

Date/Time 11/28/18, 13:00
 Received by: [Signature]
 Relinquished by: [Signature]